



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99684

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.0	4.01	12/07/2018	17:50
2	CAL2	1	Water			20.2	7.00	12/07/2018	17:52
3	CAL3	1	Water			20.1	10.04	12/07/2018	17:54
4	ICV	1	Water			19.9	7.01	12/07/2018	17:56
5	CCV1	1	Water			20.2	2.00	12/07/2018	17:58
6	J6196-02	1	Solid	20.02	20	19.9	8.29	12/07/2018	18:00
7	J6196-02DUP	1	Solid	20.03	20	20.0	8.30	12/07/2018	18:02
8	J6196-04	1	Solid	20.03	20	19.2	8.29	12/07/2018	18:04
9	J6196-06	1	Solid	20.04	20	19.2	8.44	12/07/2018	18:06
10	J6197-02	1	Solid	20.02	20	19.4	7.48	12/07/2018	18:08
11	J6214-44	1	Solid	20.02	20	20.6	8.54	12/07/2018	18:10
12	J6290-02	1	Solid	20.03	20	19.7	7.22	12/07/2018	18:12
13	J6296-01	1	Solid	20.03	20	20.2	8.51	12/07/2018	18:14
14	J6302-03	1	Solid	20.02	20	19.5	6.84	12/07/2018	18:16
15	J6306-03	1	Solid	20.02	20	19.1	7.71	12/07/2018	18:20
16	CCV2	1	Water			20.2	12.00	12/07/2018	18:22
17	J6309-01	1	Solid	20.03	20	20.1	8.99	12/07/2018	18:24
18	J6309-03	1	Solid	20.02	20	19.9	10.76	12/07/2018	18:26
19	J6309-03DUP	1	Solid	20.03	20	20.0	10.77	12/07/2018	18:28
20	CCV3	1	Water			20.1	2.00	12/07/2018	18:30

WORKLIST(Hardcopy Internal Chain)

UB99684

WorkList Name : CORROSIONITY120718

WorkList ID : 120078

Date : 12/7/2018 4:41:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2018	Solid	J6196-02	Corrosivity	Cool 4 deg C	PSEG05	L31	EO-03-120718-A	12/07/2018	9045D
12/07/2018	Solid	J6196-04	Corrosivity	Cool 4 deg C	PSEG05	L31	EO-03-120718-B	12/07/2018	9045D
12/07/2018	Solid	J6196-06	Corrosivity	Cool 4 deg C	PSEG05	L31	EO-03-120718-C	12/07/2018	9045D
12/07/2018	Solid	J6197-02	Corrosivity	Cool 4 deg C	PSEG05	L31	PL-02-120718	12/07/2018	9045D
12/06/2018	Solid	J6214-44	Corrosivity	Cool 4 deg C	AKRF02		SB-5-(12.5-14.5)	12/06/2018	9045D
12/07/2018	Solid	J6290-02	Corrosivity	Cool 4 deg C	PSEG01	L31	SOIL	12/07/2018	9045D
12/06/2018	Solid	J6296-01	Corrosivity	Cool 4 deg C	USAR03	L22	CB-1206-S-TCLP-GRID	12/06/2018	9045D
12/07/2018	Solid	J6302-03	Corrosivity	Cool 4 deg C	PSEG01	L31	K-120718-RO	12/07/2018	9045D
12/07/2018	Solid	J6306-03	Corrosivity	Cool 4 deg C	PSEG01	L31	VB-16204	12/07/2018	9045D
12/07/2018	Solid	J6309-01	Corrosivity	Cool 4 deg C	PARS02	L52	WC-20181207-SOIL	12/07/2018	9045D
12/07/2018	Solid	J6309-03	Corrosivity	Cool 4 deg C	PARS02	L52	WC-20181207-CD	12/07/2018	9045D

Date/Time 12/07/18 5:40
 Received by: SP
 Relinquished by: CP

Date/Time 12/07/18 6:40
 Received by: CP
 Relinquished by: SP