



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99685

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.0	4.01	12/07/2018	17:50
2	CAL2	1	Water			20.2	7.00	12/07/2018	17:52
3	CAL3	1	Water			20.1	10.04	12/07/2018	17:54
4	ICV	1	Water			19.9	7.01	12/07/2018	17:56
5	CCV1	1	Water			20.1	2.00	12/07/2018	18:30
6	J6269-04	1	Water			21.4	6.61	12/07/2018	18:32
7	J6269-04DUP	1	Water			21.5	6.62	12/07/2018	18:34
8	J6309-02	1	Water			16.1	7.42	12/07/2018	18:36
9	CCV2	1	Water			20.2	12.00	12/07/2018	18:38

WORKLIST(Hardcopy Internal Chain)

LB 199685

WorkList Name : corrosivity120718-1

WorkList ID : 120053

Date : 12/7/2018 10:30:35 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2018	Water	J6309-02	Corrosivity	Cool 4 deg C	PARS02	L52	WC-20181207-WATER	12/07/2018	9040C
12/06/2018	Water	J6269-04	Corrosivity	Cool 4 deg C	PSEG01	L31	OILY-WATER	12/06/2018	9040C

Date/Time 12/07/18 6:10
 Received by: SP
 Relinquished by: CP

Date/Time 12/07/18 6:40
 Received by: SP
 Relinquished by: SP