



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99693

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	12/10/2018	09:35
2	J6214-55		1	9	>100.0	>212.0	>212.0	12/10/2018	09:58
3	J6214-55DUP		1	9	>100.0	>212.0	>212.0	12/10/2018	10:20
4	J6309-02		1	8	>100.0	>212.0	>212.0	12/10/2018	10:45
5	J6317-01		1	7	>100.0	>212.0	>212.0	12/10/2018	11:10
6	J6318-02		1	8	>100.0	>212.0	>212.0	12/10/2018	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 99693

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp121018

WorkList ID : 120124

Date : 12/10/2018 9:07:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Water	J6214-55	Flash Point	Cool 4 deg C	AKRF02		TW-2	12/06/2018	1010A
12/21/2018	Water	J6309-02	Flash Point	Cool 4 deg C	PARS02	L52	WC-20181207-WATER	12/07/2018	1010A
12/21/2018	Water	J6317-01	Flash Point	Cool 4 deg C	TULL02	L23	MH-718	12/07/2018	1010A
12/21/2018	Water	J6318-02	Flash Point	Cool 4 deg C	TULL02	L22	WEST-8TH-BOX-COVER-TE	12/07/2018	1010A

Date/Time 12/10/18 9:10 AM
Received by: RS
Relinquished by: CP

Date/Time 12/10/18 11:45 AM
Received by: CP
Relinquished by: RS