

LB 99698


7471B

INSTRUMENT ID: CX2

Linear

 μ Abs.:

57997



A= 0.0000e+000

B= 1.7362e-004

C= 3.5903e-002

Rho= 0.9997221

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.036	0.036	0	0.000	0				
0.2	0.200	0.204	0.004	968	0.0 %	968				
2.5	2.500	2.476	-0.024	14056	0.0 %	14056				
5.0	5.000	5.052	0.052	28889	0.0 %	28889				
7.5	7.500	7.327	-0.173	41997	0.0 %	41997				
10.0	10.000	10.105	0.105	57997	0.0 %	57997				

LB99698 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Dec 2018 14:42:01

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	0	-	0.00007471B	PPB	10 Dec 2018 14:43:12	S	Std
0.2 - 1	50.2	968	-	0.20007471B	PPB	10 Dec 2018 14:45:31	S	Std
2.5 - 1	52.5	14056	-	2.50007471B	PPB	10 Dec 2018 14:47:49	S	Std
5.0 - 1	55	28889	-	5.00007471B	PPB	10 Dec 2018 14:50:06	S	Std
7.5 - 1	57.5	41997	-	7.50007471B	PPB	10 Dec 2018 14:52:24	S	Std
10.0 - 1	59	57997	-	10.00007471B	PPB	10 Dec 2018 15:00:06	S	Std
ICV14 - 1	ICV14	22485	3.9397	-7471B	PPB	10 Dec 2018 15:02:45	U	SMPL
ICB14 - 1	ICB14	-154	0.0092	-7471B	PPB	10 Dec 2018 15:05:00	U	SMPL
CCV22 - 1	CCV22	28294	4.9482	-7471B	PPB	10 Dec 2018 15:07:17	U	SMPL
CCB22 - 1	CCB22	-151	0.0097	-7471B	PPB	10 Dec 2018 15:09:33	U	SMPL
CRA - 1	CRA	926	0.1967	-7471B	PPB	10 Dec 2018 15:11:50	U	SMPL
HighStd - 1	HighStd	59070	10.2914	-7471B	PPB	10 Dec 2018 15:14:06	U	SMPL
ChkStd - 1	ChkStd	39168	6.8361	-7471B	PPB	10 Dec 2018 15:19:01	U	SMPL
PB115459BL - 1	PBS	-121	0.0149	-7471B	PPB	10 Dec 2018 15:21:17	U	SMPL
PB115459BS - 1	LCSS	20298	3.5600	-7471B	PPB	10 Dec 2018 15:26:21	U	SMPL
J6214-41 - 1	SB-5-(0-2)	24614	4.3093	-7471B	PPB	10 Dec 2018 15:28:38	U	SMPL
J6214-43 - 1	SB-5-(12.5-14.5)	8248	1.4679	-7471B	PPB	10 Dec 2018 15:30:56	U	SMPL
J6214-45 - 1	SB-6-(0-2)	28414	4.9690	-7471B	PPB	10 Dec 2018 15:33:16	U	SMPL
J6214-47 - 1	SB-6-(13-15)	18130	3.1836	-7471B	PPB	10 Dec 2018 15:35:34	U	SMPL
J6214-49 - 1	SB-7-(0-2)	5685	1.0229	-7471B	PPB	10 Dec 2018 15:37:51	U	SMPL
CCV23 - 1	CCV23	28524	4.9881	-7471B	PPB	10 Dec 2018 15:40:08	U	SMPL
CCB23 - 1	CCB23	-146	0.0106	-7471B	PPB	10 Dec 2018 15:42:24	U	SMPL
J6214-50 - 1	SB-7-(12.5-14.5)	4448	0.8081	-7471B	PPB	10 Dec 2018 15:44:40	U	SMPL
J6269-01 - 1	OILY-SOIL	4850	0.8779	-7471B	PPB	10 Dec 2018 15:46:56	U	SMPL
J6275-01 - 1	BF-01-120418	939	0.1989	-7471B	PPB	10 Dec 2018 15:49:13	U	SMPL
J6275-02 - 1	BF-02-120418	1127	0.2316	-7471B	PPB	10 Dec 2018 15:51:29	U	SMPL
J6275-03 - 1	BF-03-120418	719	0.1607	-7471B	PPB	10 Dec 2018 15:53:46	U	SMPL
J6275-04 - 1	BF-04-120418	403	0.1059	-7471B	PPB	10 Dec 2018 15:56:03	U	SMPL
J6275-05 - 1	BF-05-120418	896	0.1915	-7471B	PPB	10 Dec 2018 15:58:21	U	SMPL
J6282-01 - 1	AU-7-120618	2807	0.5232	-7471B	PPB	10 Dec 2018 16:00:38	U	SMPL
J6282-01DUP - 1	AU-7-120618DUP	2930	0.5446	-7471B	PPB	10 Dec 2018 16:02:56	U	SMPL
J6282-01MS - 1	AU-7-120618MS	21692	3.8020	-7471B	PPB	10 Dec 2018 16:05:12	U	SMPL
CCV24 - 1	CCV24	28272	4.9444	-7471B	PPB	10 Dec 2018 16:07:29	U	SMPL
CCB24 - 1	CCB24	-142	0.0112	-7471B	PPB	10 Dec 2018 16:09:46	U	SMPL
J6282-01MSD - 1	AU-7-120618MSD	22524	3.9464	-7471B	PPB	10 Dec 2018 16:12:03	U	SMPL
J6282-04 - 1	AU-2-120618	-33	0.0302	-7471B	PPB	10 Dec 2018 16:14:19	U	SMPL
J6282-01LX5 - 1		525	0.1271	-7471B	PPB	10 Dec 2018 16:16:36	U	SMPL
J6282-01A - 1		25212	4.4131	-7471B	PPB	10 Dec 2018 16:18:53	U	SMPL
CCV25 - 1	CCV25	28386	4.9642	-7471B	PPB	10 Dec 2018 16:21:11	U	SMPL
CCB25 - 1	CCB25	-150	0.0099	-7471B	PPB	10 Dec 2018 16:23:28	U	SMPL