



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : sweta

Run Number: LB99702

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.0 Control Limits[+/- 0.1].

True Value of CCV2 = 12.0 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.1	4.01	12/10/2018	14:35
2	CAL2	1	Water			21.0	7.00	12/10/2018	14:36
3	CAL3	1	Water			21.0	10.02	12/10/2018	14:37
4	ICV	1	Water			21.0	7.01	12/10/2018	14:38
5	CCV1	1	Water			21.1	2.01	12/10/2018	14:40
6	J6330-01	1	Water			13.0	5.86	12/10/2018	14:41
7	J6330-01DUP	1	Water			13.1	5.87	12/10/2018	14:42
8	CCV2	1	Water			21.0	12.02	12/10/2018	14:44

WORKLIST(Hardcopy Internal Chain)

LB 99702

WorkList Name : pH121018-1

WorkList ID : 120148

Date : 12/10/2018 2:30:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2018	Water	J6330-01	pH	Cool 4 deg C	STER01	M11	POOL-WATER	12/07/2018	SM 4500-H B

Date/Time 12/10/18 2:35
 Received by: SP
 Relinquished by: CP

Date/Time 12-10-18 4:00 PM
 Received by: SP
 Relinquished by: SP