

L399716


7470A

INSTRUMENT ID: CVJ

Linear

μ Abs.:

37435



A= 0.0000e+000

B= 2.6762e-004

C= -5.0628e-002

Rho= 0.9998726

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	199	0.000	199				
0.2	0.200	0.191	-0.009	904	0.0 %	904				
2.5	2.500	2.435	-0.065	9289	0.0 %	9289				
5.0	5.000	5.123	0.123	19332	0.0 %	19332				
7.5	7.500	7.480	-0.020	28141	0.0 %	28141				
10.0	10.000	9.968	-0.032	37435	0.0 %	37435				

LB99716 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Dec 2018 08:22:18

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	199	-	0.0000	7470A	PPB	11 Dec 2018 08:24:53	S	Std
0.2 - 1	52.5	904	-	0.2000	7470A	PPB	11 Dec 2018 08:27:02	S	Std
2.5 - 1	55	9289	-	2.5000	7470A	PPB	11 Dec 2018 08:29:11	S	Std
5.0 - 1	57.5	19332	-	5.0000	7470A	PPB	11 Dec 2018 08:31:19	S	Std
7.5 - 1	510	28141	-	7.5000	7470A	PPB	11 Dec 2018 08:33:28	S	Std
10.0 - 1		37435	-	10.0000	7470A	PPB	11 Dec 2018 08:35:35	S	Std
ICV15 - 1	ICV15	14909	3.9393	-	7470A	PPB	11 Dec 2018 08:38:11	U	SMPL
ICB15 - 1	ICB15	106	-0.0223	-	7470A	PPB	11 Dec 2018 08:40:17	U	SMPL
CCV26 - 1	CCV26	18956	5.0223	-	7470A	PPB	11 Dec 2018 08:42:24	U	SMPL
CCB26 - 1	CCB26	108	-0.0217	-	7470A	PPB	11 Dec 2018 08:44:30	U	SMPL
CRA - 1	CRA	880	0.1849	-	7470A	PPB	11 Dec 2018 08:46:38	U	SMPL
HighStd - 1	HighStd	36702	9.7714	-	7470A	PPB	11 Dec 2018 08:48:44	U	SMPL
ChkStd - 1	ChkStd	26003	6.9082	-	7470A	PPB	11 Dec 2018 08:50:51	U	SMPL
PB115512BL - 1	PBW	79	-0.0295	-	7470A	PPB	11 Dec 2018 08:55:37	U	SMPL
PB115512BS - 1	LCSW	13628	3.5964	-	7470A	PPB	11 Dec 2018 08:57:44	U	SMPL
J6196-02 - 1	EO-03-120718-A	-32	-0.0592	-	7470A	PPB	11 Dec 2018 08:59:51	U	SMPL
J6196-02DUP - 1	EO-03-120718-ADUP	-10	-0.0533	-	7470A	PPB	11 Dec 2018 09:02:00	U	SMPL
J6196-02MS - 1	EO-03-120718-AMS	12568	3.3128	-	7470A	PPB	11 Dec 2018 09:04:09	U	SMPL
J6196-02MSD - 1	EO-03-120718-AMSD	12765	3.3655	-	7470A	PPB	11 Dec 2018 09:06:17	U	SMPL
J6196-04 - 1	EO-03-120718-B	-37	-0.0605	-	7470A	PPB	11 Dec 2018 09:08:24	U	SMPL
CCV27 - 1	CCV27	18572	4.9195	-	7470A	PPB	11 Dec 2018 09:10:30	U	SMPL
CCB27 - 1	CCB27	47	-0.0381	-	7470A	PPB	11 Dec 2018 09:12:37	U	SMPL
J6196-06 - 1	EO-03-120718-C	-103	-0.0782	-	7470A	PPB	11 Dec 2018 09:14:44	U	SMPL
J6197-02 - 1	PL-02-120718	-36	-0.0603	-	7470A	PPB	11 Dec 2018 09:16:51	U	SMPL
J6247-19 - 1	BS-03-08	530	0.0912	-	7470A	PPB	11 Dec 2018 09:18:58	U	SMPL
J6260-03 - 1	SB-01-COMP	-2	-0.0512	-	7470A	PPB	11 Dec 2018 09:21:05	U	SMPL
J6260-06 - 1	SB-02-COMP	-58	-0.0661	-	7470A	PPB	11 Dec 2018 09:23:13	U	SMPL
J6260-09 - 1	SB-03-COMP	-21	-0.0562	-	7470A	PPB	11 Dec 2018 09:25:21	U	SMPL
J6260-12 - 1	SB-04-COMP	-19	-0.0557	-	7470A	PPB	11 Dec 2018 09:27:29	U	SMPL
J6269-02 - 1	OILY-SOIL	-31	-0.0589	-	7470A	PPB	11 Dec 2018 09:29:38	U	SMPL
J6282-02 - 1	AU-7-120618	-55	-0.0653	-	7470A	PPB	11 Dec 2018 09:31:46	U	SMPL
J6282-05 - 1	AU-2-120618	5	-0.0493	-	7470A	PPB	11 Dec 2018 09:33:53	U	SMPL
CCV28 - 1	CCV28	18383	4.8689	-	7470A	PPB	11 Dec 2018 09:36:01	U	SMPL
CCB28 - 1	CCB28	40	-0.0399	-	7470A	PPB	11 Dec 2018 09:38:07	U	SMPL
J6290-02 - 1	SOIL	0	-0.0506	-	7470A	PPB	11 Dec 2018 09:40:15	U	SMPL
J6296-01 - 1	CB-1206-S-TCLP-GRID	-6	-0.0522	-	7470A	PPB	11 Dec 2018 09:42:22	U	SMPL
J6302-03 - 1	K-120718-RO	11	-0.0477	-	7470A	PPB	11 Dec 2018 09:44:29	U	SMPL
J6306-03 - 1	VB-16204	10	-0.0480	-	7470A	PPB	11 Dec 2018 09:46:36	U	SMPL
J6309-01 - 1	WC-20181207-SOIL	6	-0.0490	-	7470A	PPB	11 Dec 2018 09:48:44	U	SMPL
J6309-03 - 1	WC-20181207-CD	8	-0.0485	-	7470A	PPB	11 Dec 2018 09:50:52	U	SMPL
PB115513BL - 1	PBW	18	-0.0458	-	7470A	PPB	11 Dec 2018 09:53:00	U	SMPL
PB115513BS - 1	LCSW	13728	3.6232	-	7470A	PPB	11 Dec 2018 09:57:53	U	SMPL
J6269-04 - 1	OILY-WATER	-43	-0.0621	-	7470A	PPB	11 Dec 2018 10:00:02	U	SMPL
J6309-02 - 1	WC-20181207-WATER	-8	-0.0528	-	7470A	PPB	11 Dec 2018 10:02:09	U	SMPL
CCV29 - 1	CCV29	18561	4.9166	-	7470A	PPB	11 Dec 2018 10:04:16	U	SMPL
CCB29 - 1	CCB29	93	-0.0257	-	7470A	PPB	11 Dec 2018 10:06:23	U	SMPL
J6309-02DUP - 1	WC-20181207-WATERDUP	-28	-0.0581	-	7470A	PPB	11 Dec 2018 10:08:31	U	SMPL
J6309-02MS - 1	WC-20181207-WATERMS	13007	3.4302	-	7470A	PPB	11 Dec 2018 10:10:38	U	SMPL
J6309-02MSD - 1	WC-20181207-WATERMSD	12683	3.3435	-	7470A	PPB	11 Dec 2018 10:12:45	U	SMPL
J6317-02 - 1	MH-718	-39	-0.0611	-	7470A	PPB	11 Dec 2018 10:14:53	U	SMPL
J6318-02 - 1	WEST-8TH-BOX-COVER-TEST-PIT	23	-0.0445	-	7470A	PPB	11 Dec 2018 10:17:00	U	SMPL
PB115514BL - 1	PBW	196	0.0018	-	7470A	PPB	11 Dec 2018 10:19:08	U	SMPL
PB115514BS - 1	LCSW	14334	3.7854	-	7470A	PPB	11 Dec 2018 10:21:16	U	SMPL
J6310-07 - 1	V26	318	0.0345	-	7470A	PPB	11 Dec 2018 10:23:25	U	SMPL
J6316-02 - 1	BORING-SAMPLE-SIGNAL-TOWER-B	218	0.0077	-	7470A	PPB	11 Dec 2018 10:25:34	U	SMPL
J6316-06 - 1	BORING-SAMPLE-CBH-591	210	0.0056	-	7470A	PPB	11 Dec 2018 10:27:41	U	SMPL
CCV30 - 1	CCV30	18395	4.8722	-	7470A	PPB	11 Dec 2018 10:29:48	U	SMPL
CCB30 - 1	CCB30	123	-0.0177	-	7470A	PPB	11 Dec 2018 10:31:55	U	SMPL
J6316-10 - 1	AVE-X-B-BALL-FIELD-S-1	202	0.0034	-	7470A	PPB	11 Dec 2018 10:34:03	U	SMPL
J6316-14 - 1	AVE-X-B-BALL-FIELD-S-2	193	0.0010	-	7470A	PPB	11 Dec 2018 10:36:10	U	SMPL
J6316-18 - 1	AVE-X-B-BALL-FIELD-S-3	227	0.0101	-	7470A	PPB	11 Dec 2018 10:38:18	U	SMPL
J6316-22 - 1	BORING-SAMPLE-SHALWELL-YARD	245	0.0149	-	7470A	PPB	11 Dec 2018 10:40:25	U	SMPL
J6316-22DUP - 1	BORING-SAMPLE-SHALWELL-YARDDUP	205	0.0042	-	7470A	PPB	11 Dec 2018 10:42:33	U	SMPL
J6316-22MS - 1	BORING-SAMPLE-SHALWELL-YARDMS	13144	3.4669	-	7470A	PPB	11 Dec 2018 10:44:40	U	SMPL
J6316-22MSD - 1	BORING-SAMPLE-SHALWELL-YARDMSD	13215	3.4859	-	7470A	PPB	11 Dec 2018 10:46:48	U	SMPL
PB115516BL - 1	PBW	-12	-0.0538	-	7470A	PPB	11 Dec 2018 10:48:57	U	SMPL
PB115516BS - 1	LCSW	14509	3.8322	-	7470A	PPB	11 Dec 2018 10:51:06	U	SMPL
J6214-51 - 1	TW-3	-2	-0.0512	-	7470A	PPB	11 Dec 2018 10:53:13	U	SMPL
CCV31 - 1	CCV31	18675	4.9471	-	7470A	PPB	11 Dec 2018 10:55:21	U	SMPL
CCB31 - 1	CCB31	96	-0.0249	-	7470A	PPB	11 Dec 2018 10:57:28	U	SMPL
J6214-52 - 1	TW-3	21	-0.0450	-	7470A	PPB	11 Dec 2018 10:59:36	U	SMPL
J6214-53 - 1	TW-5	98	-0.0244	-	7470A	PPB	11 Dec 2018 11:01:43	U	SMPL
J6214-54 - 1	TW-5	-33	-0.0595	-	7470A	PPB	11 Dec 2018 11:03:51	U	SMPL
J6288-02 - 1	BPS1-TT-MW302I2-20181205	-62	-0.0672	-	7470A	PPB	11 Dec 2018 11:05:58	U	SMPL
J6288-02DUP - 1	BPS1-TT-MW302I2-20181205DUP	-63	-0.0675	-	7470A	PPB	11 Dec 2018 11:08:06	U	SMPL
J6288-02MS - 1	BPS1-TT-MW302I2-20181205MS	13251	3.4955	-	7470A	PPB	11 Dec 2018 11:10:14	U	SMPL
J6288-02MSD - 1	BPS1-TT-MW302I2-20181205MSD	13339	3.5191	-	7470A	PPB	11 Dec 2018 11:12:22	U	SMPL
J6288-03 - 1	BPS1-TT-MW302I2-20181205	-123	-0.0835	-	7470A	PPB	11 Dec 2018 11:14:30	U	SMPL
J6288-04 - 1	BPS1-TT-MW302D-20181205	-83	-0.0728	-	7470A	PPB	11 Dec 2018 11:16:39	U	SMPL
J6288-05 - 1	BPS1-TT-MW302D-20181205	-85	-0.0734	-	7470A	PPB	11 Dec 2018 11:18:47	U	SMPL
CCV32 - 1	CCV32	18768	4.9773	-	7470A	PPB	11 Dec 2018 11:20:55	U	SMPL
CCB32 - 1	CCB32	116	-0.0196	-	7470A	PPB	11 Dec 2018 11:23:02	U	SMPL
J6288-06 - 1	BPS1-TT-MW301S-20181206	-77	-0.0712	-	7470A	PPB	11 Dec 2018 11:25:10	U	SMPL
J6288-07 - 1	BPS1-TT-MW301S-20181206	-59	-0.0664	-	7470A	PPB	11 Dec 2018 11:27:19	U	SMPL
J6288-08 - 1	BPS1-TT-MW301D-20181206	-55	-0.0653	-	7470A	PPB	11 Dec 2018 11:29:26	U	SMPL
J6288-09 - 1	BPS1-TT-MW301D-20181206	-69	-0.0691	-	7470A	PPB	11 Dec 2018 11:31:34	U	SMPL
J6288-10 - 1	BPS1-TT-MW308I-20181206	-60	-0.0667	-	7470A	PPB	11 Dec 2018 11:33:42	U	SMPL
J6288-11 - 1	BPS1-TT-MW308I-20181206	-67	-0.0686	-	7470A	PPB	11 Dec 2018 11:35:50	U	SMPL
J6288-12 - 1	BPS1-TT-MW309S-20181206	-57	-0.0659	-	7470A	PPB	11 Dec 2018 11:37:58	U	SMPL
J6288-13 - 1	BPS1-TT-MW309S-20181206	-62	-0.0672	-	7470A	PPB	11 Dec 2018 11:40:07	U	SMPL
J6288-14 - 1	BPS1-TT-MW309I-20181206	-52	-0.0645	-	7470A	PPB	11 Dec 2018 11:42:16	U	SMPL
J6288-15 - 1	BPS1-TT-MW309I-20181206	-59	-0.0664	-	7470A	PPB	11 Dec 2018 11:44:24	U	SMPL
CCV33 - 1	CCV33	18850	4.9939	-	7470A	PPB	11 Dec 2018 11:46:32	U	SMPL

LB99716 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Dec 2018 08:22:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB33 - 1	CCB33	114	-0.0201	-7470A	PPB		11 Dec 2018 11:48:40	U	SMPL
J6288-16 - 1	BPS1-TT-MW308D-20181206	-57	-0.0659	-7470A	PPB		11 Dec 2018 11:50:48	U	SMPL
J6288-17 - 1	BPS1-TT-MW308D-20181206	-40	-0.0613	-7470A	PPB		11 Dec 2018 11:52:56	U	SMPL
PB115517BL - 1	PBW	190	0.0002	-7470A	PPB		11 Dec 2018 11:55:04	U	SMPL
PB115517BS - 1	LCSW	12702	3.3486	-7470A	PPB		11 Dec 2018 11:57:12	U	SMPL
J6314-02 - 1	BPS1-TT-MW314I-20181206	81	-0.0290	-7470A	PPB		11 Dec 2018 11:59:20	U	SMPL
J6314-03 - 1	BPS1-TT-MW314I-20181206	135	-0.0145	-7470A	PPB		11 Dec 2018 12:01:27	U	SMPL
J6314-04 - 1	BPS1-TT-MW314S-20181206	94	-0.0255	-7470A	PPB		11 Dec 2018 12:03:35	U	SMPL
J6314-05 - 1	BPS1-TT-MW314S-20181206	110	-0.0212	-7470A	PPB		11 Dec 2018 12:05:44	U	SMPL
J6314-06 - 1	BPS1-TT-MW313S-20181206	134	-0.0148	-7470A	PPB		11 Dec 2018 12:07:53	U	SMPL
J6314-07 - 1	BPS1-TT-MW313S-20181206	119	-0.0188	-7470A	PPB		11 Dec 2018 12:10:01	U	SMPL
CCV34 - 1	CCV34	18133	4.8020	-7470A	PPB		11 Dec 2018 12:12:09	U	SMPL
CCB34 - 1	CCB34	139	-0.0134	-7470A	PPB		11 Dec 2018 12:14:17	U	SMPL
J6314-08 - 1	BPS1-TT-MW303S-20181207	158	-0.0083	-7470A	PPB		11 Dec 2018 12:16:25	U	SMPL
J6314-09 - 1	BPS1-TT-MW303S-20181207	173	-0.0043	-7470A	PPB		11 Dec 2018 12:18:33	U	SMPL
J6314-10 - 1	BPS1-TT-MW303I2-20181207	123	-0.0177	-7470A	PPB		11 Dec 2018 12:20:41	U	SMPL
J6314-10DUP - 1	BPS1-TT-MW303I2-20181207DUP	128	-0.0164	-7470A	PPB		11 Dec 2018 12:22:49	U	SMPL
J6314-10MS - 1	BPS1-TT-MW303I2-20181207MS	13229	3.4897	-7470A	PPB		11 Dec 2018 12:24:57	U	SMPL
J6314-10MSD - 1	BPS1-TT-MW303I2-20181207MSD	12789	3.3719	-7470A	PPB		11 Dec 2018 12:27:05	U	SMPL
J6314-11 - 1	BPS1-TT-MW303I2-20181207	73	-0.0311	-7470A	PPB		11 Dec 2018 12:29:13	U	SMPL
J6314-16 - 1	BPS1-TT-MW303D-20181207	124	-0.0174	-7470A	PPB		11 Dec 2018 12:31:21	U	SMPL
J6314-17 - 1	BPS1-TT-MW303D-20181207	112	-0.0207	-7470A	PPB		11 Dec 2018 12:33:30	U	SMPL
J6314-18 - 1	BPS1-TT-MW301I-20181207	121	-0.0182	-7470A	PPB		11 Dec 2018 12:35:39	U	SMPL
CCV35 - 1	CCV35	18372	4.8660	-7470A	PPB		11 Dec 2018 12:37:47	U	SMPL
CCB35 - 1	CCB35	134	-0.0148	-7470A	PPB		11 Dec 2018 12:39:55	U	SMPL
J6314-19 - 1	BPS1-TT-MW301I-20181207	147	-0.0113	-7470A	PPB		11 Dec 2018 12:42:03	U	SMPL
J6314-20 - 1	BPS1-TT-MW303I1-20181207	674	0.1297	-7470A	PPB		11 Dec 2018 12:44:11	U	SMPL
J6314-21 - 1	BPS1-TT-MW303I1-20181207	397	0.0556	-7470A	PPB		11 Dec 2018 12:46:19	U	SMPL
J6314-22 - 1	BP-DUP03-20181207	150	-0.0105	-7470A	PPB		11 Dec 2018 12:48:27	U	SMPL
J6314-23 - 1	BP-DUP03-20181207	146	-0.0116	-7470A	PPB		11 Dec 2018 12:50:35	U	SMPL
J6314-24 - 1	BP-EB02-20181207	150	-0.0105	-7470A	PPB		11 Dec 2018 12:52:44	U	SMPL
J6314-25 - 1	BP-EB02-20181207	137	-0.0140	-7470A	PPB		11 Dec 2018 12:54:52	U	SMPL
J6317-01 - 1	MH-718	160	-0.0078	-7470A	PPB		11 Dec 2018 12:57:01	U	SMPL
CCV36 - 1	CCV36	18687	4.9503	-7470A	PPB		11 Dec 2018 13:01:17	U	SMPL
CCB36 - 1	CCB36	119	-0.0188	-7470A	PPB		11 Dec 2018 13:03:24	U	SMPL
J6196-02LX5 - 1		21	-0.0450	-7470A	PPB		11 Dec 2018 13:05:32	U	SMPL
J6196-02A - 1		14864	3.9272	-7470A	PPB		11 Dec 2018 13:07:41	U	SMPL
J6309-02LX5 - 1		-24	-0.0571	-7470A	PPB		11 Dec 2018 13:09:49	U	SMPL
J6309-02A - 1		14625	3.8633	-7470A	PPB		11 Dec 2018 13:11:58	U	SMPL
J6316-22LX5 - 1		11	-0.0477	-7470A	PPB		11 Dec 2018 13:14:07	U	SMPL
J6316-22A - 1		18913	5.0108	-7470A	PPB		11 Dec 2018 13:16:15	U	SMPL
J6288-02LX5 - 1		-70	-0.0694	-7470A	PPB		11 Dec 2018 13:18:24	U	SMPL
J6288-02A - 1		14739	3.8938	-7470A	PPB		11 Dec 2018 13:20:32	U	SMPL
J6314-10LX5 - 1		-21	-0.0562	-7470A	PPB		11 Dec 2018 13:22:41	U	SMPL
J6314-10A - 1		14993	3.9617	-7470A	PPB		11 Dec 2018 13:24:51	U	SMPL
CCV37 - 1	CCV37	19187	5.0841	-7470A	PPB		11 Dec 2018 13:26:59	U	SMPL
CCB37 - 1	CCB37	136	-0.0142	-7470A	PPB		11 Dec 2018 13:29:08	U	SMPL
PB115512TB - 1	PB115447TB	46	-0.0383	-7470A	PPB		11 Dec 2018 13:31:16	U	SMPL
PB115513TB - 1	PB115468TB	34	-0.0415	-7470A	PPB		11 Dec 2018 13:33:25	U	SMPL
PB115514TB - 1	PB115470TB	235	0.0123	-7470A	PPB		11 Dec 2018 13:35:33	U	SMPL
CCV38 - 1	CCV38	18381	4.8684	-7470A	PPB		11 Dec 2018 13:37:42	U	SMPL
CCB38 - 1	CCB38	148	-0.0110	-7470A	PPB		11 Dec 2018 13:39:50	U	SMPL