



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99718

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	12/10/2018	16:00
2	CAL2	1	Water			20.1	7.00	12/10/2018	16:02
3	CAL3	1	Water			20.2	10.04	12/10/2018	16:04
4	ICV	1	Water			20.3	7.01	12/10/2018	16:06
5	CCV1	1	Water			20.1	2.00	12/10/2018	17:02
6	J6317-01	1	Water			16.4	7.01	12/10/2018	17:04
7	J6334-01	1	Water			9.1	7.46	12/10/2018	17:06
8	J6334-04	1	Water			10.2	7.50	12/10/2018	17:08
9	J6334-07	1	Water			11.2	9.39	12/10/2018	17:10
10	J6334-10	1	Water			11.0	6.25	12/10/2018	17:12
11	J6334-10DUP	1	Water			11.1	6.26	12/10/2018	17:14
12	CCV2	1	Water			20.2	12.00	12/10/2018	17:16

WORKLIST(Hardcopy Internal Chain)

LB 99718

WorkList Name : pH121018

WorkList ID : 120130

Date : 12/10/2018 9:47:53 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2018	Water	J6317-01	pH	Cool 4 deg C	TULL02		MH-718	12/07/2018	9040C
12/10/2018	Water	J6334-01	pH	Cool 4 deg C	PSEG01	L31	B1-121018	12/10/2018	9040C
12/10/2018	Water	J6334-04	pH	Cool 4 deg C	PSEG01	L31	B2-121018	12/10/2018	9040C
12/10/2018	Water	J6334-07	pH	Cool 4 deg C	PSEG01	L31	B3-121018	12/10/2018	9040C
12/10/2018	Water	J6334-10	pH	Cool 4 deg C	PSEG01	L31	GW-121018	12/10/2018	9040C

Date/Time 12/10/18 4:35

Received by: SP

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

12/10/18 5:20

CP

SP