



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99719

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	12/10/2018	16:00
2	CAL2	1	Water			20.1	7.00	12/10/2018	16:02
3	CAL3	1	Water			20.2	10.04	12/10/2018	16:04
4	ICV	1	Water			20.3	7.01	12/10/2018	16:06
5	CCV1	1	Water			20.1	2.00	12/10/2018	17:02
6	J6334-02	1	Water			9.1	7.46	12/10/2018	17:06
7	J6334-05	1	Water			10.2	7.50	12/10/2018	17:08
8	J6334-08	1	Water			11.2	9.39	12/10/2018	17:10
9	J6334-11	1	Water			11.0	6.25	12/10/2018	17:12
10	J6334-11DUP	1	Water			11.1	6.26	12/10/2018	17:14
11	CCV2	1	Water			20.2	12.00	12/10/2018	17:16

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY121018-1

WorkList ID : 120153

Date : 12/10/2018 3:44:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2018	Water	J6334-02	Corrosivity	Cool 4 deg C	PSEG01	L31	B1-121018	12/10/2018	9040C
12/10/2018	Water	J6334-05	Corrosivity	Cool 4 deg C	PSEG01	L31	B2-121018	12/10/2018	9040C
12/10/2018	Water	J6334-08	Corrosivity	Cool 4 deg C	PSEG01	L31	B3-121018	12/10/2018	9040C
12/10/2018	Water	J6334-11	Corrosivity	Cool 4 deg C	PSEG01	L31	GW-121018	12/10/2018	9040C

Date/Time 12/10/18 4:35
 Received by: SP
 Relinquished by: CP

Date/Time 12/10/18 5:20
 Received by: CP
 Relinquished by: SP