



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99720

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	12/11/2018	09:50
2	J6334-02		1	9	>100.0	>212.0	>212.0	12/11/2018	10:20
3	J6334-02DUP		1	9	>100.0	>212.0	>212.0	12/11/2018	10:45
4	J6334-05		1	7	>100.0	>212.0	>212.0	12/11/2018	11:25
5	J6334-08		1	6	>100.0	>212.0	>212.0	12/11/2018	11:50
6	J6334-11		1	7	>100.0	>212.0	>212.0	12/11/2018	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99720

WorkList Name : FP121118

WorkList ID : 120186

Date : 12/11/2018 9:18:34 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Water	J6334-02	Flash Point	Cool 4 deg C	PSEG01	L31	B1-121018	12/10/2018	1010A
12/13/2018	Water	J6334-05	Flash Point	Cool 4 deg C	PSEG01	L31	B2-121018	12/10/2018	1010A
12/13/2018	Water	J6334-08	Flash Point	Cool 4 deg C	PSEG01	L31	B3-121018	12/10/2018	1010A
12/13/2018	Water	J6334-11	Flash Point	Cool 4 deg C	PSEG01	L31	GW-121018	12/10/2018	1010A

Date/Time 12/11/18 9:20 AM

Received by: KS

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

12/11/18 12:25 PM
CP
KS