

Analytical Summary Report**Analysis Method:** 9045D**Analyst By :** Sweta**Parameter:** Corrosivity**Supervisor Review By :** jignesh**Run Number:** LB99722**Slope :** 99.3**BalanceID:** WC SC-4**pH Meter ID :** WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			19.8	4.01	12/11/2018	10:00
2	CAL2	1	Water			19.7	7.00	12/11/2018	10:02
3	CAL3	1	Water			20.0	10.03	12/11/2018	10:04
4	ICV	1	Water			19.8	7.01	12/11/2018	10:06
5	CCV1	1	Water			19.9	2.00	12/11/2018	10:08
6	J6193-02	1	Solid	20.02	20	17.1	10.40	12/11/2018	11:47
7	J6193-02DUP	1	Solid	20.03	20	17.2	10.41	12/11/2018	10:12
8	J6193-04	1	Solid	20.02	20	16.8	10.62	12/11/2018	10:14
9	J6193-06	1	Solid	20.02	20	16.8	10.07	12/11/2018	10:16
10	J6193-08	1	Solid	20.03	20	16.8	9.80	12/11/2018	10:18
11	J6193-10	1	Solid	20.02	20	16.8	10.02	12/11/2018	10:20
12	J6195-06	1	Solid	20.04	20	18.3	8.27	12/11/2018	10:22
13	J6195-08	1	Solid	20.02	20	18.3	8.03	12/11/2018	10:24
14	J6322-16	1	Solid	20.02	20	17.5	8.33	12/11/2018	10:26
15	J6322-18	1	Solid	20.03	20	17.2	8.15	12/11/2018	10:28
16	CCV2	1	Water			19.8	12.00	12/11/2018	10:30
17	J6322-20	1	Solid	20.03	20	17.2	8.63	12/11/2018	10:32
18	J6322-22	1	Solid	20.02	20	17.3	7.48	12/11/2018	10:34
19	J6322-22DUP	1	Solid	20.03	20	17.4	7.49	12/11/2018	10:36
20	J6322-24	1	Solid	20.04	20	17.1	6.95	12/11/2018	10:38

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			20.0	2.00	12/11/2018	10:40

WORKLIST(Hardcopy Internal Chain)

1899722

WorkList Name : corrosivity121118

WorkList ID : 120189

Date : 12/11/2018 9:45:03 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2018	Solid	J6193-02	Corrosivity	Cool 4 deg C	PSEG05	L31	CL-01-121018-A	12/10/2018	9045D
12/10/2018	Solid	J6193-04	Corrosivity	Cool 4 deg C	PSEG05	L31	CL-01-121018-B	12/10/2018	9045D
12/10/2018	Solid	J6193-06	Corrosivity	Cool 4 deg C	PSEG05	L31	CL-01-121018-C	12/10/2018	9045D
12/10/2018	Solid	J6193-08	Corrosivity	Cool 4 deg C	PSEG05	L31	CL-01-121018-D	12/10/2018	9045D
12/10/2018	Solid	J6193-10	Corrosivity	Cool 4 deg C	PSEG05	L31	CL-01-121018-E	12/10/2018	9045D
12/10/2018	Solid	J6195-06	Corrosivity	Cool 4 deg C	PSEG05		OR-02-121018-A	12/10/2018	9045D
12/10/2018	Solid	J6195-08	Corrosivity	Cool 4 deg C	PSEG05		OR-02-121018-B	12/10/2018	9045D
12/07/2018	Solid	J6322-16	Corrosivity	Cool 4 deg C	DIST01	M12	SUP-2A-C-1-120718	12/07/2018	9045D
12/07/2018	Solid	J6322-18	Corrosivity	Cool 4 deg C	DIST01	M12	SUP-2A-C-2-120718	12/07/2018	9045D
12/07/2018	Solid	J6322-20	Corrosivity	Cool 4 deg C	DIST01	M12	SUP-5A-C-1-120718	12/07/2018	9045D
12/07/2018	Solid	J6322-22	Corrosivity	Cool 4 deg C	DIST01	M12	SUP-6-C-2-120718	12/07/2018	9045D
12/06/2018	Solid	J6322-24	Corrosivity	Cool 4 deg C	DIST01	M12	SUP-6-C-1-120618	12/06/2018	9045D

Date/Time 12/11/18 9:50
 Received by: SP
 Relinquished by: SP

Date/Time 12/11/18 10:50
 Received by: SP
 Relinquished by: SP