



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99723

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			19.8	4.01	12/11/2018	10:00
2	CAL2	1	Water			19.4	7.00	12/11/2018	10:02
3	CAL3	1	Water			20.0	10.03	12/11/2018	10:04
4	ICV	1	Water			19.8	7.01	12/11/2018	10:06
5	CCV1	1	Water			20.2	2.00	12/11/2018	11:32
6	J6340-01	1	Solid	20.02	20	16.4	7.01	12/11/2018	11:34
7	J6340-01DUP	1	Solid	20.04	20	16.5	7.02	12/11/2018	11:36
8	J6340-02	1	Solid	20.02	20	16.4	6.44	12/11/2018	11:38
9	J6340-03	1	Solid	20.02	20	16.4	6.05	12/11/2018	11:40
10	J6340-04	1	Solid	20.03	20	16.3	5.82	12/11/2018	11:42
11	J6340-05	1	Solid	20.02	20	16.3	7.16	12/11/2018	11:44
12	J6340-06	1	Solid	20.02	20	16.4	6.40	12/11/2018	11:46
13	J6340-07	1	Solid	20.03	20	16.4	5.25	12/11/2018	11:48
14	J6340-09	1	Solid	20.02	20	16.1	6.03	12/11/2018	11:50
15	J6340-10	1	Solid	20.03	20	16.2	6.18	12/11/2018	11:52
16	CCV2	1	Water			20.1	12.00	12/11/2018	11:54
17	J6340-11	1	Solid	20.04	20	16.3	7.05	12/11/2018	11:56
18	J6340-11DUP	1	Solid	20.03	20	16.4	7.06	12/11/2018	11:58
19	CCV3	1	Water			19.8	2.00	12/11/2018	12:00

WORKLIST(Hardcopy Internal Chain)

LB99723

WorkList Name : pH121118

WorkList ID : 120188

Date : 12/11/2018 9:44:18 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2018	Solid	J6340-01	pH	Cool 4 deg C	BEMS01	D61	SB-02A	12/10/2018	9045D
12/10/2018	Solid	J6340-02	pH	Cool 4 deg C	BEMS01	D61	SB-02B	12/10/2018	9045D
12/10/2018	Solid	J6340-03	pH	Cool 4 deg C	BEMS01	D61	SB-03A	12/10/2018	9045D
12/10/2018	Solid	J6340-04	pH	Cool 4 deg C	BEMS01	D61	SB-03B	12/10/2018	9045D
12/10/2018	Solid	J6340-05	pH	Cool 4 deg C	BEMS01	D61	SB-04A	12/10/2018	9045D
12/10/2018	Solid	J6340-06	pH	Cool 4 deg C	BEMS01	D61	SB-04B	12/10/2018	9045D
12/10/2018	Solid	J6340-07	pH	Cool 4 deg C	BEMS01	D61	SB-05A	12/10/2018	9045D
12/10/2018	Solid	J6340-09	pH	Cool 4 deg C	BEMS01	D61	SB-05B	12/10/2018	9045D
12/10/2018	Solid	J6340-10	pH	Cool 4 deg C	BEMS01	D61	SB-06A	12/10/2018	9045D
12/10/2018	Solid	J6340-11	pH	Cool 4 deg C	BEMS01	D61	SB-10A	12/10/2018	9045D

Date/Time 12/10/18 9:50

Received by: SP

Relinquished by: SP

Date/Time

Received by:

Relinquished by:

12/11/18 12:05

CP

SP