

LB 99740


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56343



A= 0.0000e+000

B= 1.7724e-004

C= 3.5679e-003

Rho= 0.9999058

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	69	0.000	69				
0.2	0.200	0.229	0.029	1271	0.0 %	1271				
2.5	2.500	2.400	-0.100	13520	0.0 %	13520				
5.0	5.000	5.065	0.065	28555	0.0 %	28555				
7.5	7.500	7.501	0.001	42303	0.0 %	42303				
10.0	10.000	9.990	-0.010	56343	0.0 %	56343				

LB99740 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Dec 2018 17:20:51

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	69	-	0.00007471B	PPB	11 Dec 2018 17:22:02	S	Std
0.2 - 1	50.2	1271	-	0.20007471B	PPB	11 Dec 2018 17:24:21	S	Std
2.5 - 1	52.5	13520	-	2.50007471B	PPB	11 Dec 2018 17:26:39	S	Std
5.0 - 1	55	28555	-	5.00007471B	PPB	11 Dec 2018 17:28:56	S	Std
7.5 - 1	57.5	42303	-	7.50007471B	PPB	11 Dec 2018 17:31:14	S	Std
10.0 - 1	58	56343	-	10.00007471B	PPB	11 Dec 2018 17:33:32	S	Std
ICV17 - 1	ICV17	22144	3.9283	-7471B	PPB	11 Dec 2018 17:36:18	U	SMPL
ICB17 - 1	ICB17	-25	-0.0009	-7471B	PPB	11 Dec 2018 17:38:34	U	SMPL
CCV43 - 1	CCV43	28631	5.0781	-7471B	PPB	11 Dec 2018 17:40:50	U	SMPL
CCB43 - 1	CCB43	-38	-0.0032	-7471B	PPB	11 Dec 2018 17:43:06	U	SMPL
CRA - 1	CRA	1236	0.2226	-7471B	PPB	11 Dec 2018 17:45:23	U	SMPL
HighStd - 1	HighStd	55238	9.7938	-7471B	PPB	11 Dec 2018 17:47:39	U	SMPL
ChkStd - 1	ChkStd	38952	6.9073	-7471B	PPB	11 Dec 2018 17:49:55	U	SMPL
PB115528BL - 1	PBS	-77	-0.0101	-7471B	PPB	11 Dec 2018 17:52:14	U	SMPL
PB115528BS - 1	LCSS	19049	3.3798	-7471B	PPB	11 Dec 2018 17:54:33	U	SMPL
J6196-01 - 1	EO-03-120718-A	4500	0.8011	-7471B	PPB	11 Dec 2018 17:56:50	U	SMPL
J6196-03 - 1	EO-03-120718-B	3916	0.6976	-7471B	PPB	11 Dec 2018 17:59:07	U	SMPL
J6196-05 - 1	EO-03-120718-C	3626	0.6462	-7471B	PPB	11 Dec 2018 18:01:25	U	SMPL
J6197-01 - 1	PL-02-120718	26023	4.6158	-7471B	PPB	11 Dec 2018 18:03:42	U	SMPL
J6290-01 - 1	SOIL	855	0.1551	-7471B	PPB	11 Dec 2018 18:05:58	U	SMPL
CCV44 - 1	CCV44	28724	5.0945	-7471B	PPB	11 Dec 2018 18:08:15	U	SMPL
CCB44 - 1	CCB44	-34	-0.0025	-7471B	PPB	11 Dec 2018 18:10:31	U	SMPL
J6302-01 - 1	K-120718-RO	1452	0.2609	-7471B	PPB	11 Dec 2018 18:12:48	U	SMPL
J6306-01 - 1	VB-16204	11848	2.1035	-7471B	PPB	11 Dec 2018 18:15:05	U	SMPL
J6316-01 - 1	BORING-SAMPLE-SIGNAL-TOWER-B	13635	2.4202	-7471B	PPB	11 Dec 2018 18:17:21	U	SMPL
J6316-05 - 1	BORING-SAMPLE-CBH-591	25455	4.5152	-7471B	PPB	11 Dec 2018 18:19:37	U	SMPL
J6316-09 - 1	AVE-X-B-BALL-FIELD-S-1	16071	2.8520	-7471B	PPB	11 Dec 2018 18:21:54	U	SMPL
J6316-13 - 1	AVE-X-B-BALL-FIELD-S-2	10724	1.9043	-7471B	PPB	11 Dec 2018 18:24:12	U	SMPL
J6316-17 - 1	AVE-X-B-BALL-FIELD-S-3	11711	2.0792	-7471B	PPB	11 Dec 2018 18:26:30	U	SMPL
J6316-21 - 1	BORING-SAMPLE-STEAL-WELL-YARD	26384	4.6798	-7471B	PPB	11 Dec 2018 18:28:48	U	SMPL
J6331-01 - 1	TP-1	24721	4.3851	-7471B	PPB	11 Dec 2018 18:31:06	U	SMPL
J6331-07 - 1	TP-2	1023	0.1849	-7471B	PPB	11 Dec 2018 18:33:23	U	SMPL
CCV45 - 1	CCV45	28841	5.1153	-7471B	PPB	11 Dec 2018 18:35:40	U	SMPL
CCB45 - 1	CCB45	-44	-0.0042	-7471B	PPB	11 Dec 2018 18:37:57	U	SMPL
J6331-13 - 1	TP-3	715	0.1303	-7471B	PPB	11 Dec 2018 18:40:14	U	SMPL
J6331-19 - 1	TP-4	297041	52.6505	-7471B	PPB	11 Dec 2018 18:42:31	U	SMPL
J6331-25 - 1	TP-5	1036	0.1872	-7471B	PPB	11 Dec 2018 18:44:47	U	SMPL
J6331-31 - 1	TP-6	487	0.0899	-7471B	PPB	11 Dec 2018 18:47:18	U	SMPL
J6331-31DUP - 1	TP-6DUP	507	0.0934	-7471B	PPB	11 Dec 2018 18:49:36	U	SMPL
J6331-31MS - 1	TP-6MS	20606	3.6557	-7471B	PPB	11 Dec 2018 18:51:53	U	SMPL
J6331-31MSD - 1	TP-6MSD	21497	3.8137	-7471B	PPB	11 Dec 2018 18:54:10	U	SMPL
J6331-37 - 1	TP-7	1711	0.3068	-7471B	PPB	11 Dec 2018 18:56:28	U	SMPL
PB115529BL - 1	PBS	-1	0.0034	-7471B	PPB	11 Dec 2018 18:58:46	U	SMPL
PB115529BS - 1	LCSS	18992	3.3697	-7471B	PPB	11 Dec 2018 19:01:03	U	SMPL
CCV46 - 1	CCV46	28464	5.0485	-7471B	PPB	11 Dec 2018 19:03:19	U	SMPL
CCB46 - 1	CCB46	-47	-0.0048	-7471B	PPB	11 Dec 2018 19:05:36	U	SMPL
J6193-01 - 1	CL-01-121018-A	5624	1.0004	-7471B	PPB	11 Dec 2018 19:07:53	U	SMPL
J6193-01DUP - 1	CL-01-121018-ADUP	5417	0.9637	-7471B	PPB	11 Dec 2018 19:10:10	U	SMPL
J6193-01MS - 1	CL-01-121018-AMS	30032	5.3264	-7471B	PPB	11 Dec 2018 19:12:27	U	SMPL
J6193-01MSD - 1	CL-01-121018-AMSD	29230	5.1842	-7471B	PPB	11 Dec 2018 19:14:44	U	SMPL
J6193-03 - 1	CL-01-121018-B	5050	0.8986	-7471B	PPB	11 Dec 2018 19:17:01	U	SMPL
J6193-05 - 1	CL-01-121018-C	6275	1.1157	-7471B	PPB	11 Dec 2018 19:19:18	U	SMPL
J6193-07 - 1	CL-01-121018-D	5739	1.0207	-7471B	PPB	11 Dec 2018 19:21:35	U	SMPL
J6193-09 - 1	CL-01-121018-E	7801	1.3862	-7471B	PPB	11 Dec 2018 19:23:52	U	SMPL
J6195-05 - 1	OR-02-121018-A	3024	0.5395	-7471B	PPB	11 Dec 2018 19:26:09	U	SMPL
J6195-07 - 1	OR-02-121018-B	2189	0.3915	-7471B	PPB	11 Dec 2018 19:28:26	U	SMPL
CCV47 - 1	CCV47	28487	5.0525	-7471B	PPB	11 Dec 2018 19:30:43	U	SMPL
CCB47 - 1	CCB47	-41	-0.0037	-7471B	PPB	11 Dec 2018 19:32:60	U	SMPL
J6214-31 - 1	SB-2-(15-17)	24621	4.3673	-7471B	PPB	11 Dec 2018 19:34:77	U	SMPL
J6303-03 - 1	NR38-C1-110718	10294	1.8281	-7471B	PPB	11 Dec 2018 19:36:94	U	SMPL
J6322-15 - 1	SUP-2A-C-1-120718	780	0.1418	-7471B	PPB	11 Dec 2018 19:39:11	U	SMPL
J6322-17 - 1	SUP-2A-C-2-120718	2290	0.4094	-7471B	PPB	11 Dec 2018 19:41:28	U	SMPL
J6322-19 - 1	SUP-5A-C-1-120718	706	0.1287	-7471B	PPB	11 Dec 2018 19:43:45	U	SMPL
J6322-21 - 1	SUP-6-C-2-120718	1695	0.3040	-7471B	PPB	11 Dec 2018 19:46:02	U	SMPL
J6322-23 - 1	SUP-6-C-1-120618	1154	0.2081	-7471B	PPB	11 Dec 2018 19:48:19	U	SMPL
J6327-01 - 1	STOCK-PILE-(5PT-COMP)	2233	0.3993	-7471B	PPB	11 Dec 2018 19:50:36	U	SMPL
PB115530BL - 1	PBS	1	0.0037	-7471B	PPB	11 Dec 2018 19:52:53	U	SMPL
PB115530BS - 1	LCSS	18653	3.3096	-7471B	PPB	11 Dec 2018 19:55:10	U	SMPL
CCV48 - 1	CCV48	28274	5.0148	-7471B	PPB	11 Dec 2018 19:57:27	U	SMPL
CCB48 - 1	CCB48	-40	-0.0035	-7471B	PPB	11 Dec 2018 19:59:44	U	SMPL
J6340-01 - 1	SB-02A	8705	1.5464	-7471B	PPB	11 Dec 2018 20:02:01	U	SMPL
J6340-01DUP - 1	SB-02ADUP	9265	1.6457	-7471B	PPB	11 Dec 2018 20:04:18	U	SMPL
J6340-01MS - 1	SB-02AMS	31341	5.5584	-7471B	PPB	11 Dec 2018 20:06:35	U	SMPL
J6340-01MSD - 1	SB-02AMSD	31456	5.5788	-7471B	PPB	11 Dec 2018 20:08:52	U	SMPL
J6340-02 - 1	SB-02B	1744	0.3127	-7471B	PPB	11 Dec 2018 20:11:09	U	SMPL
J6340-03 - 1	SB-03A	2117	0.3788	-7471B	PPB	11 Dec 2018 20:13:26	U	SMPL
J6340-04 - 1	SB-03B	294	0.0557	-7471B	PPB	11 Dec 2018 20:15:43	U	SMPL
J6340-05 - 1	SB-04A	3340	0.5955	-7471B	PPB	11 Dec 2018 20:18:00	U	SMPL
J6340-06 - 1	SB-04B	3852	0.6863	-7471B	PPB	11 Dec 2018 20:20:17	U	SMPL
J6340-07 - 1	SB-05A	31340	5.5582	-7471B	PPB	11 Dec 2018 20:22:34	U	SMPL
CCV49 - 1	CCV49	28126	4.9886	-7471B	PPB	11 Dec 2018 20:24:51	U	SMPL
CCB49 - 1	CCB49	-90	-0.0124	-7471B	PPB	11 Dec 2018 20:27:08	U	SMPL
J6340-09 - 1	SB-05B	50675	8.9851	-7471B	PPB	11 Dec 2018 20:29:25	U	SMPL
J6340-10 - 1	SB-06A	97	0.0208	-7471B	PPB	11 Dec 2018 20:31:42	U	SMPL
J6340-11 - 1	SB-10A	2132	0.3814	-7471B	PPB	11 Dec 2018 20:33:59	U	SMPL
J6331-31LX5 - 1		250	0.0479	-7471B	PPB	11 Dec 2018 20:36:16	U	SMPL
J6331-31A - 1		22202	3.9386	-7471B	PPB	11 Dec 2018 20:38:33	U	SMPL
J6193-01LX5 - 1		1009	0.1824	-7471B	PPB	11 Dec 2018 20:40:50	U	SMPL
J6193-01A - 1		27552	4.8868	-7471B	PPB	11 Dec 2018 20:43:07	U	SMPL
J6340-01LX5 - 1		2289	0.4093	-7471B	PPB	11 Dec 2018 20:45:24	U	SMPL
J6340-01A - 1		34398	6.1002	-7471B	PPB	11 Dec 2018 20:47:41	U	SMPL
J6331-19DLX10 - 1		31633	5.6101	-7471B	PPB	11 Dec 2018 20:49:58	U	SMPL
CCV50 - 1	CCV50	28487	5.0525	-7471B	PPB	11 Dec 2018 20:52:15	U	SMPL
				-7471B	PPB	11 Dec 2018 21:00:19	U	SMPL

Method: 7471B

Operator: Admin

LB99740 INSTRUMENT ID : CV1

Date of Analysis: 11 Dec 2018 17:20:51

Sample ID

Extended ID

CCB50 - 1

CCB50

μ Abs.

Conc.

Std CondMethod

Units

-35

-0.0026

7471B

PPB

Date

Type

Type

11 Dec 2018 21:02:37

U

SMPL