



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/12/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 12/11/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 12/11/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB99741

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J6189-01	HR-05-121118-A	1	1.12	9.6	8.9	91.7
J6189-03	HR-05-121118-B	2	1.18	9.79	9.18	92.9
J6189-05	HR-05-121118-C	3	1.14	9.7	9.08	92.8
J6194-05	OK-03-121118	4	1.16	9.54	8.87	92.0
J6342-01	RICH-REC-ROAD-STONE-8	5	1.17	9.6	9.05	93.5
J6342-02	RICH-REC-ROAD-STONE-8	6	1.15	9.53	9.03	94.0
J6344-01	RBR-200027	7	1.00	2.00	2.00	100.0
J6344-02	RBR-200027	8	1.00	2.00	2.00	100.0
J6344-04	RBR-200031	9	1.00	2.00	2.00	100.0
J6344-05	RBR-200031	10	1.00	2.00	2.00	100.0
J6346-01	DW-5	11	1.11	9.51	8.84	92.0
J6346-02	DW-6	12	1.16	9.64	8.73	89.3
J6346-03	DW-8	13	1.17	9.61	9.22	95.4
J6346-04	DW-10	14	1.18	9.81	9.41	95.4
J6346-05	DW-11	15	1.12	9.7	9.12	93.2
J6349-01	VNJ-242	16	1.00	2.00	2.00	100.0
J6352-01	S32-0633	17	1.16	9.6	8.49	86.8
J6354-01	SOIL	18	1.15	9.92	9.04	90.0
J6354-02	SOIL	19	1.15	9.92	9.04	90.0
J6356-01	SOIL-CUTTINGS	20	1.16	9.57	8.59	88.3
J6357-01	SB-01A	21	1.15	9.67	8.79	89.7
J6357-02	J6357-01MS	22	1.15	9.67	8.79	89.7
J6357-03	J6357-01MSD	23	1.15	9.67	8.79	89.7
J6357-04	SB-07A	24	1.16	9.66	8.45	85.8
J6357-05	SB-07B	25	1.14	9.86	9.4	94.7
J6357-06	SB-08A	26	1.14	9.61	9.24	95.6
J6357-07	SB-08B	27	1.15	9.75	8.56	86.2
J6357-08	SB-09A	28	1.17	9.8	8.02	79.4
J6357-09	SB-09B	29	1.14	9.97	8.31	81.2
J6357-10	SB-01B	30	1.14	9.71	8.67	87.9

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$