

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/11/2018

TEMP1 IN: 103 12/11/2018 14:00 TEMP OUT: 104 12/11/2018 15:05

TEMP2 IN: 105 12/11/2018 16:00 TEMP OUT: 105 12/12/2018 07:10

TEMP3 IN: 105 12/12/2018 08:30 TEMP OUT: 105 12/12/2018 10:00

TEMP4 IN: 105 12/12/2018 11:45 TEMP OUT: 105 12/12/2018 13:20

Run Number: LB99745

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99745BL	LB99745BL	1.4127	1.4127	100	1.4127	1.4127	1.4127	0.0000	0
2	LB99745BS	LB99745BS	1.3865	1.3865	100	1.4470	1.4470	1.4470	0.0605	605
3	J6341-03	121018-JETT-F-D-B	1.4127	1.4127	250	1.4265	1.4265	1.4265	0.0138	55.2
4	J6341-03DUP	121018-JETT-F-D-BDUP	1.4246	1.4246	250	1.4385	1.4385	1.4385	0.0139	55.6
5	J6341-04	121018-JETT-F-D-S	1.3950	1.3950	300	1.4051	1.4051	1.4051	0.0101	33.7
6	J6341-05	121018-JETT-F-U-B	1.3971	1.3971	300	1.4105	1.4105	1.4105	0.0134	44.7
7	J6341-06	121018-JETT-F-U-S	1.3987	1.3987	400	1.4100	1.4100	1.4100	0.0113	28.3
8	J6350-01	SV-27781L	1.4028	1.4028	250	1.4139	1.4139	1.4139	0.0111	44.4
9	J6351-02	KY038PS027-181210	1.3970	1.3970	900	1.3996	1.3996	1.3996	0.0026	2.9

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS121218

WorkList ID : 120220

Date : 12/12/2018 7:30:49 AM

2399745

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/17/2018	Water	J6341-03 C	TSS	Cool 4 deg C	HDR103	D31	121018-JETT-F-D-B	12/10/2018	SM2540 D
12/17/2018	Water	J6341-04 C	TSS	Cool 4 deg C	HDR103	D31	121018-JETT-F-D-S	12/10/2018	SM2540 D
12/17/2018	Water	J6341-05 C	TSS	Cool 4 deg C	HDR103	D31	121018-JETT-F-U-B	12/10/2018	SM2540 D
12/17/2018	Water	J6341-06 C	TSS	Cool 4 deg C	HDR103	D31	121018-JETT-F-U-S	12/10/2018	SM2540 D
12/14/2018	Water	J6350-01 F	TSS	Cool 4 deg C	PSEG01	D11	SV-27781L	12/11/2018	SM2540 D
12/17/2018	Water	J6351-02 E	TSS	Cool 4 deg C	SHAW22	D31	KY038PS027-181210	12/10/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

12/12/18, 07:00
AS
AS

Date/Time

Received by:

Relinquished by:

12/12/18, 09:00
AS
AS