



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99751

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.0	4.01	12/11/2018	17:20
2	CAL2	1	Water			20.1	7.00	12/11/2018	17:22
3	CAL3	1	Water			20.2	10.04	12/11/2018	17:24
4	ICV	1	Water			19.9	7.01	12/11/2018	17:26
5	CCV1	1	Water			19.9	2.00	12/11/2018	17:28
6	J6350-01	1	Water			14.3	8.66	12/11/2018	17:30
7	J6350-01DUP	1	Water			14.4	8.67	12/11/2018	17:32
8	CCV2	1	Water			20.0	12.00	12/11/2018	17:36

WORKLIST(Hardcopy Internal Chain)

B 99751

WorkList Name : pH121118-1

WorkList ID : 120217

Date : 12/11/2018 4:25:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/11/2018	Water	J6350-01	pH	Cool 4 deg C	PSEG01	D11	SV-27781L	12/11/2018	9040C

Date/Time 12/11/18 5:10
 Received by: SP
 Relinquished by: CP

Date/Time 12/11/18 5:40
 Received by: CP
 Relinquished by: SP