

Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99753

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.0	4.01	12/11/2018	17:20
2	CAL2	1	Water			20.1	7.00	12/11/2018	17:22
3	CAL3	1	Water			20.2	10.04	12/11/2018	17:24
4	ICV	1	Water			19.9	7.01	12/11/2018	17:26
5	CCV1	1	Water			19.9	2.00	12/11/2018	17:40
6	J6189-02	1	Solid	20.02	20	18.4	8.56	12/11/2018	17:42
7	J6189-02DUP	1	Solid	20.03	20	18.5	8.57	12/11/2018	17:44
8	J6189-04	1	Solid	20.03	20	18.3	8.80	12/11/2018	17:46
9	J6189-06	1	Solid	20.04	20	18.1	8.74	12/11/2018	17:48
10	J6194-06	1	Solid	20.02	20	17.9	8.45	12/11/2018	17:50
11	J6340-08	1	Solid	20.03	20	16.1	6.43	12/11/2018	17:52
12	J6344-03	1	Solid	20.02	20	18.1	10.34	12/11/2018	17:54
13	J6344-06	1	Solid	20.02	20	17.9	10.97	12/11/2018	17:56
14	J6349-01	1	Solid	20.02	20	18.0	11.18	12/11/2018	17:58
15	J6354-02	1	Solid	20.03	20	17.9	9.48	12/11/2018	18:00
16	CCV2	1	Water			20.2	12.00	12/11/2018	18:02
17	J6356-01	1	Solid	20.03	20	18.0	7.79	12/11/2018	18:04
18	J6357-13	1	Solid	20.03	20	18.0	7.23	12/11/2018	18:06
19	J6357-13DUP	1	Solid	20.04	20	18.1	7.24	12/11/2018	18:08
20	CCV3	1	Water			20.1	2.00	12/12/2018	18:10

WORKLIST(Hardcopy Internal Chain)

LB 9 9753

WorkList Name : CORROSIONITY121118-1

WorkList ID : 120215

Date : 12/11/2018 4:25:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/11/2018	Solid	J6189-02	Corrosivity	Cool 4 deg C	PSEG05	D11	HR-05-121118-A	12/11/2018	9045D
12/11/2018	Solid	J6189-04	Corrosivity	Cool 4 deg C	PSEG05	D11	HR-05-121118-B	12/11/2018	9045D
12/11/2018	Solid	J6189-06	Corrosivity	Cool 4 deg C	PSEG05	D11	HR-05-121118-C	12/11/2018	9045D
12/11/2018	Solid	J6194-06	Corrosivity	Cool 4 deg C	PSEG05		OK-03-121118	12/11/2018	9045D
12/10/2018	Solid	J6340-08	Corrosivity	Cool 4 deg C	BEMS01		SB-05A	12/10/2018	9045D
12/11/2018	Solid	J6344-03	Corrosivity	Cool 4 deg C	PSEG01	D11	RBR-200027	12/11/2018	9045D
12/11/2018	Solid	J6344-06	Corrosivity	Cool 4 deg C	PSEG01	D11	RBR-200031	12/11/2018	9045D
12/11/2018	Solid	J6349-01	Corrosivity	Cool 4 deg C	PSEG01	D11	VNJ-242	12/11/2018	9045D
12/11/2018	Solid	J6354-02	Corrosivity	Cool 4 deg C	PSEG01	D12	SOIL	12/11/2018	9045D
12/11/2018	Solid	J6356-01	Corrosivity	Cool 4 deg C	SEAC01		SOIL-CUTTINGS	12/11/2018	9045D
12/11/2018	Solid	J6357-13	Corrosivity	Cool 4 deg C	BEMS01		SB-09A	12/11/2018	9045D

Date/Time 12/11/18 5:10
 Received by: SP
 Relinquished by: CP

Date/Time 12/11/18 6:14
 Received by: CP
 Relinquished by: SP