



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99755

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	12/12/2018	09:32
2	J6350-02		1	8	>100.0	>212.0	>212.0	12/12/2018	09:52
3	J6350-02DUP		1	8	>100.0	>212.0	>212.0	12/12/2018	10:20
4	J6356-02		1	9	>100.0	>212.0	>212.0	12/12/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB99755

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp121218

WorkList ID : 120245

Date : 12/12/2018 8:58:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/14/2018	Water	J6350-02 B	Flash Point	Cool 4 deg C	PSEG01	D11	SV-27781L	12/11/2018	1010A
12/24/2018	Water	J6356-02 D	Flash Point	Cool 4 deg C	SEAC01	D11	GROUND-WATER	12/11/2018	1010A

Date/Time 12/12/18 9:00 AM

Received by: KS

Relinquished by: CP

Date/Time 12/12/18 10:55 AM

Received by: CP

Relinquished by: KS