



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99758

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.0	4.01	12/11/2018	17:20
2	CAL2	1	Water			20.1	7.00	12/11/2018	17:22
3	CAL3	1	Water			20.2	10.04	12/11/2018	17:24
4	ICV	1	Water			19.9	7.01	12/11/2018	17:26
5	CCV1	1	Water			20.1	2.00	12/11/2018	18:10
6	J6357-01	1	Solid	20.02	20	17.9	9.13	12/11/2018	18:12
7	J6357-01DUP	1	Solid	20.03	20	18.0	9.14	12/11/2018	18:14
8	J6357-04	1	Solid	20.03	20	17.9	9.41	12/11/2018	18:16
9	J6357-05	1	Solid	20.02	20	17.9	8.23	12/11/2018	18:18
10	J6357-06	1	Solid	20.02	20	17.8	8.60	12/11/2018	18:20
11	J6357-07	1	Solid	20.04	20	17.9	7.87	12/11/2018	18:22
12	J6357-08	1	Solid	20.02	20	17.7	7.22	12/11/2018	18:24
13	J6357-09	1	Solid	20.02	20	17.3	6.87	12/11/2018	18:26
14	J6357-10	1	Solid	20.03	20	17.2	7.07	12/11/2018	18:28
15	CCV2	1	Water			20.2	12.00	12/11/2018	18:30

WORKLIST(Hardcopy Internal Chain)

LB 09758

WorkList Name : pH121118-2

WorkList ID : 120219

Date : 12/11/2018 4:54:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/11/2018	Solid	J6357-01	pH	Cool 4 deg C	BEMS01	D21	SB-01A	12/11/2018	9045D
12/11/2018	Solid	J6357-04	pH	Cool 4 deg C	BEMS01	D21	SB-07A	12/11/2018	9045D
12/11/2018	Solid	J6357-05	pH	Cool 4 deg C	BEMS01	D21	SB-07B	12/11/2018	9045D
12/11/2018	Solid	J6357-06	pH	Cool 4 deg C	BEMS01	D21	SB-08A	12/11/2018	9045D
12/11/2018	Solid	J6357-07	pH	Cool 4 deg C	BEMS01	D21	SB-08B	12/11/2018	9045D
12/11/2018	Solid	J6357-08	pH	Cool 4 deg C	BEMS01	D21	SB-09A	12/11/2018	9045D
12/11/2018	Solid	J6357-09	pH	Cool 4 deg C	BEMS01	D21	SB-09B	12/11/2018	9045D
12/11/2018	Solid	J6357-10	pH	Cool 4 deg C	BEMS01	D21	SB-01B	12/11/2018	9045D

Date/Time 12/11/18 5:10
 Received by: SP
 Relinquished by: CP

Date/Time 12/11/18 6:35
 Received by: CP
 Relinquished by: CSP