


7471B

LB 99770
INSTRUMENT ID: CY1

Linear

 μ Abs.:

52233



A= 0.0000e+000

B= 1.9055e-004

C= -2.8551e-002

Rho= 0.9998781

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	-4	0.000	-4				
0.2	0.200	0.183	-0.017	1108	0.0 %	1108				
2.5	2.500	2.482	-0.018	13174	0.0 %	13174				
5.0	5.000	5.108	0.108	26955	0.0 %	26955				
7.5	7.500	7.533	0.033	39684	0.0 %	39684				
10.0	10.000	9.924	-0.076	52233	0.0 %	52233				

LB99770 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Dec 2018 15:51:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-4	-	0.0000	7471B	PPB	12 Dec 2018 15:55:20	S	Std
0.2 - 1	50.2	1108	-	0.2000	7471B	PPB	12 Dec 2018 15:57:38	S	Std
2.5 - 1	52.5	13174	-	2.5000	7471B	PPB	12 Dec 2018 15:59:56	S	Std
5.0 - 1	55	26955	-	5.0000	7471B	PPB	12 Dec 2018 16:02:14	S	Std
7.5 - 1	57.5	39684	-	7.5000	7471B	PPB	12 Dec 2018 16:04:32	S	Std
10.0 - 1	510	52233	-	10.0000	7471B	PPB	12 Dec 2018 16:09:24	S	Std
ICV19 - 1	ICV19	21344	4.0385	-7471B	PPB	12 Dec 2018 16:12:07	U	SMPL	
ICB19 - 1	ICB19	-124	-0.0522	-7471B	PPB	12 Dec 2018 16:14:22	U	SMPL	
CCV61 - 1	CCV61	26606	5.0411	-7471B	PPB	12 Dec 2018 16:16:39	U	SMPL	
CCB61 - 1	CCB61	-134	-0.0541	-7471B	PPB	12 Dec 2018 16:18:54	U	SMPL	
CRA - 1	CRA	1067	0.1748	-7471B	PPB	12 Dec 2018 16:21:11	U	SMPL	
HighStd - 1	HighStd	52676	10.0087	-7471B	PPB	12 Dec 2018 16:23:27	U	SMPL	
ChkStd - 1	ChkStd	36165	6.8626	-7471B	PPB	12 Dec 2018 16:25:44	U	SMPL	
PB115579BL - 1	PBS	-246	-0.0754	-7471B	PPB	12 Dec 2018 16:28:02	U	SMPL	
PB115579BS - 1	LCSS	19148	3.6200	-7471B	PPB	12 Dec 2018 16:30:21	U	SMPL	
J6189-01 - 1	HR-05-121118-A	10339	1.9415	-7471B	PPB	12 Dec 2018 16:32:38	U	SMPL	
J6189-03 - 1	HR-05-121118-B	5217	0.9655	-7471B	PPB	12 Dec 2018 16:34:56	U	SMPL	
J6189-05 - 1	HR-05-121118-C	3718	0.6799	-7471B	PPB	12 Dec 2018 16:37:13	U	SMPL	
J6194-05 - 1	OK-03-121118	1177	0.1957	-7471B	PPB	12 Dec 2018 16:39:31	U	SMPL	
J6342-01 - 1	RICH-REC-ROAD-STONE-8	232	0.0157	-7471B	PPB	12 Dec 2018 16:41:47	U	SMPL	
CCV62 - 1	CCV62	26391	5.0002	-7471B	PPB	12 Dec 2018 16:44:03	U	SMPL	
CCB62 - 1	CCB62	-116	-0.0507	-7471B	PPB	12 Dec 2018 16:46:19	U	SMPL	
J6344-01 - 1	RBR-200027	45993	8.7353	-7471B	PPB	12 Dec 2018 16:48:36	U	SMPL	
J6344-04 - 1	RBR-200031	128	-0.0042	-7471B	PPB	12 Dec 2018 16:50:52	U	SMPL	
J6354-01 - 1	SOIL	7385	1.3786	-7471B	PPB	12 Dec 2018 16:53:11	U	SMPL	
J6357-01 - 1	SB-01A	1122	0.1852	-7471B	PPB	12 Dec 2018 16:55:28	U	SMPL	
J6357-01DUP - 1	SB-01ADUP	1140	0.1887	-7471B	PPB	12 Dec 2018 17:00:32	U	SMPL	
J6357-02 - 1	J6357-01MS	20754	3.9261	-7471B	PPB	12 Dec 2018 17:02:49	U	SMPL	
J6357-03 - 1	J6357-01MSD	20276	3.8350	-7471B	PPB	12 Dec 2018 17:05:06	U	SMPL	
J6357-04 - 1	SB-07A	24742	4.6860	-7471B	PPB	12 Dec 2018 17:07:24	U	SMPL	
J6357-05 - 1	SB-07B	10401	1.9533	-7471B	PPB	12 Dec 2018 17:09:42	U	SMPL	
J6357-06 - 1	SB-08A	3402	0.6197	-7471B	PPB	12 Dec 2018 17:12:01	U	SMPL	
CCV63 - 1	CCV63	26461	5.0135	-7471B	PPB	12 Dec 2018 17:14:17	U	SMPL	
CCB63 - 1	CCB63	-137	-0.0547	-7471B	PPB	12 Dec 2018 17:16:33	U	SMPL	
J6357-07 - 1	SB-08B	323	0.0330	-7471B	PPB	12 Dec 2018 17:18:51	U	SMPL	
J6357-08 - 1	SB-09A	8514	1.5938	-7471B	PPB	12 Dec 2018 17:21:08	U	SMPL	
J6357-09 - 1	SB-09B	1111	0.1831	-7471B	PPB	12 Dec 2018 17:23:24	U	SMPL	
J6357-10 - 1	SB-01B	220	0.0134	-7471B	PPB	12 Dec 2018 17:25:41	U	SMPL	
J6357-01LX5 - 1		213	0.0120	-7471B	PPB	12 Dec 2018 17:27:58	U	SMPL	
J6357-01A - 1		23552	4.4592	-7471B	PPB	12 Dec 2018 17:30:15	U	SMPL	
CCV64 - 1	CCV64	26473	5.0158	-7471B	PPB	12 Dec 2018 17:32:33	U	SMPL	
CCB64 - 1	CCB64	-131	-0.0535	-7471B	PPB	12 Dec 2018 17:34:51	U	SMPL	