

LB 99775


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

58503



A= 0.0000e+000

B= 1.7210e-004

C= -7.8937e-002

Rho= 0.9998866

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	704	0.000	704				
0.2	0.200	0.235	0.035	1825	0.0 %	1825				
2.5	2.500	2.401	-0.099	14412	0.0 %	14412				
5.0	5.000	4.965	-0.035	29307	0.0 %	29307				
7.5	7.500	7.567	0.067	44425	0.0 %	44425				
10.0	10.000	9.990	-0.010	58503	0.0 %	58503				

LB99775 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Dec 2018 07:27:00

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	704	-	0.0000	7471B	PPB	13 Dec 2018 07:56:42	S	Std
0.2 - 1	SD.2	1825	-	0.2000	7471B	PPB	13 Dec 2018 07:59:00	S	Std
2.5 - 1	SD.5	14412	-	2.5000	7471B	PPB	13 Dec 2018 08:01:18	S	Std
5.0 - 1	SD	29307	-	5.0000	7471B	PPB	13 Dec 2018 08:03:35	S	Std
7.5 - 1	SD.5	44425	-	7.5000	7471B	PPB	13 Dec 2018 08:05:53	S	Std
10.0 - 1	SD	58503	-	10.0000	7471B	PPB	13 Dec 2018 08:08:11	S	Std
ICV20 - 1	ICV20	23033	3.8851	-	7471B	PPB	13 Dec 2018 08:11:00	U	SMPL
ICB20 - 1	ICB20	597	0.0238	-	7471B	PPB	13 Dec 2018 08:13:16	U	SMPL
CCV65 - 1	CCV65	27728	4.6931	-	7471B	PPB	13 Dec 2018 08:15:32	U	SMPL
CCB65 - 1	CCB65	605	0.0252	-	7471B	PPB	13 Dec 2018 08:17:48	U	SMPL
CRA - 1	CRA	1788	0.2288	-	7471B	PPB	13 Dec 2018 08:20:06	U	SMPL
HighStd - 1	HighStd	57033	9.7366	-	7471B	PPB	13 Dec 2018 08:22:22	U	SMPL
ChkStd - 1	ChkStd	39248	6.6758	-	7471B	PPB	13 Dec 2018 08:24:38	U	SMPL
PB115621BS - 1	LCSS	19667	3.3058	-	7471B	PPB	13 Dec 2018 08:31:50	U	SMPL
J6199-05 - 1	SU-04-121218-A	5067	0.7931	-	7471B	PPB	13 Dec 2018 08:34:08	U	SMPL
J6199-05DUP - 1	SU-04-121218-ADUP	4947	0.7725	-	7471B	PPB	13 Dec 2018 08:36:25	U	SMPL
J6199-05MS - 1	SU-04-121218-AMS	23986	4.0491	-	7471B	PPB	13 Dec 2018 08:38:43	U	SMPL
J6199-05MSD - 1	SU-04-121218-AMSD	24752	4.1810	-	7471B	PPB	13 Dec 2018 08:41:01	U	SMPL
J6199-07 - 1	SU-04-121218-B	3060	0.4477	-	7471B	PPB	13 Dec 2018 08:43:18	U	SMPL
CCV66 - 1	CCV66	27997	4.7394	-	7471B	PPB	13 Dec 2018 08:45:35	U	SMPL
CCB66 - 1	CCB66	606	0.0254	-	7471B	PPB	13 Dec 2018 08:47:51	U	SMPL
J6319-01 - 1	055_BT-1-1	14097	2.3472	-	7471B	PPB	13 Dec 2018 08:50:08	U	SMPL
J6319-02 - 1	056_BT-1-2	16467	2.7551	-	7471B	PPB	13 Dec 2018 08:52:25	U	SMPL
J6319-03 - 1	057_BT-2-1	18792	3.1552	-	7471B	PPB	13 Dec 2018 08:54:41	U	SMPL
J6319-05 - 1	059_BT-3-1	51175	8.7284	-	7471B	PPB	13 Dec 2018 08:56:58	U	SMPL
J6319-07 - 1	061_SW-N-1	30691	5.2031	-	7471B	PPB	13 Dec 2018 08:59:15	U	SMPL
J6319-09 - 1	063_SW-E-1	18502	3.1053	-	7471B	PPB	13 Dec 2018 09:01:34	U	SMPL
J6319-10 - 1	064_SW-E-2	47185	8.0417	-	7471B	PPB	13 Dec 2018 09:03:53	U	SMPL
J6319-14 - 1	068_DUP-2	29800	5.0497	-	7471B	PPB	13 Dec 2018 09:06:11	U	SMPL
J6329-01 - 1	071-PIPE-1	1740	0.2205	-	7471B	PPB	13 Dec 2018 09:08:31	U	SMPL
J6329-02 - 1	072-DUP-1	1394	0.1610	-	7471B	PPB	13 Dec 2018 09:10:49	U	SMPL
CCV67 - 1	CCV67	28011	4.7418	-	7471B	PPB	13 Dec 2018 09:13:05	U	SMPL
CCB67 - 1	CCB67	635	0.0303	-	7471B	PPB	13 Dec 2018 09:15:22	U	SMPL
J6347-03 - 1	P001-SD001-0612-01	1588	0.1944	-	7471B	PPB	13 Dec 2018 09:17:40	U	SMPL
J6347-04 - 1	P001-SD002-0006-01	4785	0.7446	-	7471B	PPB	13 Dec 2018 09:19:56	U	SMPL
J6347-05 - 1	P001-SD002-0612-01	9547	1.5641	-	7471B	PPB	13 Dec 2018 09:22:12	U	SMPL
J6347-20 - 1	P001-SD001-0006-01	1258	0.1376	-	7471B	PPB	13 Dec 2018 09:24:29	U	SMPL
J6362-01 - 1	P-1	1550	0.1878	-	7471B	PPB	13 Dec 2018 09:26:46	U	SMPL
J6362-02 - 1	P-2	1354	0.1541	-	7471B	PPB	13 Dec 2018 09:29:03	U	SMPL
J6362-03 - 1	P-3	1458	0.1720	-	7471B	PPB	13 Dec 2018 09:31:20	U	SMPL
J6362-04 - 1	P-4	1189	0.1257	-	7471B	PPB	13 Dec 2018 09:33:39	U	SMPL
PB115622BL - 1	PBS	267	-0.0330	-	7471B	PPB	13 Dec 2018 09:35:56	U	SMPL
PB115622BS - 1	LCSS	20199	3.3974	-	7471B	PPB	13 Dec 2018 09:38:13	U	SMPL
CCV68 - 1	CCV68	28117	4.7601	-	7471B	PPB	13 Dec 2018 09:40:30	U	SMPL
CCB68 - 1	CCB68	499	0.0069	-	7471B	PPB	13 Dec 2018 09:42:47	U	SMPL
J6353-01 - 1	P001-SS001-0002-01	4289	0.6592	-	7471B	PPB	13 Dec 2018 09:45:04	U	SMPL
J6353-01DUP - 1	P001-SS001-0002-01DUP	4786	0.7447	-	7471B	PPB	13 Dec 2018 09:47:21	U	SMPL
J6353-02 - 1	J6353-01MS	24349	4.1116	-	7471B	PPB	13 Dec 2018 09:52:31	U	SMPL
J6353-03 - 1	J6353-01MSD	25560	4.3200	-	7471B	PPB	13 Dec 2018 09:54:47	U	SMPL
J6353-04 - 1	P001-SS001-0002-02	3401	0.5064	-	7471B	PPB	13 Dec 2018 09:57:05	U	SMPL
J6353-05 - 1	P001-SS002-0002-01	9759	1.6006	-	7471B	PPB	13 Dec 2018 09:59:23	U	SMPL
J6353-06 - 1	P001-SS003-0002-01	6004	0.9544	-	7471B	PPB	13 Dec 2018 10:01:40	U	SMPL
J6353-07 - 1	P001-SS004-0002-01	2955	0.4296	-	7471B	PPB	13 Dec 2018 10:03:57	U	SMPL
J6353-08 - 1	P001-SS005-0002-01	607	0.0255	-	7471B	PPB	13 Dec 2018 10:06:16	U	SMPL
J6353-09 - 1	P001-SS006-0002-01	4263	0.6547	-	7471B	PPB	13 Dec 2018 10:08:33	U	SMPL
CCV69 - 1	CCV69	28262	4.7850	-	7471B	PPB	13 Dec 2018 10:10:49	U	SMPL
CCB69 - 1	CCB69	499	0.0069	-	7471B	PPB	13 Dec 2018 10:13:06	U	SMPL
J6353-10 - 1	P001-SS007-0002-01	7973	1.2932	-	7471B	PPB	13 Dec 2018 10:15:24	U	SMPL
J6361-01 - 1	TN4803-121218	1514	0.1816	-	7471B	PPB	13 Dec 2018 10:17:40	U	SMPL
J6368-01 - 1	SV26391L	562	0.0178	-	7471B	PPB	13 Dec 2018 10:19:57	U	SMPL
J6371-02 - 1	EB07-1-2	13816	2.2988	-	7471B	PPB	13 Dec 2018 10:22:13	U	SMPL
J6371-03 - 1	EB07-11-12	993	0.0920	-	7471B	PPB	13 Dec 2018 10:24:30	U	SMPL
J6371-04 - 1	EB04-0-2	11530	1.9054	-	7471B	PPB	13 Dec 2018 10:26:47	U	SMPL
J6371-05 - 1	EB04-14-16	921	0.0796	-	7471B	PPB	13 Dec 2018 10:29:04	U	SMPL
J6372-01 - 1	FP-SOIL-1	19383	3.2569	-	7471B	PPB	13 Dec 2018 10:31:22	U	SMPL
J6373-01 - 1	TP-A-FRAME	16240	2.7160	-	7471B	PPB	13 Dec 2018 10:33:40	U	SMPL
J6373-05 - 1	TP-T1	8119	1.3184	-	7471B	PPB	13 Dec 2018 10:35:57	U	SMPL
CCV70 - 1	CCV70	28438	4.8153	-	7471B	PPB	13 Dec 2018 10:38:14	U	SMPL
CCB70 - 1	CCB70	495	0.0063	-	7471B	PPB	13 Dec 2018 10:40:31	U	SMPL
J6374-01 - 1	FK-GF-02-029-A	6031	0.9590	-	7471B	PPB	13 Dec 2018 10:42:49	U	SMPL
J6374-03 - 1	FK-GF-02-029-B	5740	0.9089	-	7471B	PPB	13 Dec 2018 10:45:06	U	SMPL
J6199-05LX5 - 1		1446	0.1699	-	7471B	PPB	13 Dec 2018 10:47:23	U	SMPL
J6199-05A - 1		27854	4.7148	-	7471B	PPB	13 Dec 2018 10:49:40	U	SMPL
J6353-01LX5 - 1		1228	0.1324	-	7471B	PPB	13 Dec 2018 10:51:57	U	SMPL
J6353-01A - 1		26310	4.4491	-	7471B	PPB	13 Dec 2018 10:54:15	U	SMPL
PB115621BL - 1	PBS	180	-0.0480	-	7471B	PPB	13 Dec 2018 10:56:32	U	SMPL
CCV71 - 1	CCV71	27791	4.7040	-	7471B	PPB	13 Dec 2018 10:58:51	U	SMPL
CCB71 - 1	CCB71	526	0.0116	-	7471B	PPB	13 Dec 2018 11:01:09	U	SMPL