

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/13/2018

TEMP1 IN: 103 12/13/2018 07:00 TEMP OUT: 104 12/13/2018 08:05

TEMP2 IN: 104 12/13/2018 08:45 TEMP OUT: 104 12/13/2018 10:00

TEMP3 IN: 105 12/13/2018 11:30 TEMP OUT: 105 12/13/2018 13:05

TEMP4 IN: 105 12/13/2018 13:40 TEMP OUT: 105 12/13/2018 15:20

Run Number: LB99776

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99776BL	LB99776BL	1.4000	1.4000	100	1.4000	1.4000	1.4000	0.0000	0
2	LB99776BS	LB99776BS	1.3956	1.3956	100	1.4457	1.4457	1.4457	0.0501	501
3	J6363-01	121118-JETT-F-D-B	1.3757	1.3757	300	1.3846	1.3846	1.3846	0.0089	29.7
4	J6363-02	121118-JETT-F-D-S	1.4272	1.4272	400	1.4374	1.4374	1.4374	0.0102	25.5
5	J6363-03	121118-JETT-F-U-B	1.3825	1.3825	400	1.3953	1.3953	1.3953	0.0128	32
6	J6363-04	121118-JETT-F-U-S	1.3778	1.3778	500	1.3905	1.3905	1.3905	0.0127	25.4
7	J6373-10	OILY-WATER	1.3895	1.3895	50	1.4146	1.4146	1.4146	0.0251	502
8	J6373-10DUP	OILY-WATERDUP	1.4220	1.4220	50	1.4480	1.4480	1.4480	0.0260	520
9	J6375-01	COMP	1.3863	1.3863	50	1.3908	1.3908	1.3908	0.0045	90
10	J6376-07	TEST-PIT-1-WATER	1.4263	1.4263	30	1.4544	1.4544	1.4544	0.0281	936.7
11	J6380-01	RAV-006	1.4052	1.4052	500	1.4062	1.4062	1.4062	0.0010	2
12	J6380-02	RAV-01D	1.3684	1.3684	500	1.3689	1.3689	1.3689	0.0005	1
13	J6380-04	RAV-RAINEY-BEFORE	1.3803	1.3803	500	1.3808	1.3808	1.3808	0.0005	1
14	J6380-05	RAV-RAINEY-AFTER	1.3883	1.3883	500	1.3902	1.3902	1.3902	0.0019	3.8



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/13/2018

TEMP1 IN: 103 12/13/2018 07:00 TEMP OUT: 104 12/13/2018 08:05

TEMP2 IN: 104 12/13/2018 08:45 TEMP OUT: 104 12/13/2018 10:00

TEMP3 IN: 105 12/13/2018 11:30 TEMP OUT: 105 12/13/2018 13:05

TEMP4 IN: 105 12/13/2018 13:40 TEMP OUT: 105 12/13/2018 15:20

Run Number: LB99776

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS121318

WorkList ID : 120284

Date : 12/13/2018 7:54:11 AM

12/13/18 7:54:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2018	Water	J6363-01	C TSS	Cool 4 deg C	HDR103	D41	121118-JETT-F-D-B	12/11/2018	SM2540 D
12/18/2018	Water	J6363-02	C TSS	Cool 4 deg C	HDR103	D41	121118-JETT-F-D-S	12/11/2018	SM2540 D
12/18/2018	Water	J6363-03	C TSS	Cool 4 deg C	HDR103	D41	121118-JETT-F-U-B	12/11/2018	SM2540 D
12/18/2018	Water	J6363-04	C TSS	Cool 4 deg C	HDR103	D41	121118-JETT-F-U-S	12/11/2018	SM2540 D
12/18/2018	Water	J6373-10	C TSS	Cool 4 deg C	PSEG01	D12	OILY-WATER	12/12/2018	SM2540 D
12/19/2018	Water	J6375-01	- TSS	Cool 4 deg C	ARAM01	D32	COMP	12/12/2018	SM2540 D
12/17/2018	Water	J6376-07	F TSS	Cool 4 deg C	PSEG01	D12	TEST-PIT-1-WATER	12/12/2018	SM2540 D
12/19/2018	Water	J6380-01	A TSS	Cool 4 deg C	FOST03	D31	RAV-006	12/12/2018	SM2540 D
12/19/2018	Water	J6380-02	A TSS	Cool 4 deg C	FOST03	D31	RAV-01D	12/12/2018	SM2540 D
12/19/2018	Water	J6380-04	A TSS	Cool 4 deg C	FOST03	D31	RAV-RAINEY-BEFORE	12/12/2018	SM2540 D
12/19/2018	Water	J6380-05	A TSS	Cool 4 deg C	FOST03	D31	RAV-RAINEY-AFTER	12/12/2018	SM2540 D

SP

Date/Time 12/13/18, 08:00
 Received by: AS
 Relinquished by: CS

Date/Time 12/13/18 12:00
 Received by: AS
 Relinquished by: AS