



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99782

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/12/2018	16:30
2	CAL2	1	Water			20.1	7.00	12/12/2018	16:32
3	CAL3	1	Water			20.2	10.04	12/12/2018	16:34
4	ICV	1	Water			20.3	7.01	12/12/2018	16:36
5	CCV1	1	Water			20.0	2.00	12/12/2018	16:46
6	J6373-10	1	Water			15.0	6.62	12/12/2018	16:48
7	J6373-10DUP	1	Water			15.1	6.63	12/12/2018	16:50
8	J6376-07	1	Water			16.4	6.47	12/12/2018	16:52
9	CCV2	1	Water			20.1	12.00	12/12/2018	16:54

WORKLIST(Hardcopy Internal Chain)

UB99782

WorkList Name : pH121218-1

WorkList ID : 120280

Date : 12/12/2018 4:10:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/12/2018	Water	J6373-10	pH	Cool 4 deg C	PSEG01	D12	OILY-WATER	12/12/2018	9040C
12/12/2018	Water	J6376-07	pH	Cool 4 deg C	PSEG01	D12	TEST-PIT-1-WATER	12/12/2018	9040C

Date/Time 12/12/18 4:20

Received by: SP

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

12/12/18 5:00

CP SP