



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99783

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2390         |
| buffer solution pH 7 yellow    | W2358         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2323         |
| buffer solution pH 7 yellow    | W2322         |
| pH 2.00 Buffer                 | W2289         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |             |             | 20.1             | 4.01        | 12/12/2018 | 16:30     |
| 2   | CAL2        | 1  | Water  |             |             | 20.1             | 7.00        | 12/12/2018 | 16:32     |
| 3   | CAL3        | 1  | Water  |             |             | 20.2             | 10.04       | 12/12/2018 | 16:34     |
| 4   | ICV         | 1  | Water  |             |             | 20.3             | 7.01        | 12/12/2018 | 16:36     |
| 5   | CCV1        | 1  | Water  |             |             | 20.1             | 2.00        | 12/12/2018 | 16:38     |
| 6   | J6368-01    | 1  | Solid  | 20.02       | 20          | 19.0             | 8.72        | 12/12/2018 | 16:40     |
| 7   | J6368-01DUP | 1  | Solid  | 20.04       | 20          | 19.1             | 8.73        | 12/12/2018 | 16:42     |
| 8   | CCV2        | 1  | Water  |             |             | 20.1             | 12.00       | 12/12/2018 | 16:44     |

# WORKLIST(Hardcopy Internal Chain)

LB 99783

WorkList Name : pH121218

WorkList ID : 120279

Date : 12/12/2018 4:10:25 PM

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 12/12/2018 | Solid  | J6368-01 | pH   | Cool 4 deg C | PSEG01   | D12              | SV26391L        | 12/12/2018   | 9045D  |

Date/Time 12/12/18 4:20  
 Received by: SP  
 Relinquished by: CP

Date/Time 12/12/18 5:00  
 Received by: CP  
 Relinquished by: SP