



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99784

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/12/2018	16:30
2	CAL2	1	Water			20.1	7.00	12/12/2018	16:32
3	CAL3	1	Water			20.2	10.04	12/12/2018	16:34
4	ICV	1	Water			20.3	7.01	12/12/2018	16:36
5	CCV1	1	Water			20.2	2.00	12/12/2018	17:30
6	J6199-06	1	Solid	20.02	20	19.6	8.88	12/12/2018	17:32
7	J6199-06DUP	1	Solid	20.03	20	19.7	8.89	12/12/2018	17:34
8	J6199-08	1	Solid	20.03	20	19.0	8.77	12/12/2018	17:36
9	J6361-03	1	Solid	20.02	20	20.2	5.90	12/12/2018	17:38
10	J6372-03	1	Solid	20.02	20	19.5	11.43	12/12/2018	17:40
11	J6373-04	1	Solid	20.04	20	19.3	7.74	12/12/2018	17:42
12	J6373-08	1	Solid	20.02	20	19.2	7.89	12/12/2018	17:44
13	J6376-03	1	Solid	20.03	20	19.3	8.07	12/12/2018	17:46
14	J6376-06	1	Solid	20.02	20	19.3	7.61	12/12/2018	17:48
15	CCV2	1	Water			20.1	12.00	12/12/2018	17:50

WORKLIST(Hardcopy Internal Chain)

1399784

WorkList Name : corrosivity121218

WorkList ID : 120277

Date : 12/12/2018 4:10:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/12/2018	Solid	J6199-06	Corrosivity	Cool 4 deg C	PSEG05		SU-04-121218-A	12/12/2018	9045D
12/12/2018	Solid	J6199-08	Corrosivity	Cool 4 deg C	PSEG05		SU-04-121218-B	12/12/2018	9045D
12/12/2018	Solid	J6361-03	Corrosivity	Cool 4 deg C	PSEG01	D12	TN4803-121218	12/12/2018	9045D
12/12/2018	Solid	J6372-03	Corrosivity	Cool 4 deg C	PSEG01	D12	FP-SOIL-1	12/12/2018	9045D
12/12/2018	Solid	J6373-04	Corrosivity	Cool 4 deg C	PSEG01	D12	TP-A-FRAME	12/12/2018	9045D
12/12/2018	Solid	J6373-08	Corrosivity	Cool 4 deg C	PSEG01	D12	TP-T1	12/12/2018	9045D
12/12/2018	Solid	J6376-03	Corrosivity	Cool 4 deg C	PSEG01	D12	TEST-PIT-1	12/12/2018	9045D
12/12/2018	Solid	J6376-06	Corrosivity	Cool 4 deg C	PSEG01	D12	TEST-PIT-2	12/12/2018	9045D

Date/Time 12/12/18 5:20
 Received by: SP
 Relinquished by: CP

5:55
 Date/Time 12/12/18
 Received by: CP
 Relinquished by: SP