

LB 99785


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36831



A= 0.0000e+000

B= 2.6872e-004

C= -5.6548e-003

Rho= 0.9997700

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	26	0.000	26				
0.2	0.200	0.203	0.003	777	0.0 %	777				
2.5	2.500	2.425	-0.075	9044	0.0 %	9044				
5.0	5.000	5.044	0.044	18792	0.0 %	18792				
7.5	7.500	7.635	0.135	28433	0.0 %	28433				
10.0	10.000	9.892	-0.108	36831	0.0 %	36831				

LB99785 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Dec 2018 11:27:47

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	26	-	0.00007470A	PPB	13 Dec 2018 11:29:15	S	Std
0.2 - 1	50.2	777	-	0.20007470A	PPB	13 Dec 2018 11:31:24	S	Std
2.5 - 1	52.5	9044	-	2.50007470A	PPB	13 Dec 2018 11:33:33	S	Std
5.0 - 1	55	18792	-	5.00007470A	PPB	13 Dec 2018 11:35:41	S	Std
7.5 - 1	57.5	28433	-	7.50007470A	PPB	13 Dec 2018 11:37:49	S	Std
10.0 - 1	57.5	36831	-	10.00007470A	PPB	13 Dec 2018 11:42:23	S	Std
ICV21 - 1	ICV21	14028	3.7640	-7470A	PPB	13 Dec 2018 11:44:56	U	SMPL
ICB21 - 1	ICB21	-73	-0.0253	-7470A	PPB	13 Dec 2018 11:47:02	U	SMPL
CCV72 - 1	CCV72	18472	4.9582	-7470A	PPB	13 Dec 2018 11:49:09	U	SMPL
CCB72 - 1	CCB72	-71	-0.0247	-7470A	PPB	13 Dec 2018 11:51:15	U	SMPL
CRA - 1	CRA	777	0.2031	-7470A	PPB	13 Dec 2018 11:53:23	U	SMPL
HighStd - 1	HighStd	36727	9.8637	-7470A	PPB	13 Dec 2018 11:55:29	U	SMPL
ChkStd - 1	ChkStd	26009	6.9836	-7470A	PPB	13 Dec 2018 11:57:37	U	SMPL
PB115609BL - 1	PBW	-80	-0.0272	-7470A	PPB	13 Dec 2018 11:59:45	U	SMPL
PB115609BS - 1	LCSW	13982	3.7516	-7470A	PPB	13 Dec 2018 12:01:54	U	SMPL
J6189-02 - 1	HR-05-121118-A	-49	-0.0188	-7470A	PPB	13 Dec 2018 12:04:02	U	SMPL
J6189-02DUP - 1	HR-05-121118-ADUP	14	-0.0019	-7470A	PPB	13 Dec 2018 12:06:10	U	SMPL
J6189-02MS - 1	HR-05-121118-AMS	12947	3.4735	-7470A	PPB	13 Dec 2018 12:08:19	U	SMPL
J6189-02MSD - 1	HR-05-121118-AMSD	12913	3.4644	-7470A	PPB	13 Dec 2018 12:10:28	U	SMPL
J6189-04 - 1	HR-05-121118-B	12	-0.0024	-7470A	PPB	13 Dec 2018 12:12:35	U	SMPL
CCV73 - 1	CCV73	18432	4.9475	-7470A	PPB	13 Dec 2018 12:14:42	U	SMPL
CCB73 - 1	CCB73	-74	-0.0255	-7470A	PPB	13 Dec 2018 12:16:49	U	SMPL
J6189-06 - 1	HR-05-121118-C	61	0.0107	-7470A	PPB	13 Dec 2018 12:18:58	U	SMPL
J6193-02 - 1	CL-01-121018-A	79	0.0156	-7470A	PPB	13 Dec 2018 12:21:05	U	SMPL
J6193-04 - 1	CL-01-121018-B	72	0.0137	-7470A	PPB	13 Dec 2018 12:23:12	U	SMPL
J6193-06 - 1	CL-01-121018-C	70	0.0132	-7470A	PPB	13 Dec 2018 12:25:20	U	SMPL
J6193-08 - 1	CL-01-121018-D	30	0.0024	-7470A	PPB	13 Dec 2018 12:27:28	U	SMPL
J6193-10 - 1	CL-01-121018-E	30	0.0024	-7470A	PPB	13 Dec 2018 12:29:36	U	SMPL
J6194-06 - 1	OK-03-121118	30	0.0024	-7470A	PPB	13 Dec 2018 12:31:44	U	SMPL
J6340-08 - 1	SB-05A	40	0.0051	-7470A	PPB	13 Dec 2018 12:33:53	U	SMPL
J6344-03 - 1	RBR-200027	37	0.0043	-7470A	PPB	13 Dec 2018 12:36:01	U	SMPL
J6344-06 - 1	RBR-200031	46	0.0067	-7470A	PPB	13 Dec 2018 12:38:09	U	SMPL
CCV74 - 1	CCV74	18499	4.9655	-7470A	PPB	13 Dec 2018 12:40:16	U	SMPL
CCB74 - 1	CCB74	-65	-0.0231	-7470A	PPB	13 Dec 2018 12:42:23	U	SMPL
J6349-01 - 1	VNJ-242	30	0.0024	-7470A	PPB	13 Dec 2018 12:44:31	U	SMPL
J6354-02 - 1	SOIL	41	0.0054	-7470A	PPB	13 Dec 2018 12:46:38	U	SMPL
J6356-01 - 1	SOIL-CUTTINGS	38	0.0046	-7470A	PPB	13 Dec 2018 12:48:46	U	SMPL
J6357-13 - 1	SB-09A	55	0.0091	-7470A	PPB	13 Dec 2018 12:50:53	U	SMPL
PB115610BL - 1	PBW	19	-0.0005	-7470A	PPB	13 Dec 2018 12:53:00	U	SMPL
PB115610BS - 1	LCSW	14947	4.0110	-7470A	PPB	13 Dec 2018 12:55:08	U	SMPL
J6343-02 - 1	BP-HN-MW29IR-20181209	-97	-0.0317	-7470A	PPB	13 Dec 2018 12:57:17	U	SMPL
J6343-02DUP - 1	BP-HN-MW29IR-20181209DUP	-46	-0.0180	-7470A	PPB	13 Dec 2018 12:59:25	U	SMPL
J6343-02MS - 1	BP-HN-MW29IR-20181209MS	12747	3.4198	-7470A	PPB	13 Dec 2018 13:01:34	U	SMPL
J6343-02MSD - 1	BP-HN-MW29IR-20181209MSD	12472	3.3459	-7470A	PPB	13 Dec 2018 13:03:41	U	SMPL
CCV75 - 1	CCV75	18161	4.8746	-7470A	PPB	13 Dec 2018 13:05:49	U	SMPL
CCB75 - 1	CCB75	-47	-0.0183	-7470A	PPB	13 Dec 2018 13:07:56	U	SMPL
J6343-03 - 1	BP-HN-MW29IR-20181209	-60	-0.0218	-7470A	PPB	13 Dec 2018 13:10:04	U	SMPL
J6343-04 - 1	BP-HN-MW29D-20181209	-45	-0.0177	-7470A	PPB	13 Dec 2018 13:12:11	U	SMPL
J6343-05 - 1	BP-HN-MW29D-20181209	-42	-0.0169	-7470A	PPB	13 Dec 2018 13:14:18	U	SMPL
J6343-06 - 1	BPS1-TT-MW311S-20181209	-28	-0.0132	-7470A	PPB	13 Dec 2018 13:16:25	U	SMPL
J6343-07 - 1	BPS1-TT-MW311S-20181209	-55	-0.0204	-7470A	PPB	13 Dec 2018 13:18:33	U	SMPL
J6343-08 - 1	BPS1-TT-MW311I-20181209	-59	-0.0215	-7470A	PPB	13 Dec 2018 13:20:41	U	SMPL
J6343-09 - 1	BPS1-TT-MW311I-20181209	-45	-0.0177	-7470A	PPB	13 Dec 2018 13:22:49	U	SMPL
J6343-10 - 1	BP-HN-MW24IR-20181209	-38	-0.0159	-7470A	PPB	13 Dec 2018 13:24:58	U	SMPL
J6343-11 - 1	BP-HN-MW24IR-20181209	-27	-0.0129	-7470A	PPB	13 Dec 2018 13:27:06	U	SMPL
J6343-12 - 1	BPSI-TT-MW310S-20181209	-41	-0.0167	-7470A	PPB	13 Dec 2018 13:29:14	U	SMPL
CCV76 - 1	CCV76	18800	5.0463	-7470A	PPB	13 Dec 2018 13:31:21	U	SMPL
CCB76 - 1	CCB76	-60	-0.0218	-7470A	PPB	13 Dec 2018 13:33:29	U	SMPL
J6343-13 - 1	BPSI-TT-MW310S-20181209	-40	-0.0164	-7470A	PPB	13 Dec 2018 13:35:36	U	SMPL
J6343-14 - 1	BP-DUP05-20181209	-47	-0.0183	-7470A	PPB	13 Dec 2018 13:37:43	U	SMPL
J6343-15 - 1	BP-DUP05-20181209	-44	-0.0175	-7470A	PPB	13 Dec 2018 13:39:51	U	SMPL
J6343-16 - 1	BPS1-TT-MW312I-20181210	-48	-0.0186	-7470A	PPB	13 Dec 2018 13:41:58	U	SMPL
J6343-17 - 1	BPS1-TT-MW312I-20181210	-52	-0.0196	-7470A	PPB	13 Dec 2018 13:44:06	U	SMPL
J6343-18 - 1	BPS1-TT-MW309D-20181210	-46	-0.0180	-7470A	PPB	13 Dec 2018 13:46:14	U	SMPL
J6343-19 - 1	BPS1-TT-MW309D-20181210	-53	-0.0199	-7470A	PPB	13 Dec 2018 13:48:22	U	SMPL
J6343-20 - 1	BPS1-TT-MW312S-20181210	-50	-0.0191	-7470A	PPB	13 Dec 2018 13:50:30	U	SMPL
J6343-21 - 1	BPS1-TT-MW312S-20181210	-45	-0.0177	-7470A	PPB	13 Dec 2018 13:52:40	U	SMPL
PB115611BL - 1	PBW	13	-0.0022	-7470A	PPB	13 Dec 2018 13:54:47	U	SMPL
CCV77 - 1	CCV77	18604	4.9937	-7470A	PPB	13 Dec 2018 13:56:54	U	SMPL
CCB77 - 1	CCB77	-41	-0.0167	-7470A	PPB	13 Dec 2018 13:59:02	U	SMPL
PB115611BS - 1	LCSW	14402	3.8645	-7470A	PPB	13 Dec 2018 14:01:10	U	SMPL
J6305-01 - 1	P001-SW001-01	9348	2.5064	-7470A	PPB	13 Dec 2018 14:03:17	U	SMPL
J6305-02 - 1	P001-SW002-01	288	0.0717	-7470A	PPB	13 Dec 2018 14:05:24	U	SMPL
J6305-03 - 1	P001-SW003-01	181	0.0430	-7470A	PPB	13 Dec 2018 14:07:33	U	SMPL
J6305-04 - 1	P001-SW004-01	278	0.0691	-7470A	PPB	13 Dec 2018 14:09:40	U	SMPL
J6305-04DUP - 1	P001-SW004-01DUP	272	0.0674	-7470A	PPB	13 Dec 2018 14:11:48	U	SMPL
J6305-05 - 1	J6305-04MS	13831	3.7111	-7470A	PPB	13 Dec 2018 14:13:56	U	SMPL
J6305-06 - 1	J6305-04MSD	12490	3.3507	-7470A	PPB	13 Dec 2018 14:16:05	U	SMPL
J6305-07 - 1	P001-SW004-02	156	0.0363	-7470A	PPB	13 Dec 2018 14:18:15	U	SMPL
J6305-08 - 1	P001-SW005-01	8	-0.0035	-7470A	PPB	13 Dec 2018 14:20:24	U	SMPL
CCV78 - 1	CCV78	18556	4.9808	-7470A	PPB	13 Dec 2018 14:22:32	U	SMPL
CCB78 - 1	CCB78	-72	-0.0250	-7470A	PPB	13 Dec 2018 14:24:40	U	SMPL
J6350-01 - 1	SV-27781L	-28	-0.0132	-7470A	PPB	13 Dec 2018 14:26:48	U	SMPL
J6356-02 - 1	GROUND-WATER	-40	-0.0164	-7470A	PPB	13 Dec 2018 14:28:56	U	SMPL
J6357-11 - 1	TWP-02	34	0.0035	-7470A	PPB	13 Dec 2018 14:31:04	U	SMPL
J6189-02LX5 - 1		25	0.0011	-7470A	PPB	13 Dec 2018 14:33:12	U	SMPL
J6189-02A - 1		14380	3.8586	-7470A	PPB	13 Dec 2018 14:35:20	U	SMPL
J6343-02LX5 - 1		-40	-0.0164	-7470A	PPB	13 Dec 2018 14:37:29	U	SMPL
J6343-02A - 1		14335	3.8465	-7470A	PPB	13 Dec 2018 14:39:37	U	SMPL
J6305-04LX5 - 1		11	-0.0027	-7470A	PPB	13 Dec 2018 14:41:46	U	SMPL
J6305-04A - 1		15672	4.2058	-7470A	PPB	13 Dec 2018 14:43:54	U	SMPL
PB115609TB - 1	PB115577TB	-139	-0.0430	-7470A	PPB	13 Dec 2018 14:46:02	U	SMPL
CCV79 - 1	CCV79	18231	4.8934	-7470A	PPB	13 Dec 2018 14:48:11	U	SMPL

LB99785 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Dec 2018 11:27:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB79 - 1	CCB79	-66	-0.0234		7470A	PPB	13 Dec 2018 14:50:20	U	SMPL