

2 B99801

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/13/2018 15:36

Test: Phenolics-

mg/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.041	0.0	0.108	
ICB1	0.003	0.0	0.003	
CCV1	1.013	0.0	0.105	
CCB1	0.007	0.0	0.003	
PB115511BL	0.002	0.0	0.003	
PB115511BS	1.058	0.0	0.110	
J6214-55	0.005	0.0	0.003	
J6214-55DUP	0.006	0.0	0.003	
J6318-01	0.004	0.0	0.003	
J6214-55MS	0.979	0.0	0.102	
J6214-55MSD	0.971	0.0	0.101	
CCV2	0.986	0.0	0.103	
CCB2	0.003	0.0	0.003	

N	13
Mean	0.468
SD	0.5214
CV%	111.50

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/13/2018 15:03


Test Phenolics-

Accepted 12/13/2018 15:03

Factor 9.863
Bias 0.003

Coeff. of det. 0.999769

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0537	0.0500	
2	Phenol-2	0.013	0.1018	0.1000	
3	Phenol-2	0.028	0.2550	0.2500	
4	Phenol-2	0.053	0.5020	0.5000	
5	Phenol-2	0.102	0.9776	1.0000	
6	Phenol-2	0.206	2.0099	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 13 15:03:11 2018

Thu Dec 13 15:34:23 2018

✓AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0537	mg/l	12/13/2018 15:03:11	
0.1PPM	A	Phenolics-	P	0.1018	mg/l	12/13/2018 15:03:12	
0.25PPM	A	Phenolics-	P	0.255	mg/l	12/13/2018 15:03:13	
0.50PPM	A	Phenolics-	P	0.502	mg/l	12/13/2018 15:03:14	
1.0PPM	A	Phenolics-	P	0.9776	mg/l	12/13/2018 15:03:15	
2.0PPM	A	Phenolics-	P	2.0099	mg/l	12/13/2018 15:03:16	
ICV1	S	Phenolics-	P	1.0415	mg/l	12/13/2018 15:26:33	
ICB1	S	Phenolics-	P	0.003	mg/l	12/13/2018 15:26:34	
CCV1	S	Phenolics-	P	1.013	mg/l	12/13/2018 15:26:36	
CCB1	S	Phenolics-	P	0.0066	mg/l	12/13/2018 15:26:38	
PB115511BL	S	Phenolics-	P	0.0021	mg/l	12/13/2018 15:26:40	
PB115511BS	S	Phenolics-	P	1.058	mg/l	12/13/2018 15:26:41	
J6214-55	S	Phenolics-	P	0.0054	mg/l	12/13/2018 15:26:42	
J6214-55DUP	S	Phenolics-	P	0.006	mg/l	12/13/2018 15:26:43	
J6318-01	S	Phenolics-	P	0.0042	mg/l	12/13/2018 15:34:16	
J6214-55MS	S	Phenolics-	P	0.9788	mg/l	12/13/2018 15:34:17	
J6214-55MSD	S	Phenolics-	P	0.9713	mg/l	12/13/2018 15:34:18	
CCV2	S	Phenolics-	P	0.986	mg/l	12/13/2018 15:34:20	
CCB2	S	Phenolics-	P	0.0027	mg/l	12/13/2018 15:34:21	