



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99805

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/13/2018	17:00
2	CAL2	1	Water			20.1	7.00	12/13/2018	17:02
3	CAL3	1	Water			20.2	10.04	12/13/2018	17:04
4	ICV	1	Water			20.3	7.01	12/13/2018	17:06
5	CCV1	1	Water			20.1	2.00	12/13/2018	17:08
6	J6390-04	1	Water			10.7	5.05	12/13/2018	17:10
7	J6390-04DUP	1	Water			10.8	5.06	12/13/2018	17:12
8	J6390-05	1	Water			12.7	5.78	12/13/2018	17:14
9	J6390-06	1	Water			15.2	5.77	12/13/2018	17:16
10	J6390-07	1	Water			15.7	6.04	12/13/2018	17:18
11	J6390-08	1	Water			13.4	6.58	12/13/2018	17:20
12	J6390-09	1	Water			14.2	6.76	12/13/2018	17:22
13	J6403-65	1	Water			13.5	7.09	12/13/2018	17:24
14	CCV2	1	Water			20.1	12.00	12/13/2018	17:28

WORKLIST(Hardcopy Internal Chain)

UB99805

WorkList Name : pH121318

WorkList ID : 120331

Date : 12/13/2018 4:25:05 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Water	J6390-04	pH	Cool 4 deg C	PSEG01	D21	30098	12/13/2018	9040C
12/13/2018	Water	J6390-05	pH	Cool 4 deg C	PSEG01	D21	30104	12/13/2018	9040C
12/13/2018	Water	J6390-06	pH	Cool 4 deg C	PSEG01	D21	30099	12/13/2018	9040C
12/13/2018	Water	J6390-07	pH	Cool 4 deg C	PSEG01	D21	30082	12/13/2018	9040C
12/13/2018	Water	J6390-08	pH	Cool 4 deg C	PSEG01	D21	30081	12/13/2018	9040C
12/13/2018	Water	J6390-09	pH	Cool 4 deg C	PSEG01	D21	30080	12/13/2018	9040C
12/13/2018	Water	J6403-65	pH	Cool 4 deg C	PSEG01	D22	TP-3-OILY-WATER	12/13/2018	9040C

Date/Time 12/13/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 12/13/18 5:30
 Received by: CP
 Relinquished by: SP