



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB99806

Analyst By : sweta

Supervisor Review By : jignesh

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/13/2018	17:00
2	CAL2	1	Water			20.1	7.00	12/13/2018	17:02
3	CAL3	1	Water			20.2	10.04	12/13/2018	17:04
4	ICV	1	Water			20.3	7.01	12/13/2018	17:06
5	CCV1	1	Water			20.1	2.00	12/13/2018	17:08
6	J6403-66	1	Water			13.5	7.09	12/13/2018	17:24
7	J6403-66DUP	1	Water			13.6	7.10	12/13/2018	17:26
8	CCV2	1	Water			20.1	12.00	12/13/2018	17:28

WORKLIST(Hardcopy Internal Chain)

U99806

WorkList Name : corrosivity121318-1

WorkList ID : 120330

Date : 12/13/2018 4:24:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Water	J6403-66	Corrosivity	Cool 4 deg C	PSEG01	D22	TP-3-OILY-WATER	12/13/2018	9040C

Date/Time 12/13/18 5:00
 Received by: SP
 Relinquished by: SP

Date/Time 12/13/18 5:30
 Received by: SP
 Relinquished by: SP