



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99807

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/13/2018	17:00
2	CAL2	1	Water			20.1	7.00	12/13/2018	17:02
3	CAL3	1	Water			20.2	10.04	12/13/2018	17:04
4	ICV	1	Water			20.3	7.01	12/13/2018	17:06
5	CCV1	1	Water			20.3	2.00	12/13/2018	17:32
6	J6390-01	1	Solid	20.02	20	20.1	6.85	12/13/2018	17:34
7	J6390-01DUP	1	Solid	20.03	20	20.2	6.86	12/13/2018	17:36
8	CCV2	1	Water			20.1	12.00	12/13/2018	17:38

WORKLIST(Hardcopy Internal Chain)

VB 99807

WorkList Name : pH121318-1

WorkList ID : 120332

Date : 12/13/2018 4:25:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Solid	J6390-01	pH	Cool 4 deg C	PSEG01	D21	20380	12/13/2018	9045D

Date/Time 12/13/18 5:00
 Received by: SP
 Relinquished by: [Signature]

Date/Time 12/13/18 5:40
 Received by: [Signature]
 Relinquished by: [Signature]