



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB99808

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/13/2018	17:00
2	CAL2	1	Water			20.1	7.00	12/13/2018	17:02
3	CAL3	1	Water			20.2	10.04	12/13/2018	17:04
4	ICV	1	Water			20.3	7.01	12/13/2018	17:06
5	CCV1	1	Water			20.2	2.00	12/13/2018	17:40
6	J6403-04	1	Solid	20.02	20	19.5	7.85	12/13/2018	17:42
7	J6403-04DUP	1	Solid	20.03	20	19.6	7.86	12/13/2018	17:44
8	J6403-08	1	Solid	20.03	20	19.4	7.84	12/13/2018	17:46
9	J6403-12	1	Solid	20.02	20	19.8	7.67	12/13/2018	17:48
10	J6403-16	1	Solid	20.03	20	20.0	8.39	12/13/2018	17:50
11	J6403-20	1	Solid	20.03	20	20.0	8.04	12/13/2018	17:52
12	J6403-24	1	Solid	20.02	20	19.8	7.81	12/13/2018	17:54
13	J6403-28	1	Solid	20.02	20	20.1	8.17	12/13/2018	17:56
14	J6403-32	1	Solid	20.02	20	20.1	8.48	12/13/2018	17:58
15	J6403-36	1	Solid	20.04	20	19.9	8.48	12/13/2018	18:00
16	CCV2	1	Water			20.1	12.00	12/13/2018	18:02
17	J6403-40	1	Solid	20.04	20	20.0	8.62	12/13/2018	18:04
18	J6403-44	1	Solid	20.02	20	20.2	9.55	12/13/2018	18:06
19	J6403-44DUP	1	Solid	20.03	20	20.3	9.56	12/13/2018	18:08
20	J6403-48	1	Solid	20.02	20	20.2	9.69	12/13/2018	18:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J6403-52	1	Solid	20.02	20	20.0	9.09	12/13/2018	18:12
22	J6403-56	1	Solid	20.02	20	20.2	9.39	12/13/2018	18:14
23	J6403-60	1	Solid	20.02	20	19.9	8.74	12/13/2018	18:16
24	J6403-64	1	Solid	20.03	20	19.9	8.62	12/13/2018	18:18
25	CCV3	1	Water			20.2	2.00	12/13/2018	18:20

WORKLIST(Hardcopy Internal Chain)

1399808

WorkList Name : corrosivity121318

WorkList ID : 120329

Date : 12/13/2018 4:24:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Solid	J6403-04	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-1-A	12/13/2018	9045D
12/13/2018	Solid	J6403-08	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-1-B	12/13/2018	9045D
12/13/2018	Solid	J6403-12	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-2-A	12/13/2018	9045D
12/13/2018	Solid	J6403-16	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-2-B	12/13/2018	9045D
12/13/2018	Solid	J6403-20	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-3-A	12/13/2018	9045D
12/13/2018	Solid	J6403-24	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-3-B	12/13/2018	9045D
12/13/2018	Solid	J6403-28	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-4-A	12/13/2018	9045D
12/13/2018	Solid	J6403-32	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-4-B	12/13/2018	9045D
12/13/2018	Solid	J6403-36	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-5-A	12/13/2018	9045D
12/13/2018	Solid	J6403-40	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-5-B	12/13/2018	9045D
12/13/2018	Solid	J6403-44	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-6-A	12/13/2018	9045D
12/13/2018	Solid	J6403-48	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-6-B	12/13/2018	9045D
12/13/2018	Solid	J6403-52	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-7-A	12/13/2018	9045D
12/13/2018	Solid	J6403-56	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-7-B	12/13/2018	9045D
12/13/2018	Solid	J6403-60	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-8-A	12/13/2018	9045D
12/13/2018	Solid	J6403-64	Corrosivity	Cool 4 deg C	PSEG01	D21	TP-8-B	12/13/2018	9045D

Date/Time 12/13/18 5:50

Received by: SP

Relinquished by: SP

Date/Time 12/13/18 6:22

Received by: CP

Relinquished by: CP