

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/13/2018 16:27

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.044	0.0	0.108	
ICB1	-0.000	0.0	0.003	
CCV1	0.988	0.0	0.103	
CCB1	-0.001	0.0	0.002	
PB115451BL	0.001	0.0	0.003	
PB115451BS	1.021	0.0	0.106	
J6218-01	0.012	0.0	0.004	
J6218-01DUP	0.012	0.0	0.004	
J6218-01MS	0.904	0.0	0.094	
J6218-01MSD	0.909	0.0	0.095	
J6218-02	0.005	0.0	0.003	
J6218-03	0.006	0.0	0.003	
J6218-04	0.002	0.0	0.003	
J6218-05	0.009	0.0	0.004	
CCV2	0.981	0.0	0.102	
CCB2	0.004	0.0	0.003	
J6235-01	-0.000	0.0	0.003	
J6235-02	0.004	0.0	0.003	
J6235-03	0.005	0.0	0.003	
J6235-04	0.006	0.0	0.003	
J6235-05	0.005	0.0	0.003	
J6235-06	0.006	0.0	0.003	
CCV3	0.982	0.0	0.102	
CCB3	0.002	0.0	0.003	

N	24
Mean	0.288
SD	0.4518
CV%	156.98

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 13 16:02:35 2018

Thu Dec 13 16:16:36 2018

AS

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.05PPM	A	Phenolics-	P	0.0537	mg/l	12/13/2018 15:03:11
0.1PPM	A	Phenolics-	P	0.1018	mg/l	12/13/2018 15:03:12
0.25PPM	A	Phenolics-	P	0.255	mg/l	12/13/2018 15:03:13
0.50PPM	A	Phenolics-	P	0.502	mg/l	12/13/2018 15:03:14
1.0PPM	A	Phenolics-	P	0.9776	mg/l	12/13/2018 15:03:15
2.0PPM	A	Phenolics-	P	2.0099	mg/l	12/13/2018 15:03:16
ICV1	S	Phenolics-	P	1.0445	mg/l	12/13/2018 16:02:35
ICB1	S	Phenolics-	P	0	mg/l	12/13/2018 16:02:36
CCV1	S	Phenolics-	P	0.9882	mg/l	12/13/2018 16:02:37
CCB1	S	Phenolics-	P	-0.001	mg/l	12/13/2018 16:02:38
PB115451BL	S	Phenolics-	P	0.0009	mg/l	12/13/2018 16:02:39
PB115451BS	S	Phenolics-	P	1.0214	mg/l	12/13/2018 16:02:40
J6218-01	S	Phenolics-	P	0.0122	mg/l	12/13/2018 16:02:41
J6218-01DUP	S	Phenolics-	P	0.012	mg/l	12/13/2018 16:02:42
J6218-01MS	S	Phenolics-	P	0.9044	mg/l	12/13/2018 16:02:43
J6218-01MSD	S	Phenolics-	P	0.9091	mg/l	12/13/2018 16:02:44
J6218-02	S	Phenolics-	P	0.0053	mg/l	12/13/2018 16:02:45
J6218-03	S	Phenolics-	P	0.0056	mg/l	12/13/2018 16:02:46
J6218-04	S	Phenolics-	P	0.0019	mg/l	12/13/2018 16:12:37
J6218-05	S	Phenolics-	P	0.0094	mg/l	12/13/2018 16:12:38
CCV2	S	Phenolics-	P	0.9813	mg/l	12/13/2018 16:12:39
CCB2	S	Phenolics-	P	0.0035	mg/l	12/13/2018 16:12:40
J6235-01	S	Phenolics-	P	-0.0004	mg/l	12/13/2018 16:12:41
J6235-02	S	Phenolics-	P	0.0037	mg/l	12/13/2018 16:12:42
J6235-03	S	Phenolics-	P	0.005	mg/l	12/13/2018 16:12:43
J6235-04	S	Phenolics-	P	0.0062	mg/l	12/13/2018 16:12:44
J6235-05	S	Phenolics-	P	0.0046	mg/l	12/13/2018 16:12:45
J6235-06	S	Phenolics-	P	0.0058	mg/l	12/13/2018 16:12:46
CCV3	S	Phenolics-	P	0.9816	mg/l	12/13/2018 16:16:35
CCB3	S	Phenolics-	P	0.002	mg/l	12/13/2018 16:16:36

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/13/2018 15:45

Test Phenolics-


Accepted 12/13/2018 15:03

Factor 9.863

Bias 0.003

Coeff. of det. 0.999769

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0537	0.0500	
2	Phenol-2	0.013	0.1018	0.1000	
3	Phenol-2	0.028	0.2550	0.2500	
4	Phenol-2	0.053	0.5020	0.5000	
5	Phenol-2	0.102	0.9776	1.0000	
6	Phenol-2	0.206	2.0099	2.0000	