

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/13/2018

TEMP1 IN: 103 12/13/2018 14:00 TEMP OUT: 104 12/13/2018 15:20

TEMP2 IN: 104 12/13/2018 16:00 TEMP OUT: 104 12/14/2018 07:20

TEMP3 IN: 105 12/14/2018 09:00 TEMP OUT: 105 12/14/2018 10:40

TEMP4 IN: 105 12/14/2018 11:30 TEMP OUT: 105 12/14/2018 13:20

Run Number: LB99814

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99814BL	LB99814BL	1.4012	1.4012	100	1.4012	1.4012	1.4012	0.0000	0
2	LB99814BS	LB99814BS	1.3986	1.3986	100	1.4490	1.4490	1.4490	0.0504	504
3	J6385-01	121218-JETT-F-D-B	1.3801	1.3801	250	1.3869	1.3869	1.3869	0.0068	27.2
4	J6385-01DUP	121218-JETT-F-D-BDUP	1.3710	1.3710	250	1.3777	1.3777	1.3777	0.0067	26.8
5	J6385-02	121218-JETT-F-D-S	1.3957	1.3957	400	1.4044	1.4044	1.4044	0.0087	21.8
6	J6385-03	121218-JETT-F-U-B	1.3871	1.3871	400	1.3967	1.3967	1.3967	0.0096	24
7	J6385-04	121218-JETT-F-U-S	1.3820	1.3820	500	1.3924	1.3924	1.3924	0.0104	20.8
8	J6399-01	SPDES-1	1.3916	1.3916	900	1.3926	1.3926	1.3926	0.0010	1.1
9	J6403-65	TP-3-OILY-WATER	1.4049	1.4049	10	1.4485	1.4485	1.4485	0.0436	4360

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS121418

WorkList ID : 120336

Date : 12/14/2018 7:49:34 AM

21399814

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/19/2018	Water	J6385-01 C	TSS	Cool 4 deg C	HDR103	D51	121218-JETT-F-D-B	12/12/2018	SM2540 D
12/19/2018	Water	J6385-02 C	TSS	Cool 4 deg C	HDR103	D51	121218-JETT-F-D-S	12/12/2018	SM2540 D
12/19/2018	Water	J6385-03 C	TSS	Cool 4 deg C	HDR103	D51	121218-JETT-F-U-B	12/12/2018	SM2540 D
12/19/2018	Water	J6385-04 C	TSS	Cool 4 deg C	HDR103	D51	121218-JETT-F-U-S	12/12/2018	SM2540 D
12/14/2018	Water	J6399-01 D	TSS	Cool 4 deg C	SKAN06	D32	SPDES-1	12/13/2018	SM2540 D
12/18/2018	Water	J6403-65 C	TSS	Cool 4 deg C	PSEG01	D22	TP-3-OILY-WATER	12/13/2018	SM2540 D

PR-12791048

Date/Time 12/14/18, 08:00
 Received by: AS
 Relinquished by: AS

Date/Time 12/14/18, 10:30
 Received by: AS
 Relinquished by: AS