

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB99825

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP70281
calibration std. phosphate 0.5 ppm	WP70280
calibration std. phosphate 0.3 ppm	WP70279
calibration std. 0 ppm	WP70276
phosphate ICV std.	WP70284
phosphate CCV std.	WP70283
calibration std. phosphate 0.1 ppm	WP70278
calibration std. phosphate 0.05 ppm	WP70277
Combined reagent	WP70287
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0008Slope: 0.6506Regression: 0.99981

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/14/2018	15:20
2	CAL2	0.05	1	50	50	0.033	12/14/2018	15:20
3	CAL3	0.10	1	50	50	0.065	12/14/2018	15:21
4	CAL4	0.30	1	50	50	0.186	12/14/2018	15:21
5	CAL5	0.50	1	50	50	0.331	12/14/2018	15:22
6	CAL6	1.00	1	50	50	0.649	12/14/2018	15:23



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB99825

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.327	0.504	12/14/2018	15:23
2	ICB		1	50	50	0.000	0.001	12/14/2018	15:23
3	CCV1	0.50	1	50	50	0.328	0.505	12/14/2018	15:24
4	CCB1		1	50	50	0.000	0.001	12/14/2018	15:24
5	PB115704BL		1	50	50	0.000	0.001	12/14/2018	15:25
6	PB115704BS	0.50	1	50	50	0.320	0.493	12/14/2018	15:25
7	J6421-01		1	50	50	0.369	0.568	12/14/2018	15:26
8	J6421-01DUP		1	50	50	0.375	0.578	12/14/2018	15:26
9	J6421-01MS	0.50	1	50	50	0.645	0.993	12/14/2018	15:27
10	J6421-01MSD	0.50	1	50	50	0.648	0.997	12/14/2018	15:27
11	CCV2	0.50	1	50	50	0.327	0.504	12/14/2018	15:28
12	CCB2		1	50	50	0.000	0.001	12/14/2018	15:28