

LB 99839


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39073



A= 0.0000e+000

B= 2.5754e-004

C= 3.0355e-002

Rho= 0.9998186

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	-16	0.000	-16				
0.2	0.200	0.233	0.033	786	0.0 %	786				
2.5	2.500	2.527	0.027	9695	0.0 %	9695				
5.0	5.000	4.876	-0.124	18816	0.0 %	18816				
7.5	7.500	7.444	-0.056	28787	0.0 %	28787				
10.0	10.000	10.093	0.093	39073	0.0 %	39073				

LB99839 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Dec 2018 11:27:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-16	-	0.0000	7470A	PPB	15 Dec 2018 11:36:08	S	Std
0.2 - 1	SO	786	-	0.2000	7470A	PPB	15 Dec 2018 11:38:17	S	Std
2.5 - 1	52.5	9695	-	2.5000	7470A	PPB	15 Dec 2018 11:40:26	S	Std
5.0 - 1	55	18816	-	5.0000	7470A	PPB	15 Dec 2018 11:42:34	S	Std
7.5 - 1	57.5	28787	-	7.5000	7470A	PPB	15 Dec 2018 11:44:42	S	Std
10.0 - 1	570	39073	-	10.0000	7470A	PPB	15 Dec 2018 11:46:50	S	Std
ICV26 - 1	ICV26	14891	3.8654	-	7470A	PPB	15 Dec 2018 11:49:23	U	SMPL
ICB26 - 1	ICB26	-74	0.0113	-	7470A	PPB	15 Dec 2018 11:51:29	U	SMPL
CCV12 - 1	CCV12	18709	4.8487	-	7470A	PPB	15 Dec 2018 11:53:36	U	SMPL
CCB12 - 1	CCB12	-58	0.0154	-	7470A	PPB	15 Dec 2018 11:55:43	U	SMPL
CRA - 1	CRA	773	0.2294	-	7470A	PPB	15 Dec 2018 11:57:50	U	SMPL
HighStd - 1	HighStd	39097	10.0995	-	7470A	PPB	15 Dec 2018 11:59:56	U	SMPL
ChkStd - 1	ChkStd	26664	6.8974	-	7470A	PPB	15 Dec 2018 12:02:04	U	SMPL
PB115706BS - 1	LCSW	13823	3.5904	-	7470A	PPB	15 Dec 2018 12:06:21	U	SMPL
J6390-01 - 1	20380	279	0.1022	-	7470A	PPB	15 Dec 2018 12:08:29	U	SMPL
J6403-04 - 1	TP-1-A	515	0.1630	-	7470A	PPB	15 Dec 2018 12:10:37	U	SMPL
J6403-04DUP - 1	TP-1-ADUP	527	0.1661	-	7470A	PPB	15 Dec 2018 12:15:10	U	SMPL
J6403-04MS - 1	TP-1-AMS	13663	3.5491	-	7470A	PPB	15 Dec 2018 12:19:47	U	SMPL
J6403-04MSD - 1	TP-1-AMSD	13485	3.5033	-	7470A	PPB	15 Dec 2018 12:21:54	U	SMPL
CCV13 - 1	CCV13	18929	4.9054	-	7470A	PPB	15 Dec 2018 12:24:01	U	SMPL
CCB13 - 1	CCB13	-49	0.0177	-	7470A	PPB	15 Dec 2018 12:26:07	U	SMPL
J6403-08 - 1	TP-1-B	446	0.1452	-	7470A	PPB	15 Dec 2018 12:28:14	U	SMPL
J6403-12 - 1	TP-2-A	535	0.1681	-	7470A	PPB	15 Dec 2018 12:30:20	U	SMPL
J6403-16 - 1	TP-2-B	944	0.2735	-	7470A	PPB	15 Dec 2018 12:32:27	U	SMPL
J6403-20 - 1	TP-3-A	722	0.2163	-	7470A	PPB	15 Dec 2018 12:34:34	U	SMPL
J6403-24 - 1	TP-3-B	10	0.0329	-	7470A	PPB	15 Dec 2018 12:36:42	U	SMPL
J6403-28 - 1	TP-4-A	42	0.0412	-	7470A	PPB	15 Dec 2018 12:38:50	U	SMPL
J6403-32 - 1	TP-4-B	206	0.0834	-	7470A	PPB	15 Dec 2018 12:40:57	U	SMPL
J6403-36 - 1	TP-5-A	503	0.1599	-	7470A	PPB	15 Dec 2018 12:43:06	U	SMPL
J6403-40 - 1	TP-5-B	807	0.2382	-	7470A	PPB	15 Dec 2018 12:45:14	U	SMPL
J6403-44 - 1	TP-6-A	891	0.2598	-	7470A	PPB	15 Dec 2018 12:47:21	U	SMPL
CCV14 - 1	CCV14	18720	4.8515	-	7470A	PPB	15 Dec 2018 12:49:29	U	SMPL
CCB14 - 1	CCB14	-37	0.0208	-	7470A	PPB	15 Dec 2018 12:51:36	U	SMPL
J6403-48 - 1	TP-6-B	217	0.0862	-	7470A	PPB	15 Dec 2018 12:53:43	U	SMPL
J6403-52 - 1	TP-7-A	216	0.0860	-	7470A	PPB	15 Dec 2018 12:55:50	U	SMPL
J6403-56 - 1	TP-7-B	178	0.0762	-	7470A	PPB	15 Dec 2018 12:57:57	U	SMPL
J6403-60 - 1	TP-8-A	287	0.1043	-	7470A	PPB	15 Dec 2018 13:00:04	U	SMPL
J6403-64 - 1	TP-8-B	292	0.1056	-	7470A	PPB	15 Dec 2018 13:02:12	U	SMPL
PB115706BL - 1	PBW	-7	0.0286	-	7470A	PPB	15 Dec 2018 13:04:21	U	SMPL
PB115706TB - 1	PB115666TB	103	0.0569	-	7470A	PPB	15 Dec 2018 13:09:41	U	SMPL
J6403-04LX5 - 1		139	0.0662	-	7470A	PPB	15 Dec 2018 13:11:49	U	SMPL
J6403-04A - 1		16109	4.1791	-	7470A	PPB	15 Dec 2018 13:13:57	U	SMPL
CCV15 - 1	CCV15	18398	4.7686	-	7470A	PPB	15 Dec 2018 13:16:05	U	SMPL
CCB15 - 1	CCB15	-36	0.0211	-	7470A	PPB	15 Dec 2018 13:18:12	U	SMPL