


LB99840

7471B

INSTRUMENT ID: CV1

 μ Abs.:

55111



Linear

A= 0.0000e+000

B= 1.8009e-004

C= -1.0428e-003

Rho= 0.9998158

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	44	0.000	44				
0.2	0.200	0.203	0.003	1132	0.0 %	1132				
2.5	2.500	2.484	-0.016	13798	0.0 %	13798				
5.0	5.000	4.939	-0.061	27433	0.0 %	27433				
7.5	7.500	7.643	0.143	42443	0.0 %	42443				
10.0	10.000	9.924	-0.076	55111	0.0 %	55111				

LB99840 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Dec 2018 13:49:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	44	-	0.0000	7471B	PPB	15 Dec 2018 13:56:15	S	Std
0.2 - 1	50.2	1132	-	0.2000	7471B	PPB	15 Dec 2018 13:58:33	S	Std
2.5 - 1	52.5	13798	-	2.5000	7471B	PPB	15 Dec 2018 14:00:51	S	Std
5.0 - 1	55	27433	-	5.0000	7471B	PPB	15 Dec 2018 14:03:09	S	Std
7.5 - 1	57.5	42443	-	7.5000	7471B	PPB	15 Dec 2018 14:05:26	S	Std
10.0 - 1	59.1	55111	-	10.0000	7471B	PPB	15 Dec 2018 14:07:44	S	Std
ICV27 - 1	ICV27	22098	3.9787	-7471B	PPB	PPB	15 Dec 2018 14:10:32	U	SMPL
ICB27 - 1	ICB27	-34	-0.0072	-7471B	PPB	PPB	15 Dec 2018 14:17:57	U	SMPL
CCV16 - 1	CCV16	28699	5.1675	-7471B	PPB	PPB	15 Dec 2018 14:20:12	U	SMPL
CCB16 - 1	CCB16	-131	-0.0246	-7471B	PPB	PPB	15 Dec 2018 14:22:28	U	SMPL
CRA - 1	CRA	1139	0.2041	-7471B	PPB	PPB	15 Dec 2018 14:24:45	U	SMPL
HighStd - 1	HighStd	54861	9.8791	-7471B	PPB	PPB	15 Dec 2018 14:27:01	U	SMPL
ChkStd - 1	ChkStd	37462	6.7457	-7471B	PPB	PPB	15 Dec 2018 14:29:18	U	SMPL
PB115714BL - 1	PBS	-7	-0.0023	-7471B	PPB	PPB	15 Dec 2018 14:31:36	U	SMPL
PB115714BS - 1	LCSS	21199	3.8168	-7471B	PPB	PPB	15 Dec 2018 14:36:29	U	SMPL
J6386-01 - 1	P001-SS006-1824-02	4758	0.8558	-7471B	PPB	PPB	15 Dec 2018 15:00:53	U	SMPL
J6386-02 - 1	P001-SS015-0612-02	8856	1.5939	-7471B	PPB	PPB	15 Dec 2018 15:03:10	U	SMPL
J6386-03 - 1	P001-SS016-0006-02	5364	0.9650	-7471B	PPB	PPB	15 Dec 2018 15:05:27	U	SMPL
J6386-05 - 1	P001-SS005-0006-01	3906	0.7024	-7471B	PPB	PPB	15 Dec 2018 15:07:45	U	SMPL
J6386-06 - 1	P001-SS005-0612-01	1336	0.2396	-7471B	PPB	PPB	15 Dec 2018 15:10:01	U	SMPL
CCV17 - 1	CCV17	29050	5.2307	-7471B	PPB	PPB	15 Dec 2018 15:12:16	U	SMPL
CCB17 - 1	CCB17	-142	-0.0266	-7471B	PPB	PPB	15 Dec 2018 15:14:32	U	SMPL
J6386-07 - 1	P001-SS005-1218-01	1540	0.2763	-7471B	PPB	PPB	15 Dec 2018 15:16:50	U	SMPL
J6386-08 - 1	P001-SS005-1824-01	1182	0.2118	-7471B	PPB	PPB	15 Dec 2018 15:19:06	U	SMPL
J6386-09 - 1	P001-SS006-0006-01	5035	0.9057	-7471B	PPB	PPB	15 Dec 2018 15:21:22	U	SMPL
J6386-10 - 1	P001-SS006-0612-01	6127	1.1024	-7471B	PPB	PPB	15 Dec 2018 15:23:38	U	SMPL
J6386-11 - 1	P001-SS006-1218-01	5701	1.0257	-7471B	PPB	PPB	15 Dec 2018 15:25:55	U	SMPL
J6386-12 - 1	P001-SS006-1824-01	5666	1.0194	-7471B	PPB	PPB	15 Dec 2018 15:28:12	U	SMPL
J6386-12DUP - 1	P001-SS006-1824-01DUP	5346	0.9617	-7471B	PPB	PPB	15 Dec 2018 15:30:29	U	SMPL
J6386-13 - 1	J6386-12MS	23916	4.3061	-7471B	PPB	PPB	15 Dec 2018 15:32:46	U	SMPL
J6386-14 - 1	J6386-12MSD	23731	4.2728	-7471B	PPB	PPB	15 Dec 2018 15:35:04	U	SMPL
J6386-15 - 1	P001-SS007-0006-01	6551	1.1788	-7471B	PPB	PPB	15 Dec 2018 15:37:21	U	SMPL
CCV18 - 1	CCV18	28541	5.1390	-7471B	PPB	PPB	15 Dec 2018 15:39:38	U	SMPL
CCB18 - 1	CCB18	-147	-0.0275	-7471B	PPB	PPB	15 Dec 2018 15:41:55	U	SMPL
J6386-16 - 1	P001-SS007-0612-01	6491	1.1680	-7471B	PPB	PPB	15 Dec 2018 15:44:13	U	SMPL
J6386-17 - 1	P001-SS007-1218-01	8183	1.4727	-7471B	PPB	PPB	15 Dec 2018 15:46:29	U	SMPL
J6386-18 - 1	P001-SS007-1824-01	7433	1.3376	-7471B	PPB	PPB	15 Dec 2018 15:48:46	U	SMPL
J6386-19 - 1	P001-SS008-0006-01	7593	1.3664	-7471B	PPB	PPB	15 Dec 2018 15:51:02	U	SMPL
J6386-20 - 1	P001-SS008-0612-01	8277	1.4896	-7471B	PPB	PPB	15 Dec 2018 15:53:19	U	SMPL
PB115715BL - 1	PBS	-27	-0.0059	-7471B	PPB	PPB	15 Dec 2018 15:55:35	U	SMPL
PB115715BS - 1	LCSS	22158	3.9895	-7471B	PPB	PPB	15 Dec 2018 15:57:53	U	SMPL
J6391-01 - 1	P001-SS015-0612-01	6746	1.2139	-7471B	PPB	PPB	15 Dec 2018 16:00:10	U	SMPL
J6391-02 - 1	J6391-01MS	26856	4.8356	-7471B	PPB	PPB	15 Dec 2018 16:05:16	U	SMPL
J6391-03 - 1	J6391-01MSD	27191	4.8959	-7471B	PPB	PPB	15 Dec 2018 16:07:32	U	SMPL
CCV19 - 1	CCV19	28633	5.1556	-7471B	PPB	PPB	15 Dec 2018 16:09:49	U	SMPL
CCB19 - 1	CCB19	-104	-0.0198	-7471B	PPB	PPB	15 Dec 2018 16:12:07	U	SMPL
J6391-04 - 1	P001-SS015-1218-01	5292	0.9520	-7471B	PPB	PPB	15 Dec 2018 16:14:24	U	SMPL
J6391-05 - 1	P001-SS015-1824-01	1490	0.2673	-7471B	PPB	PPB	15 Dec 2018 16:16:41	U	SMPL
J6391-06 - 1	P001-SS016-0006-01	4395	0.7905	-7471B	PPB	PPB	15 Dec 2018 16:18:57	U	SMPL
J6391-06DUP - 1	P001-SS016-0006-01DUP	4694	0.8443	-7471B	PPB	PPB	15 Dec 2018 16:21:13	U	SMPL
J6391-07 - 1	J6391-06MS	22727	4.0920	-7471B	PPB	PPB	15 Dec 2018 16:23:30	U	SMPL
J6391-08 - 1	J6391-06MSD	22611	4.0711	-7471B	PPB	PPB	15 Dec 2018 16:25:47	U	SMPL
J6391-09 - 1	P001-SS016-0612-01	3263	0.5866	-7471B	PPB	PPB	15 Dec 2018 16:28:05	U	SMPL
J6391-10 - 1	P001-SS016-1218-01	6587	1.1852	-7471B	PPB	PPB	15 Dec 2018 16:30:23	U	SMPL
J6391-11 - 1	P001-SS016-1824-01	7636	1.3742	-7471B	PPB	PPB	15 Dec 2018 16:32:41	U	SMPL
J6391-12 - 1	P001-SS017-0006-01	3549	0.6381	-7471B	PPB	PPB	15 Dec 2018 16:34:58	U	SMPL
CCV20 - 1	CCV20	27976	5.0373	-7471B	PPB	PPB	15 Dec 2018 16:37:14	U	SMPL
CCB20 - 1	CCB20	-86	-0.0165	-7471B	PPB	PPB	15 Dec 2018 16:39:31	U	SMPL
J6391-13 - 1	P001-SS017-0612-01	9210	1.6576	-7471B	PPB	PPB	15 Dec 2018 16:41:48	U	SMPL
J6391-14 - 1	P001-SS017-1218-01	8254	1.4855	-7471B	PPB	PPB	15 Dec 2018 16:44:04	U	SMPL
J6391-15 - 1	P001-SS017-1824-01	4201	0.7555	-7471B	PPB	PPB	15 Dec 2018 16:46:21	U	SMPL
J6391-16 - 1	P001-SS018-0006-01	4390	0.7896	-7471B	PPB	PPB	15 Dec 2018 16:48:37	U	SMPL
J6391-17 - 1	P001-SS018-0612-01	5127	0.9223	-7471B	PPB	PPB	15 Dec 2018 16:50:54	U	SMPL
J6391-18 - 1	P001-SS018-1218-01	5774	1.0388	-7471B	PPB	PPB	15 Dec 2018 16:53:11	U	SMPL
J6391-19 - 1	P001-SS018-1824-01	3457	0.6215	-7471B	PPB	PPB	15 Dec 2018 16:55:28	U	SMPL
CCV21 - 1	CCV21	27731	4.9932	-7471B	PPB	PPB	15 Dec 2018 16:57:46	U	SMPL
CCB21 - 1	CCB21	-58	-0.0115	-7471B	PPB	PPB	15 Dec 2018 17:00:04	U	SMPL
J6386-12LX5 - 1		1547	0.2776	-7471B	PPB	PPB	15 Dec 2018 17:02:20	U	SMPL
J6386-12A - 1		26880	4.8399	-7471B	PPB	PPB	15 Dec 2018 17:04:37	U	SMPL
J6391-06LX5 - 1		1161	0.2080	-7471B	PPB	PPB	15 Dec 2018 17:06:54	U	SMPL
J6391-06A - 1		25314	4.5579	-7471B	PPB	PPB	15 Dec 2018 17:09:12	U	SMPL
CCV22 - 1	CCV22	28263	5.0890	-7471B	PPB	PPB	15 Dec 2018 17:11:28	U	SMPL
CCB22 - 1	CCB22	-156	-0.0291	-7471B	PPB	PPB	15 Dec 2018 17:13:47	U	SMPL