



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB99842

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/15/2018	12:10
2	CAL2	1	Water			20.1	7.00	12/15/2018	12:11
3	CAL3	1	Water			20.1	10.03	12/15/2018	12:13
4	ICV	1	Water			20.1	7.00	12/15/2018	12:15
5	CCV1	1	Water			20.2	2.01	12/15/2018	12:16
6	J6438-05	1	Water			18.5	5.36	12/15/2018	12:21
7	J6438-09	1	Water			18.4	5.89	12/15/2018	12:23
8	J6438-13	1	Water			18.8	5.75	12/15/2018	12:25
9	J6438-17	1	Water			19.0	6.63	12/15/2018	12:26
10	J6440-02	1	Water			20.2	6.60	12/15/2018	12:28
11	J6440-05	1	Water			18.2	5.70	12/15/2018	12:30
12	J6440-08	1	Water			18.5	6.02	12/15/2018	12:31
13	J6440-08DUP	1	Water			18.6	6.03	12/15/2018	12:32
14	CCV2	1	Water			20.2	12.02	12/15/2018	12:34

2099842

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J6425

WorkList ID : 120426

Date : 12/15/2018 11:24:42 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/14/2018	Water	J6438-05 -13	pH	Cool 4 deg C	PSEG01	E21	COMP-1	12/14/2018	9040C
12/14/2018	Water	J6438-09 -13	pH	Cool 4 deg C	PSEG01	E21	COMP-2	12/14/2018	9040C
12/14/2018	Water	J6438-13 -13	pH	Cool 4 deg C	PSEG01	E21	COMP-3	12/14/2018	9040C
12/14/2018	Water	J6438-17 -13	pH	Cool 4 deg C	PSEG01	E21	COMP-4	12/14/2018	9040C
12/14/2018	Water	J6440-02 -13	pH	Cool 4 deg C	PSEG01	E12	COMP-2	12/14/2018	9040C
12/14/2018	Water	J6440-05 -13	pH	Cool 4 deg C	PSEG01	E12	COMP-3	12/14/2018	9040C
12/14/2018	Water	J6440-08 -13	pH	Cool 4 deg C	PSEG01	E12	COMP-4	12/14/2018	9040C

Date/Time 12-15-18 11:31 PM
Received by: 30
Relinquished by: 69

Date/Time 12-15-18 2:00 PM
Received by: 60
Relinquished by: 30