



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB99843

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.0 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/15/2018	12:10
2	CAL2	1	Water			20.1	7.00	12/15/2018	12:11
3	CAL3	1	Water			20.1	10.03	12/15/2018	12:13
4	ICV	1	Water			20.1	7.00	12/15/2018	12:15
5	CCV1	1	Water			20.2	2.01	12/15/2018	12:40
6	J6429-04	1	Solid	20.03	20	19.9	7.91	12/15/2018	12:41
7	J6429-04DUP	1	Solid	20.04	20	19.8	7.92	12/15/2018	12:43
8	CCV2	1	Water			20.1	12.02	12/15/2018	12:44

2B 09043

WORKLIST(Hardcopy Internal Chain)

Date : 12/15/2018 11:25:59 AM

WorkList ID : 120428

WorkList Name : CORROSIONITY J6429

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/14/2018	Solid	J6429-04	Corrosivity	Cool 4 deg C	PSEG01	E11	OILY-STONES	12/14/2018	9045D

Date/Time 12-5-18 11:31 AM
Received by: J
Relinquished by: C

Date/Time 12-15-18 3:30 PM
Received by: S
Relinquished by: R