


LB 99844

INSTRUMENT 10: CV 1

$$\mu \text{ Abs.} :$$

40865



A= 0.0000e+000

B= 2.4750e-004

C= -1.5695e-001

Rho= 0.9999259

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.025	-0.025	535	0.000	535					
0.2	0.200	0.185	-0.015	1382	0.0 %	1382					
2.5	2.500	2.492	-0.008	10701	0.0 %	10701					
5.0	5.000	5.097	0.097	21228	0.0 %	21228					
7.5	7.500	7.493	-0.007	30908	0.0 %	30908					
10.0	10.000	9.958	-0.042	40869	0.0 %	40869					

$\frac{\%D}{-}$
 $\frac{1}{-8}$
 $\frac{0}{2}$
 $\frac{0}{0}$
 MB
 12/15/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	535	-		0 ISM02.4	PPB	12/15/2018 17:29	Std
	0.2 S0.2	1382	-		0.2 ISM02.4	PPB	12/15/2018 17:32	Std
	2.5 S2.5	10701	-		2.5 ISM02.4	PPB	12/15/2018 17:34	Std
	5 S5	21228	-		5 ISM02.4	PPB	12/15/2018 17:36	Std
	7.5 S7.5	30908	-		7.5 ISM02.4	PPB	12/15/2018 17:38	Std
	10 S10.0	40869	-		10 ISM02.4	PPB	12/15/2018 17:41	Std
ICV	ICV	16846	4.0124 -		ISM02.4	PPB	12/15/2018 17:44	SMPL
ICB	ICB	479	-0.0384 -		ISM02.4	PPB	12/15/2018 17:46	SMPL
CCV023	CCV023	21265	5.1061 -		ISM02.4	PPB	12/15/2018 17:48	SMPL
CCB023	CCB023	478	-0.0386 -		ISM02.4	PPB	12/15/2018 17:50	SMPL
J6276-01	MC0AT8	1395	0.1883 -		ISM02.4	PPB	12/15/2018 17:57	SMPL
J6276-02	MC0AW2	1219	0.1448 -		ISM02.4	PPB	12/15/2018 18:00	SMPL
J6276-03	MC0AW5	1171	0.1329 -		ISM02.4	PPB	12/15/2018 18:02	SMPL
J6276-04	MC0AW7	1174	0.1336 -		ISM02.4	PPB	12/15/2018 18:04	SMPL
J6276-05	MC0AW3	1236	0.149 -		ISM02.4	PPB	12/15/2018 18:06	SMPL
J6276-06	MC0AW4	1420	0.1945 -		ISM02.4	PPB	12/15/2018 18:09	SMPL
J6276-07	MC0AW9	692	0.0143 -		ISM02.4	PPB	12/15/2018 18:11	SMPL
J6276-08	MC0AX0	386	-0.0614 -		ISM02.4	PPB	12/15/2018 18:13	SMPL
J6276-09	MC0AX1	354	-0.0693 -		ISM02.4	PPB	12/15/2018 18:16	SMPL
J6276-10	MC0AX2	357	-0.0686 -		ISM02.4	PPB	12/15/2018 18:18	SMPL
J6276-11	MC0AX3	359	-0.0681 -		ISM02.4	PPB	12/15/2018 18:20	SMPL
J6276-12	MC0AX4	416	-0.054 -		ISM02.4	PPB	12/15/2018 18:22	SMPL
J6276-13	MC0AX5	398	-0.0584 -		ISM02.4	PPB	12/15/2018 18:25	SMPL
J6276-14	MC0AX5D	390	-0.0604 -		ISM02.4	PPB	12/15/2018 18:27	SMPL
J6276-15	MC0AX5S	4525	0.963 -		ISM02.4	PPB	12/15/2018 18:29	SMPL
J6276-16	MC0AX7	361	-0.0676 -		ISM02.4	PPB	12/15/2018 18:32	SMPL
J6276-17	MC0AX8	328	-0.0758 -		ISM02.4	PPB	12/15/2018 18:34	SMPL
J6276-18	MC0AX9	326	-0.0763 -		ISM02.4	PPB	12/15/2018 18:36	SMPL
J6276-19	MC0AY0	385	-0.0617 -		ISM02.4	PPB	12/15/2018 18:38	SMPL
CCV024	CCV024	21786	5.2351 -		ISM02.4	PPB	12/15/2018 18:41	SMPL
CCB024	CCB024	467	-0.0414 -		ISM02.4	PPB	12/15/2018 18:43	SMPL
PB115725BL	PBW725	551	-0.0206 -		ISM02.4	PPB	12/15/2018 18:48	SMPL
J6320-01	MC0AZ4	459	-0.0433 -		ISM02.4	PPB	12/15/2018 18:50	SMPL
J6320-02	MC0AZ4D	460	-0.0431 -		ISM02.4	PPB	12/15/2018 18:52	SMPL
J6320-03	MC0AZ4S	4550	0.9692 -		ISM02.4	PPB	12/15/2018 18:55	SMPL
J6320-04	MC0AZ6	432	-0.05 -		ISM02.4	PPB	12/15/2018 18:57	SMPL
J6320-05	MC0AZ7	3985	0.8293 -		ISM02.4	PPB	12/15/2018 18:59	SMPL
J6320-06	MC0AZ8	1456	0.2034 -		ISM02.4	PPB	12/15/2018 19:01	SMPL
J6320-07	MC0AZ9	465	-0.0419 -		ISM02.4	PPB	12/15/2018 19:04	SMPL
J6320-08	MC0B00	397	-0.0587 -		ISM02.4	PPB	12/15/2018 19:06	SMPL
J6320-09	MC0B04	399	-0.0582 -		ISM02.4	PPB	12/15/2018 19:08	SMPL
J6320-10	MC0B05	398	-0.0584 -		ISM02.4	PPB	12/15/2018 19:11	SMPL
J6320-11	MC0B06	509	-0.031 -		ISM02.4	PPB	12/15/2018 19:13	SMPL
J6320-12	MC0B07	364	-0.0669 -		ISM02.4	PPB	12/15/2018 19:15	SMPL
J6320-13	MC0B08	365	-0.0666 -		ISM02.4	PPB	12/15/2018 19:17	SMPL
J6320-14	MC0B09	354	-0.0693 -		ISM02.4	PPB	12/15/2018 19:20	SMPL

INSTRUMENT ID : CV1

J6320-15	MC0B10	859	0.0557 -	ISM02.4 PPB	12/15/2018 19:22 SMPL
J6320-16	MC0B11	336	-0.0738 -	ISM02.4 PPB	12/15/2018 19:24 SMPL
J6320-17	MC0B13	348	-0.0708 -	ISM02.4 PPB	12/15/2018 19:27 SMPL
J6320-18	MC0B14	413	-0.0547 -	ISM02.4 PPB	12/15/2018 19:29 SMPL
J6320-19	MC0B15	359	-0.0681 -	ISM02.4 PPB	12/15/2018 19:31 SMPL
CCV025	CCV025	21587	5.1858 -	ISM02.4 PPB	12/15/2018 19:33 SMPL
CCB025	CCB025	465	-0.0419 -	ISM02.4 PPB	12/15/2018 19:36 SMPL
J6320-20	MC0B16	799	0.0408 -	ISM02.4 PPB	12/15/2018 19:38 SMPL
J6320-21	MC0B17	1614	0.2425 -	ISM02.4 PPB	12/15/2018 19:40 SMPL
J6320-22	MC0B18	1651	0.2517 -	ISM02.4 PPB	12/15/2018 19:43 SMPL
PB115726BL	PBW726	535	-0.0245 -	ISM02.4 PPB	12/15/2018 19:45 SMPL
J6365-01	MC0B19	315	-0.079 -	ISM02.4 PPB	12/15/2018 19:47 SMPL
J6365-02	MC0B20	359	-0.0681 -	ISM02.4 PPB	12/15/2018 19:49 SMPL
J6365-03	MC0B21	434	-0.0495 -	ISM02.4 PPB	12/15/2018 19:52 SMPL
J6365-04	MC0B22	319	-0.078 -	ISM02.4 PPB	12/15/2018 19:54 SMPL
J6365-05	MC0B23	384	-0.0619 -	ISM02.4 PPB	12/15/2018 19:56 SMPL
J6365-06	MC0B24	334	-0.0743 -	ISM02.4 PPB	12/15/2018 19:58 SMPL
J6365-07	MC0B26	282	-0.0872 -	ISM02.4 PPB	12/15/2018 20:01 SMPL
J6365-08	MC0B27	388	-0.0609 -	ISM02.4 PPB	12/15/2018 20:03 SMPL
J6365-09	MC0B28	412	-0.055 -	ISM02.4 PPB	12/15/2018 20:05 SMPL
J6365-10	MC0B29	342	-0.0723 -	ISM02.4 PPB	12/15/2018 20:08 SMPL
J6365-11	MC0B30	338	-0.0733 -	ISM02.4 PPB	12/15/2018 20:10 SMPL
J6365-12	MC0B31	3715	0.7625 -	ISM02.4 PPB	12/15/2018 20:12 SMPL
J6365-13	MC0B32	278	-0.0881 -	ISM02.4 PPB	12/15/2018 20:14 SMPL
J6365-14	MC0B33	315	-0.079 -	ISM02.4 PPB	12/15/2018 20:17 SMPL
J6365-15	MC0B33D	402	-0.0575 -	ISM02.4 PPB	12/15/2018 20:19 SMPL
J6365-16	MC0B33S	4482	0.9523 -	ISM02.4 PPB	12/15/2018 20:21 SMPL
CCV026	CCV026	21831	5.2462 -	ISM02.4 PPB	12/15/2018 20:24 SMPL
CCB026	CCB026	497	-0.0339 -	ISM02.4 PPB	12/15/2018 20:26 SMPL
J6365-17	MC0B34	364	-0.0669 -	ISM02.4 PPB	12/15/2018 20:28 SMPL
PB115724BL	PBW724	504	-0.0322 -	ISM02.4 PPB	12/15/2018 20:30 SMPL
CCV027	CCV027	21981	5.2833 -	ISM02.4 PPB	12/15/2018 20:35 SMPL
CCB027	CCB027	486	-0.0367 -	ISM02.4 PPB	12/15/2018 20:37 SMPL