



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/17/2018

TEMP1 IN: 103 12/17/2018 07:00 TEMP OUT: 104 12/17/2018 08:10

TEMP2 IN: 104 12/17/2018 09:00 TEMP OUT: 104 12/17/2018 10:05

TEMP3 IN: 105 12/17/2018 11:30 TEMP OUT: 105 12/17/2018 13:10

TEMP4 IN: 105 12/17/2018 14:00 TEMP OUT: 105 12/17/2018 15:30

Run Number: LB99848

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99848BL	LB99848BL	1.4123	1.4123	100	1.4123	1.4123	1.4123	0.0000	0
2	LB99848BS	LB99848BS	1.4015	1.4015	100	1.4520	1.4520	1.4520	0.0505	505
3	J6417-01	121318-TIPC-F-D-B	1.4021	1.4021	250	1.4069	1.4069	1.4069	0.0048	19.2
4	J6417-01DUP	121318-TIPC-F-D-BDUP	1.3785	1.3785	250	1.3832	1.3832	1.3832	0.0047	18.8
5	J6417-02	121318-TIPC-F-D-S	1.4022	1.4022	400	1.4093	1.4093	1.4093	0.0071	17.8
6	J6417-03	121318-TIPC-F-U-B	1.3999	1.3999	500	1.4075	1.4075	1.4075	0.0076	15.2
7	J6417-04	121318-TIPC-F-U-S	1.4272	1.4272	500	1.4354	1.4354	1.4354	0.0082	16.4
8	J6421-01	EFFLUENT	1.3983	1.3983	20	1.4181	1.4181	1.4181	0.0198	990
9	J6421-02	AER-TANK-1	1.3987	1.3987	10	1.5109	1.5109	1.5109	0.1122	11220
10	J6433-01	001M-WILLETS-PT-BLVD (NOV)	1.3988	1.3988	300	1.4086	1.4086	1.4086	0.0098	32.7
11	J6433-02	001M-WILLETS-PT-BLVD (DEC)	1.4052	1.4052	200	1.4179	1.4179	1.4179	0.0127	63.5
12	J6433-03	002M-35TH-AVE (DEC)	1.4078	1.4078	300	1.4178	1.4178	1.4178	0.0100	33.3
13	J6438-05	COMP-1	1.3810	1.3810	50	1.3957	1.3957	1.3957	0.0147	294
14	J6438-09	COMP-2	1.4040	1.4040	50	1.4179	1.4179	1.4179	0.0139	278
15	J6438-13	COMP-3	1.4052	1.4052	20	1.4473	1.4473	1.4473	0.0421	2105

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/17/2018

TEMP1 IN: 103 12/17/2018 07:00 TEMP OUT: 104 12/17/2018 08:10

TEMP2 IN: 104 12/17/2018 09:00 TEMP OUT: 104 12/17/2018 10:05

TEMP3 IN: 105 12/17/2018 11:30 TEMP OUT: 105 12/17/2018 13:10

TEMP4 IN: 105 12/17/2018 14:00 TEMP OUT: 105 12/17/2018 15:30

Run Number: LB99848

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J6438-17	COMP-4	1.4283	1.4283	200	1.4376	1.4376	1.4376	0.0093	46.5
17	J6440-02	COMP-2	1.4233	1.4233	200	1.4564	1.4564	1.4564	0.0331	165.5
18	J6440-05	COMP-3	1.4210	1.4210	100	1.4392	1.4392	1.4392	0.0182	182
19	J6440-08	COMP-4	1.3844	1.3844	150	1.4048	1.4048	1.4048	0.0204	136
20	J6440-08DUP	COMP-4DUP	1.3882	1.3882	150	1.4092	1.4092	1.4092	0.0210	140

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss121718

WorkList ID : 120452

Date : 12/17/2018 7:57:06 AM

2399848

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2018	Water	J6417-01 C	TSS	Cool 4 deg C	HDR03	E11	121318-TIPC-F-D-B	12/14/2018	SM2540 D
12/18/2018	Water	J6417-02 C	TSS	Cool 4 deg C	HDR03	E11	121318-TIPC-F-D-S	12/14/2018	SM2540 D
12/18/2018	Water	J6417-03 C	TSS	Cool 4 deg C	HDR03	E11	121318-TIPC-F-U-B	12/14/2018	SM2540 D
12/18/2018	Water	J6417-04 C	TSS	Cool 4 deg C	HDR03	E11	121318-TIPC-F-U-S	12/14/2018	SM2540 D
12/20/2018	Water	J6421-01 B	TSS	Cool 4 deg C	HOLL01	E12	EFFLUENT	12/13/2018	SM2540 D
12/20/2018	Water	J6421-02 ✓	TSS	Cool 4 deg C	HOLL01	E12	AER-TANK-1	12/13/2018	SM2540 D
12/20/2018	Water	J6433-01 ✓	TSS	Cool 4 deg C	TULL01	D22	001M-WILLETS-PT-BLVD(NO	12/13/2018	SM2540 D
12/20/2018	Water	J6433-02 ✓	TSS	Cool 4 deg C	TULL01	D22	001M-WILLETS-PT-BLVD(DE	12/13/2018	SM2540 D
12/20/2018	Water	J6433-03 ✓	TSS	Cool 4 deg C	TULL01	D22	002M-35TH-AVE(DEC)	12/13/2018	SM2540 D
12/19/2018	Water	J6438-05 0	TSS	Cool 4 deg C	PSEG01	E21	COMP-1	12/14/2018	SM2540 D
12/19/2018	Water	J6438-09 E	TSS	Cool 4 deg C	PSEG01	E21	COMP-2	12/14/2018	SM2540 D
12/19/2018	Water	J6438-13 D	TSS	Cool 4 deg C	PSEG01	E21	COMP-3	12/14/2018	SM2540 D
12/19/2018	Water	J6438-17 D	TSS	Cool 4 deg C	PSEG01	E21	COMP-4	12/14/2018	SM2540 D
12/19/2018	Water	J6440-02 C	TSS	Cool 4 deg C	PSEG01	E12	COMP-2	12/14/2018	SM2540 D
12/19/2018	Water	J6440-05 C	TSS	Cool 4 deg C	PSEG01	E12	COMP-3	12/14/2018	SM2540 D
12/19/2018	Water	J6440-08 F	TSS	Cool 4 deg C	PSEG01	E12	COMP-4	12/14/2018	SM2540 D

Date/Time 12/17/18 08:00
 Received by: 
 Relinquished by: 

Date/Time 12/17/18 12:10
 Received by: 
 Relinquished by: 