

LB 99850


7470A

INSTRUMENT ID: CW

Linear

μ Abs.:

36864



A= 0.0000e+000

B= 2.6870e-004

C= 1.7142e-003

Rho= 0.9997875

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	7	0.000	7				
0.2	0.200	0.201	0.001	743	0.0 %	743				
2.5	2.500	2.404	-0.096	8941	0.0 %	8941				
5.0	5.000	5.098	0.098	18966	0.0 %	18966				
7.5	7.500	7.586	0.086	28224	0.0 %	28224				
10.0	10.000	9.907	-0.093	36864	0.0 %	36864				

LB99850 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Dec 2018 09:27:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7470A	PPB	17 Dec 2018 09:28:43	S	Std
0.2 - 1	50.2	743	-	0.2000	7470A	PPB	17 Dec 2018 09:32:13	S	Std
2.5 - 1	52.5	8941	-	2.5000	7470A	PPB	17 Dec 2018 09:34:23	S	Std
5.0 - 1	55	18966	-	5.0000	7470A	PPB	17 Dec 2018 09:36:31	S	Std
7.5 - 1	57.5	28224	-	7.5000	7470A	PPB	17 Dec 2018 09:38:39	S	Std
10.0 - 1	57.5	36864	-	10.0000	7470A	PPB	17 Dec 2018 09:43:11	S	Std
ICV28 - 1	ICV28	14956	4.0205	-	7470A	PPB	17 Dec 2018 09:45:39	U	SMPL
ICB28 - 1	ICB28	44	0.0135	-	7470A	PPB	17 Dec 2018 09:47:45	U	SMPL
CCV28 - 1	CCV28	19308	5.1899	-	7470A	PPB	17 Dec 2018 09:49:52	U	SMPL
CCB28 - 1	CCB28	73	0.0213	-	7470A	PPB	17 Dec 2018 09:51:58	U	SMPL
CRA - 1	CRA	815	0.2207	-	7470A	PPB	17 Dec 2018 09:54:05	U	SMPL
HighStd - 1	HighStd	37331	10.0327	-	7470A	PPB	17 Dec 2018 09:56:12	U	SMPL
ChkStd - 1	ChkStd	26164	7.0321	-	7470A	PPB	17 Dec 2018 09:58:19	U	SMPL
PB115707BL - 1	PBW	50	0.0151	-	7470A	PPB	17 Dec 2018 10:00:28	U	SMPL
PB115707BS - 1	LCSW	13627	3.6634	-	7470A	PPB	17 Dec 2018 10:02:36	U	SMPL
J6386-04 - 1	P001-RB-181212	-31	-0.0066	-	7470A	PPB	17 Dec 2018 10:04:44	U	SMPL
J6403-65 - 1	TP-3-OILY-WATER	829634	HIGH	-	7470A	PPB	17 Dec 2018 10:06:52	X	
J6426-06 - 1	P001-RB-181213	230	0.0635	-	7470A	PPB	17 Dec 2018 10:12:01	U	SMPL
J6438-05 - 1	COMP-1	420	0.1146	-	7470A	PPB	17 Dec 2018 10:14:09	U	SMPL
J6438-09 - 1	COMP-2	12	0.0049	-	7470A	PPB	17 Dec 2018 10:16:15	U	SMPL
CCV29 - 1	CCV29	19128	5.1415	-	7470A	PPB	17 Dec 2018 10:18:22	U	SMPL
CCB29 - 1	CCB29	31	0.0100	-	7470A	PPB	17 Dec 2018 10:20:29	U	SMPL
J6438-13 - 1	COMP-3	640	0.1737	-	7470A	PPB	17 Dec 2018 10:22:35	U	SMPL
J6438-17 - 1	COMP-4	-56	-0.0133	-	7470A	PPB	17 Dec 2018 10:24:42	U	SMPL
J6440-02 - 1	COMP-2	193	0.0536	-	7470A	PPB	17 Dec 2018 10:26:48	U	SMPL
J6440-05 - 1	COMP-3	5	0.0031	-	7470A	PPB	17 Dec 2018 10:28:56	U	SMPL
J6440-08 - 1	COMP-4	-16	-0.0026	-	7470A	PPB	17 Dec 2018 10:31:03	U	SMPL
J6440-08DUP - 1	COMP-4DUP	-4	0.0006	-	7470A	PPB	17 Dec 2018 10:33:11	U	SMPL
J6440-08MS - 1	COMP-4MS	13417	3.6069	-	7470A	PPB	17 Dec 2018 10:35:19	U	SMPL
J6440-08MSD - 1	COMP-4MSD	13349	3.5887	-	7470A	PPB	17 Dec 2018 10:37:27	U	SMPL
J6440-08LX5 - 1		-4	0.0006	-	7470A	PPB	17 Dec 2018 10:39:36	U	SMPL
J6440-08A - 1		14067	3.7816	-	7470A	PPB	17 Dec 2018 10:41:43	U	SMPL
CCV30 - 1	CCV30	19255	5.1756	-	7470A	PPB	17 Dec 2018 10:43:49	U	SMPL
CCB30 - 1	CCB30	3	0.0025	-	7470A	PPB	17 Dec 2018 10:45:57	U	SMPL
J6403-65DLX10 - 1		173702	46.6762	-	7470A	PPB	17 Dec 2018 10:48:04	U	SMPL
J6403-65DLX100 - 1		16902	4.5434	-	7470A	PPB	17 Dec 2018 10:50:11	U	SMPL
CCV31 - 1	CCV31	19250	5.1743	-	7470A	PPB	17 Dec 2018 10:52:23	U	SMPL
CCB31 - 1	CCB31	60	0.0178	-	7470A	PPB	17 Dec 2018 10:54:30	U	SMPL