



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99853

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	12/14/2018	09:50
2	J6373-11		1	9	>100.0	>212.0	>212.0	12/14/2018	10:20
3	J6373-11DUP		1	9	>100.0	>212.0	>212.0	12/14/2018	10:45
4	J6376-08		1	7	>100.0	>212.0	>212.0	12/14/2018	11:15
5	J6390-04		1	6	>100.0	>212.0	>212.0	12/14/2018	11:40
6	J6390-05		1	9	>100.0	>212.0	>212.0	12/14/2018	12:05
7	J6390-06		1	8	>100.0	>212.0	>212.0	12/14/2018	12:35
8	J6390-07		1	8	>100.0	>212.0	>212.0	12/14/2018	13:10
9	J6390-08		1	8	>100.0	>212.0	>212.0	12/14/2018	13:45
10	J6390-09		1	9	>100.0	>212.0	>212.0	12/14/2018	14:15
11	J6403-66		1	7	>100.0	>212.0	>212.0	12/14/2018	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 99833

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp121418

WorkList ID : 120351

Date : 12/14/2018 9:17:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2018	Water	J6373-11	Flash Point	Cool 4 deg C	PSEG01	D12	OILY-WATER	12/12/2018	1010A
12/17/2018	Water	J6376-08	Flash Point	Cool 4 deg C	PSEG01	D12	TEST-PIT-1-WATER	12/12/2018	1010A
12/20/2018	Water	J6390-04	Flash Point	Cool 4 deg C	PSEG01	D21	30098	12/13/2018	1010A
12/20/2018	Water	J6390-05	Flash Point	Cool 4 deg C	PSEG01	D21	30104	12/13/2018	1010A
12/20/2018	Water	J6390-06	Flash Point	Cool 4 deg C	PSEG01	D21	30099	12/13/2018	1010A
12/20/2018	Water	J6390-07	Flash Point	Cool 4 deg C	PSEG01	D21	30082	12/13/2018	1010A
12/20/2018	Water	J6390-08	Flash Point	Cool 4 deg C	PSEG01	D21	30081	12/13/2018	1010A
12/20/2018	Water	J6390-09	Flash Point	Cool 4 deg C	PSEG01	D21	30080	12/13/2018	1010A
12/18/2018	Water	J6403-66	Flash Point	Cool 4 deg C	PSEG01	D22	TP-3-OILY-WATER	12/13/2018	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: