



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99856

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.2	12/17/2018	09:05
2	J6390-02		1	4	21.0	69.8	70.4	12/17/2018	09:40
3	J6390-02DUP		1	4	21.0	69.8	70.4	12/17/2018	10:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 99856

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp121418-1

WorkList ID : 120352

Date : 12/14/2018 9:18:41 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/20/2018	Solid	J6390-02 B	Flash Point	Cool 4 deg C	PSEG01	D21	50051	12/13/2018	1010A

Date/Time 12/17/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 12/17/18 10:15 AM
Received by: CP
Relinquished by: KS