

2B99862

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/17/2018 12:22

Test: *Deactive*
Total CN

Sample Id	Result <i>ug/l</i>	Dil. 1 +	Response	Errors
ICV1	95.069	0.0	0.088	
ICB1	0.919	0.0	0.004	
CCV1	239.024	0.0	0.216	
CCB1	0.706	0.0	0.004	
PB115769BL	0.656	0.0	0.004	
J6425-10	0.748	0.0	0.004	
J6425-10DUP	0.729	0.0	0.004	
J6438-06	0.876	0.0	0.004	
J6438-10	1.502	0.0	0.004	
J6438-14	2.086	0.0	0.005	
J6440-03	0.616	0.0	0.004	
J6440-06	1.232	0.0	0.004	
J6440-09	0.768	0.0	0.004	
J6438-18	0.669	0.0	0.004	
CCV2	253.249	0.0	0.229	
CCB2	0.589	0.0	0.004	

N	16
Mean	37.465
SD	84.7982
CV%	226.34

Aquakem v. 7.2AQ1

LA

Results from time period:

Mon Dec 17 11:52:00 2018

Mon Dec 17 12:22:15 2018

Sample Id	Sam/Ctr/c	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.1334	µg/l	12/17/2018 7:37:39	
5.OPPBCN	A	Reactive CN	P	4.7569	µg/l	12/17/2018 7:37:40	
10PPBCN	A	Reactive CN	P	9.8163	µg/l	12/17/2018 7:37:41	
50PPBCN	A	Reactive CN	P	50.4651	µg/l	12/17/2018 7:37:42	
100PPBCN	A	Reactive CN	P	100.1044	µg/l	12/17/2018 7:37:43	
250PPBCN	A	Reactive CN	P	249.5705	µg/l	12/17/2018 7:37:44	
500PPBCN	A	Reactive CN	P	500.1534	µg/l	12/17/2018 7:37:45	
ICV1	S	Reactive CN	P	95.069	µg/l	12/17/2018 11:52:00	
ICB1	S	Reactive CN	P	0.9192	µg/l	12/17/2018 11:52:01	
CCV1	S	Reactive CN	P	239.0237	µg/l	12/17/2018 11:52:02	
CCB1	S	Reactive CN	P	0.7059	µg/l	12/17/2018 11:52:03	
PB115769BL	S	Reactive CN	P	0.6564	µg/l	12/17/2018 11:52:04	
J6425-10	S	Reactive CN	P	0.7478	µg/l	12/17/2018 11:52:05	
J6425-10DUP	S	Reactive CN	P	0.7294	µg/l	12/17/2018 11:52:06	
J6438-06	S	Reactive CN	P	0.8758	µg/l	12/17/2018 11:52:07	
J6438-10	S	Reactive CN	P	1.5018	µg/l	12/17/2018 11:52:08	
J6438-14	S	Reactive CN	P	2.0865	µg/l	12/17/2018 11:52:09	
J6440-03	S	Reactive CN	P	0.6163	µg/l	12/17/2018 11:56:47	
J6440-06	S	Reactive CN	P	1.2318	µg/l	12/17/2018 11:56:48	
J6440-09	S	Reactive CN	P	0.7682	µg/l	12/17/2018 11:56:49	
J6438-18	S	Reactive CN	P	0.6691	µg/l	12/17/2018 12:22:13	
CCV2	S	Reactive CN	P	253.2488	µg/l	12/17/2018 12:22:14	
CCB2	S	Reactive CN	P	0.5888	µg/l	12/17/2018 12:22:15	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/17/2018 7:38


Test Total CN

Accepted 12/17/2018 7:37

Factor 1120
Bias 0.003

Coeff. of det. 0.999997

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1334	0.0000	
2	5.0PPBCN	0.007	4.7569	5.0000	
3	10PPBCN	0.012	9.8163	10.0000	
4	50PPBCN	0.048	50.4651	50.0000	
5	100PPBCN	0.092	100.1044	100.0000	
6	250PPBCN	0.226	249.5705	250.0000	
7	500PPBCN	0.450	500.1534	500.0000	