



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99871

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	81.2	12/17/2018	11:40
2	J6425-10		1	7	>100.0	>212.0	>212.0	12/17/2018	12:05
3	J6425-10DUP		1	7	>100.0	>212.0	>212.0	12/17/2018	12:45
4	J6438-06		1	9	>100.0	>212.0	>212.0	12/17/2018	13:10
5	J6438-10		1	7	>100.0	>212.0	>212.0	12/17/2018	13:45
6	J6438-14		1	9	>100.0	>212.0	>212.0	12/17/2018	14:15
7	J6438-18		1	8	>100.0	>212.0	>212.0	12/17/2018	14:40
8	J6440-03		1	6	>100.0	>212.0	>212.0	12/17/2018	15:05
9	J6440-06		1	8	>100.0	>212.0	>212.0	12/17/2018	15:32
10	J6440-09		1	7	>100.0	>212.0	>212.0	12/17/2018	15:45
11	J6440-11		1	8	>100.0	>212.0	>212.0	12/17/2018	16:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99871

WorkList Name : fp121718

WorkList ID : 120474

Date : 12/17/2018 11:03:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/21/2018	Water	J6425-10	Flash Point	Cool 4 deg C	PARS02	D62	WC-20181214	12/14/2018	1010A
12/19/2018	Water	J6438-06	Flash Point	Cool 4 deg C	PSEG01	E21	COMP-1	12/14/2018	1010A
12/19/2018	Water	J6438-10	Flash Point	Cool 4 deg C	PSEG01	E21	COMP-2	12/14/2018	1010A
12/19/2018	Water	J6438-14	Flash Point	Cool 4 deg C	PSEG01	E21	COMP-3	12/14/2018	1010A
12/19/2018	Water	J6438-18	Flash Point	Cool 4 deg C	PSEG01	E21	COMP-4	12/14/2018	1010A
12/19/2018	Water	J6440-03	Flash Point	Cool 4 deg C	PSEG01	E12	COMP-2	12/14/2018	1010A
12/19/2018	Water	J6440-06	Flash Point	Cool 4 deg C	PSEG01	E12	COMP-3	12/14/2018	1010A
12/19/2018	Water	J6440-09	Flash Point	Cool 4 deg C	PSEG01	E12	COMP-4	12/14/2018	1010A
12/19/2018	Water	J6440-11	Flash Point	Cool 4 deg C	PSEG01	E12	Diesel	12/14/2018	1010A

Date/Time 12/17/18 11:10 AM
 Received by: KS
 Relinquished by: CP

Date/Time 12/17/18 16:20
 Received by: CP
 Relinquished by: KS