



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99872

Slope : 99.5

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/17/2018	14:00
2	CAL2	1	Water			20.1	7.00	12/17/2018	14:02
3	CAL3	1	Water			20.2	10.04	12/17/2018	14:04
4	ICV	1	Water			20.3	7.01	12/17/2018	14:06
5	CCV1	1	Water			20.2	2.00	12/17/2018	14:08
6	J6459-01	1	Water			18.9	5.71	12/17/2018	14:10
7	J6459-01DUP	1	Water			19.0	5.72	12/17/2018	14:12
8	J6460-01	1	Water			19.2	4.59	12/17/2018	14:14
9	CCV2	1	Water			20.1	12.00	12/17/2018	14:16

WORKLIST(Hardcopy Internal Chain)

LB 99 872

WorkList Name : pH121718

WorkList ID : 120484

Date : 12/17/2018 12:08:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/14/2018	Water	J6459-01	pH	Cool 4 deg C	STER08	D31	POOL-WATER	12/14/2018	SM 4500-H B
12/13/2018	Water	J6460-01	pH	Cool 4 deg C	STER04	D41	POOL-WATER	12/13/2018	SM 4500-H B

Date/Time 12/17/18 1:50
 Received by: SP
 Relinquished by: SP

Date/Time 12/17/18 2:20
 Received by: SP
 Relinquished by: SP