



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99889

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/17/2018	16:30
2	CAL2	1	Water			20.1	7.00	12/17/2018	16:32
3	CAL3	1	Water			20.2	10.04	12/17/2018	16:34
4	ICV	1	Water			20.3	7.01	12/17/2018	16:36
5	CCV1	1	Water			20.2	2.00	12/17/2018	16:38
6	J6462-03	1	Water			12.3	8.42	12/17/2018	16:40
7	J6462-04	1	Water			12.0	7.17	12/17/2018	16:42
8	J6462-05	1	Water			13.6	8.54	12/17/2018	16:44
9	J6462-06	1	Water			14.9	8.97	12/17/2018	16:46
10	J6462-06DUP	1	Water			15.0	8.98	12/17/2018	16:48
11	CCV2	1	Water			20.2	12.00	12/17/2018	16:50

WORKLIST(Hardcopy Internal Chain)

LB 99889

WorkList Name : pH121718-1

WorkList ID : 120497

Date : 12/17/2018 3:38:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/17/2018	Water	J6462-03	pH	Cool 4 deg C	PSEG01	E21	OW-1-5	12/17/2018	9040C
12/17/2018	Water	J6462-04	pH	Cool 4 deg C	PSEG01	E21	OW-6-11	12/17/2018	9040C
12/17/2018	Water	J6462-05	pH	Cool 4 deg C	PSEG01	E21	OW-12-17	12/17/2018	9040C
12/17/2018	Water	J6462-06	pH	Cool 4 deg C	PSEG01	E21	OILY-WATER	12/17/2018	9040C

Date/Time 12/17/18 4:15
 Received by: SP
 Relinquished by: SP

Date/Time 12/17/18 5:00
 Received by: CP
 Relinquished by: SP