



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99897

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	12/18/2018	09:50
2	J6462-07		1	8	>100.0	>212.0	>212.0	12/18/2018	10:25
3	J6462-07DUP		1	8	>100.0	>212.0	>212.0	12/18/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB99897

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP121818

WorkList ID : 120527

Date : 12/18/2018 9:12:31 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/20/2018	Water	J6462-07	Flash Point	Cool 4 deg C	PSEG01	E21	OILY-WATER	12/17/2018	1010A

Date/Time

Received by:

Relinquished by:

12/18/18 9:15 AM
KS
CP

Date/Time

Received by:

Relinquished by:

12/18/18 11:00 AM
CP
KS