

21399904

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/18/2018 14:49

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.138	0.0	0.085	
ICB1	0.603	0.0	0.004	
CCV1	231.458	0.0	0.209	
CCB1	0.329	0.0	0.004	
PB115813BL	0.773	0.0	0.004	
PB115813BS	93.586	0.0	0.086	
J6462-06	19.096	0.0	0.020	
J6462-06DUP	19.334	0.0	0.020	
J6462-06MS	58.520	0.0	0.055	
J6462-06MSD	58.750	0.0	0.055	
LOWPB115813	9.619	0.0	0.012	
HIGHPB115813	483.740	0.0	0.433	
CCV2	244.830	0.0	0.221	
CCB2	0.416	0.0	0.004	

N 14
Mean 93.799
SD 138.4803
CV% 147.63

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 18 14:14:26 2018

Tue Dec 18 14:33:20 2018

12

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.5977	µg/l	12/18/2018 9:36:39	
5.0PPBCN	A	Total CN	P	4.7648	µg/l	12/18/2018 9:36:40	
10PPBCN	A	Total CN	P	9.8972	µg/l	12/18/2018 9:36:41	
50PPBCN	A	Total CN	P	47.9767	µg/l	12/18/2018 9:36:42	
100PPBCN	A	Total CN	P	100.9307	µg/l	12/18/2018 9:36:43	
250PPBCN	A	Total CN	P	251.6245	µg/l	12/18/2018 9:36:44	
500PPBCN	A	Total CN	P	499.2083	µg/l	12/18/2018 9:36:45	
ICV1	S	Total CN	P	92.1382	µg/l	12/18/2018 14:14:26	
ICB1	S	Total CN	P	0.6032	µg/l	12/18/2018 14:14:27	
CCV1	S	Total CN	P	231.4576	µg/l	12/18/2018 14:14:28	
CCB1	S	Total CN	P	0.3291	µg/l	12/18/2018 14:14:29	
PB115813BL	S	Total CN	P	0.7734	µg/l	12/18/2018 14:14:30	
PB115813BS	S	Total CN	P	93.5857	µg/l	12/18/2018 14:14:31	
J6462-06	S	Total CN	P	19.0961	µg/l	12/18/2018 14:14:32	
J6462-06DUP	S	Total CN	P	19.3337	µg/l	12/18/2018 14:14:33	
J6462-06MS	S	Total CN	P	58.52	µg/l	12/18/2018 14:14:34	
J6462-06MSD	S	Total CN	P	58.7497	µg/l	12/18/2018 14:14:35	
LOWPB115813	S	Total CN	P	9.6185	µg/l	12/18/2018 14:19:14	
HIGHPB115813	S	Total CN	P	483.7404	µg/l	12/18/2018 14:19:15	
CCV2	S	Total CN	P	244.8297	µg/l	12/18/2018 14:19:16	
CCB2	S	Total CN	P	0.416	µg/l	12/18/2018 14:19:17	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AB

12/18/2018 9:39


Test Total CN

Accepted 12/18/2018 9:39

Factor 1125
Bias 0.003

Coeff. of det. 0.999958

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.5977	0.0000	
2	5.0PPBCN	0.007	4.7648	5.0000	
3	10PPBCN	0.012	9.8972	10.0000	
4	50PPBCN	0.046	47.9767	50.0000	
5	100PPBCN	0.093	100.9307	100.0000	
6	250PPBCN	0.227	251.6245	250.0000	
7	500PPBCN	0.447	499.2083	500.0000	