

LB 99907


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

58202



A= 0.0000e+000

B= 1.7276e-004

C= 8.5426e-003

Rho= 0.9999336

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	62	0.000	62				
0.2	0.200	0.203	0.003	1127	0.0 %	1127				
2.5	2.500	2.519	0.019	14534	0.0 %	14534				
5.0	5.000	4.962	-0.038	28671	0.0 %	28671				
7.5	7.500	7.432	-0.068	42971	0.0 %	42971				
10.0	10.000	10.064	0.064	58202	0.0 %	58202				

LB99907 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Dec 2018 16:36:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	62	-	0.00007471B	PPB		18 Dec 2018 16:37:56	S	Std
0.2 - 1	50.2	1127	-	0.20007471B	PPB		18 Dec 2018 16:40:14	S	Std
2.5 - 1	52.5	14534	-	2.50007471B	PPB		18 Dec 2018 16:42:32	S	Std
5.0 - 1	55	28671	-	5.00007471B	PPB		18 Dec 2018 16:44:50	S	Std
7.5 - 1	57.5	42971	-	7.50007471B	PPB		18 Dec 2018 16:47:07	S	Std
10.0 - 1	510	58202	-	10.00007471B	PPB		18 Dec 2018 16:49:25	S	Std
ICV31 - 1	ICV31	21988	3.8073	-7471B	PPB		18 Dec 2018 16:52:10	U	SMPL
ICB31 - 1	ICB31	-70	-0.0036	-7471B	PPB		18 Dec 2018 16:54:26	U	SMPL
CCV42 - 1	CCV42	27432	4.7478	-7471B	PPB		18 Dec 2018 16:59:46	U	SMPL
CCB42 - 1	CCB42	-82	-0.0056	-7471B	PPB		18 Dec 2018 17:02:01	U	SMPL
CRA - 1	CRA	1061	0.1918	-7471B	PPB		18 Dec 2018 17:04:18	U	SMPL
HighStd - 1	HighStd	56312	9.7372	-7471B	PPB		18 Dec 2018 17:06:34	U	SMPL
ChkStd - 1	ChkStd	38605	6.6781	-7471B	PPB		18 Dec 2018 17:08:51	U	SMPL
PB115794BL - 1	PBS	-211	-0.0279	-7471B	PPB		18 Dec 2018 17:11:09	U	SMPL
PB115794BS - 1	LCSS	21931	3.7974	-7471B	PPB		18 Dec 2018 17:15:59	U	SMPL
J6195-09 - 1	OR-03-121718	2716	0.4778	-7471B	PPB		18 Dec 2018 17:18:15	U	SMPL
J6210-07 - 1	C-1-112918	12790	2.2182	-7471B	PPB		18 Dec 2018 17:20:33	U	SMPL
J6210-09 - 1	C-2-112918	614	0.1146	-7471B	PPB		18 Dec 2018 17:22:50	U	SMPL
J6210-09DUP - 1	C-2-112918DUP	711	0.1314	-7471B	PPB		18 Dec 2018 17:25:08	U	SMPL
J6210-09MS - 1	C-2-112918MS	21673	3.7529	-7471B	PPB		18 Dec 2018 17:27:24	U	SMPL
CCV43 - 1	CCV43	27549	4.7680	-7471B	PPB		18 Dec 2018 17:29:40	U	SMPL
CCB43 - 1	CCB43	-107	-0.0099	-7471B	PPB		18 Dec 2018 17:31:56	U	SMPL
J6210-09MSD - 1	C-2-112918MSD	21946	3.8000	-7471B	PPB		18 Dec 2018 17:34:13	U	SMPL
J6319-08 - 1	062_SW-S-1	1401	0.2506	-7471B	PPB		18 Dec 2018 17:36:29	U	SMPL
J6319-11 - 1	065_SW-W-1	23276	4.0298	-7471B	PPB		18 Dec 2018 17:38:46	U	SMPL
J6319-12 - 1	066_SW-W-2	12522	2.1719	-7471B	PPB		18 Dec 2018 17:41:03	U	SMPL
J6453-01 - 1	SS-SOIL	2231	0.3940	-7471B	PPB		18 Dec 2018 17:43:21	U	SMPL
J6461-01 - 1	NCC-KEARNY-TOPSOIL	20686	3.5823	-7471B	PPB		18 Dec 2018 17:45:37	U	SMPL
J6464-01 - 1	290-RT4008-RBR251686	21021	3.6402	-7471B	PPB		18 Dec 2018 17:47:55	U	SMPL
J6464-04 - 1	VB16202	2612	0.4598	-7471B	PPB		18 Dec 2018 17:50:13	U	SMPL
J6210-09LX5 - 1		105	0.0267	-7471B	PPB		18 Dec 2018 17:52:31	U	SMPL
J6210-09A - 1		20260	3.5087	-7471B	PPB		18 Dec 2018 17:54:47	U	SMPL
CCV44 - 1	CCV44	28201	4.8807	-7471B	PPB		18 Dec 2018 17:57:03	U	SMPL
CCB44 - 1	CCB44	-99	-0.0086	-7471B	PPB		18 Dec 2018 17:59:20	U	SMPL