


LB 99917

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.5845e-004

C= 3.4740e-002

Rho= 0.9998948

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	-20	0.000	-20				
0.2	0.200	0.210	0.010	679	0.0 %	679				
2.5	2.500	2.430	-0.070	9269	0.0 %	9269				
5.0	5.000	4.982	-0.018	19141	0.0 %	19141				
7.5	7.500	7.594	0.094	29250	0.0 %	29250				
10.0	10.000	9.954	-0.046	38380	0.0 %	38380				

LB99917 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Dec 2018 09:58:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-20	-	0.0000	7470A	PPB	19 Dec 2018 10:03:16	S	Std
0.2 - 1	50.2	679	-	0.2000	7470A	PPB	19 Dec 2018 10:05:25	S	Std
2.5 - 1	52.5	9269	-	2.5000	7470A	PPB	19 Dec 2018 10:07:34	S	Std
5.0 - 1	55	19141	-	5.0000	7470A	PPB	19 Dec 2018 10:09:42	S	Std
7.5 - 1	57.5	29250	-	7.5000	7470A	PPB	19 Dec 2018 10:11:50	S	Std
10.0 - 1	510	38380	-	10.0000	7470A	PPB	19 Dec 2018 10:16:21	S	Std
ICV32 - 1	ICV32	14834	3.8685	-	7470A	PPB	19 Dec 2018 10:18:55	U	SMPL
ICB32 - 1	ICB32	-99	0.0092	-	7470A	PPB	19 Dec 2018 10:21:02	U	SMPL
CCV45 - 1	CCV45	19957	5.1926	-	7470A	PPB	19 Dec 2018 10:23:09	U	SMPL
CCB45 - 1	CCB45	-112	0.0058	-	7470A	PPB	19 Dec 2018 10:25:15	U	SMPL
CRA - 1	CRA	670	0.2079	-	7470A	PPB	19 Dec 2018 10:27:22	U	SMPL
HighStd - 1	HighStd	37968	9.8475	-	7470A	PPB	19 Dec 2018 10:29:29	U	SMPL
ChkStd - 1	ChkStd	27044	7.0242	-	7470A	PPB	19 Dec 2018 10:31:36	U	SMPL
PB115816BL - 1	PBW	-150	-0.0040	-	7470A	PPB	19 Dec 2018 10:33:44	U	SMPL
J6195-10 - 1	OR-03-121718	-58	0.0197	-	7470A	PPB	19 Dec 2018 10:38:00	U	SMPL
J6195-10DUP - 1	OR-03-121718DUP	-46	0.0229	-	7470A	PPB	19 Dec 2018 10:40:08	U	SMPL
J6195-10MS - 1	OR-03-121718MS	14976	3.9052	-	7470A	PPB	19 Dec 2018 10:45:05	U	SMPL
J6195-10MSD - 1	OR-03-121718MSD	14638	3.8179	-	7470A	PPB	19 Dec 2018 10:47:13	U	SMPL
J6453-03 - 1	SS-SOIL	-92	0.0110	-	7470A	PPB	19 Dec 2018 10:49:20	U	SMPL
PB115816BS - 1	LCSW	14817	3.8641	-	7470A	PPB	19 Dec 2018 10:51:43	U	SMPL
CCV46 - 1	CCV46	20260	5.2709	-	7470A	PPB	19 Dec 2018 10:54:09	U	SMPL
CCB46 - 1	CCB46	-101	0.0086	-	7470A	PPB	19 Dec 2018 10:56:16	U	SMPL
J6462-01 - 1	OILY-DEBRIS	-45	0.0231	-	7470A	PPB	19 Dec 2018 10:58:22	U	SMPL
J6462-02 - 1	PAINT-MATERIAL	-33	0.0262	-	7470A	PPB	19 Dec 2018 11:00:29	U	SMPL
J6464-03 - 1	290-RT4008-RBR251686	-28	0.0275	-	7470A	PPB	19 Dec 2018 11:02:37	U	SMPL
J6464-06 - 1	VB16202	-23	0.0288	-	7470A	PPB	19 Dec 2018 11:04:44	U	SMPL
PB115818BL - 1	PBW	-17	0.0303	-	7470A	PPB	19 Dec 2018 11:06:52	U	SMPL
PB115818BS - 1	LCSW	14620	3.8132	-	7470A	PPB	19 Dec 2018 11:09:00	U	SMPL
J6446-07 - 1	P001-RB-181214	-139	-0.0012	-	7470A	PPB	19 Dec 2018 11:11:08	U	SMPL
J6462-06 - 1	OILY-WATER	103	0.0614	-	7470A	PPB	19 Dec 2018 11:13:17	U	SMPL
J6462-06DUP - 1	OILY-WATERDUP	76	0.0544	-	7470A	PPB	19 Dec 2018 11:15:26	U	SMPL
J6462-06MS - 1	OILY-WATERMS	14962	3.9016	-	7470A	PPB	19 Dec 2018 11:17:33	U	SMPL
CCV47 - 1	CCV47	20172	5.2481	-	7470A	PPB	19 Dec 2018 11:19:40	U	SMPL
CCB47 - 1	CCB47	-103	0.0081	-	7470A	PPB	19 Dec 2018 11:21:47	U	SMPL
J6462-06MSD - 1	OILY-WATERMSD	14674	3.8272	-	7470A	PPB	19 Dec 2018 11:23:55	U	SMPL
PB115817BL - 1	PBW	-84	0.0130	-	7470A	PPB	19 Dec 2018 11:26:03	U	SMPL
PB115817BS - 1	LCSW	14827	3.8667	-	7470A	PPB	19 Dec 2018 11:28:10	U	SMPL
J6319-17 - 1	061_SW-N-1	377	0.1322	-	7470A	PPB	19 Dec 2018 11:30:18	U	SMPL
J6319-18 - 1	063_SW-E-1	-109	0.0066	-	7470A	PPB	19 Dec 2018 11:32:27	U	SMPL
J6319-19 - 1	064_SW-E-2	-85	0.0128	-	7470A	PPB	19 Dec 2018 11:34:35	U	SMPL
J6319-20 - 1	068-DUP-2	-72	0.0161	-	7470A	PPB	19 Dec 2018 11:36:44	U	SMPL
PB115816TB - 1	PB115788TB	-21	0.0293	-	7470A	PPB	19 Dec 2018 11:38:53	U	SMPL
PB115817TB - 1	PB115790TB	-75	0.0154	-	7470A	PPB	19 Dec 2018 11:41:02	U	SMPL
CCV48 - 1	CCV48	20441	5.3177	-	7470A	PPB	19 Dec 2018 11:43:10	U	SMPL
CCB48 - 1	CCB48	-66	0.0177	-	7470A	PPB	19 Dec 2018 11:45:18	U	SMPL
J6195-10LX5 - 1		-25	0.0283	-	7470A	PPB	19 Dec 2018 11:47:26	U	SMPL
J6195-10A - 1		13933	3.6357	-	7470A	PPB	19 Dec 2018 11:49:34	U	SMPL
J6462-06LX5 - 1		-48	0.0223	-	7470A	PPB	19 Dec 2018 11:51:41	U	SMPL
J6462-06A - 1		14802	3.8603	-	7470A	PPB	19 Dec 2018 11:53:49	U	SMPL
CCV49 - 1	CCV49	19083	4.9667	-	7470A	PPB	19 Dec 2018 11:55:56	U	SMPL
CCB49 - 1	CCB49	-100	0.0089	-	7470A	PPB	19 Dec 2018 11:58:04	U	SMPL